



Analytical Resources, Incorporated
Analytical Chemists and Consultants

04 January 2019

Karen Mixon
AECOM
1111 Third Avenue, Suite 1600
Seattle, WA 98101

RE: Portland Harbor Pre-Remedial Design

Please find enclosed sample receipt documentation and analytical results for samples from the project referenced above.

Sample analyses were performed according to ARI's Quality Assurance Plan and any provided project specific Quality Assurance Plan. Each analytical section of this report has been approved and reviewed by an analytical peer, the appropriate Laboratory Supervisor or qualified substitute, and a technical reviewer.

Should you have any questions or problems, please feel free to contact us at your convenience.

Associated Work Order(s)
18L0044

Associated SDG ID(s)
N/A

I certify that this data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed in the enclose Narrative. ARI, an accredited laboratory, certifies that the report results for which ARI is accredited meets all the requirements of the accrediting body. A list of certified analyses, accreditations, and expiration dates is included in this report.

Release of the data contained in this hardcopy data package has been authorized by the Laboratory Manager or his/her designee, as verified by the following signature.

Analytical Resources, Inc.

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Chain of Custody Record & Laboratory Analysis Request

ARI Assigned Number: 18L0044	Turn-around Requested:
ARI Client Company: AECOM	Phone: (206) 438-2261 / (206) 438-2010

Date:	12/03/2018
Page:	1 of 1
No. of Coolers:	1
Cooler Temps:	2.6



Analytical Resources, Incorporated
Analytical Chemists and Consultants
4611 South 134th Place, Suite 100
Tukwila, WA 98168
206-695-6200 206-695-6201 (fax)

Client Contact:	Amy Dahl/Chelsey Cook
-----------------	------------------------------

Client Project Name:	Portland Harbor Pre-Remedial Design Investigation and Baseline Sampling
----------------------	---

Client Project #:	60566335
Samplers:	ED and BW

Sample ID	Date	Time	Matrix	No. Containers
PDI-WS-T02-1811	11/30/18	15:06	W	2
PDI-WS-T06-1811	11/30/18	16:26	W	2
PDI-WS-T04-1811	12/01/18	13:10	W	2

Analysis Requested										Notes/Comments	
PCP 8041											

Comments/Special Instructions	Relinquished by: (Signature) [Signature]	Received by: (Signature) [Signature]	Relinquished by: (Signature) [Signature]	Received by: (Signature) [Signature]
	Printed Name: REBECCA TORTORELLO	Printed Name: Fedex	Printed Name: Stephanie Fisher	Printed Name: Stephanie Fisher
	Company: AECOM	Company: Fed ex	Company: ARI	Company: ARI
	Date & Time: 12/3/2018 1500	Date & Time: 12/3/2018 1500	Date & Time: 12-4-18 1350	Date & Time: 12-4-18 1350

Limits of Liability: ARI will perform all requested services in accordance with appropriate methodology following ARI Standard Operating Procedures and the ARI Quality Assurance Program. This program meets standards for the industry. The total liability of ARI, its officers, agents, employees, or successors, arising out of or in connection with the requested services, shall not exceed the Invoiced amount for said services. The acceptance by the client of a proposal for services by ARI release ARI from any liability in excess thereof, notwithstanding any provision to the contrary in any contract, purchase order or co-signed agreement between ARI and the Client.

Sample Retention Policy: Unless specified by workorder or contract, all water/soil samples submitted to ARI will be discarded or returned, no sooner than 90 days after receipt or 60 days after submission of hardcopy data, whichever is longer. Sediment samples submitted under PSDDA/PSEP/SMS protocol will be stored frozen for up to one year and then discarded.

RE: 18L0044 Portland Harbor Sample Acknowledgment

Dahl, Amy

Wed 12/5/2018 6:53 PM

To: Amanda Volgardsen <amanda.volgardsen@arilabs.com>;

Cc: Cook, Chelsey <Chelsey.Cook@aecom.com>; Mixon, Karen <karen.mixon@aecom.com>;

📎 3 attachments (1 MB)

18L0044_coc.pdf; 18L0044_crf.pdf; 18L0044_info.pdf;

Hi Amanda, I just noticed something from a different COC, please change the following sample ID and mark on COC "per AECOM":

18L0044-03 "PDI-WS-T04-181¹" should be "PDI-WS-T04-181²"

Thank you,

PRIVILEGED AND CONFIDENTIAL / JOINT DEFENSE COMMUNICATION / ATTORNEY CLIENT WORK PRODUCT

Amy Dahl, PhD

Chemist, Environment, Pacific Northwest

D +1-206-438-2261

amy.dahl@aecom.com

AECOM

1111 Third Avenue, Suite 1600

Seattle, WA 98101, United States

T +1-206-438-2700

aecom.com

From: Dahl, Amy

Sent: Wednesday, December 05, 2018 6:04 PM

To: 'Amanda Volgardsen'; Cook, Chelsey; Mixon, Karen

Subject: RE: 18L0044 Portland Harbor Sample Acknowledgment

Looks good, thanks Amanda!

Amy Dahl, PhD

Chemist, Environment, Pacific Northwest

D +1-206-438-2261

amy.dahl@aecom.com

AECOM

1111 Third Avenue, Suite 1600

Seattle, WA 98101, United States

ARI Assigned Number: 1060044	Turn-around Requested:
ARI Client Company: AECOM	Phone: (206) 438-2261 / (206) 438-2010
Client Contact: Amy Dahl/Chelsey Cook	

Page: 1 of 1

No. of Coolers: 1 Cooler Temps: 2.6



Analytical Resources, Incorporated
Analytical Chemists and Consultants
4611 South 134th Place, Suite 100
Tukwila, WA 98168
206-695-6200 206-695-6201 (fax)

Client Project Name: Portland Harbor Pre-Remedial Design Investigation and Baseline Sampling

Client Project #: 60566335	Samplers: ED and BW
----------------------------	---------------------


Analysis Requested

Notes/Comments

PCP 8041

[illegible]

Comments/Special Instructions	
-------------------------------	--

Relinquished by: (Signature)	
Printed Name:	REBECCA TORTORELLO
Company:	AELOM
Date & Time:	12/3/2018 1500

Received by:	
(Signature)	
Printed Name:	Fed ex
Company:	Fed ex
Date & Time:	12/3/2019 1500

Relinquished by:	
(Signature)	
Printed Name:	
Company:	
Date & Time:	

Received by:	(Signature) <i>Stephanie Fisher</i>
Printed Name:	<i>Stephanie Fisher</i>
Company:	<i>AOL</i>
Date & Time:	<i>12-4-16 1350</i>

Sample Retention Policy: Unless specified by workorder or contract, all water/soil samples submitted to ARI will be discarded or returned, no sooner than 90 days after receipt or 60 days after submission of hardcopy data, whichever is longer. Sediment samples submitted under PSDDA/PSEP/SMS protocol will be stored frozen for up to one year and then discarded.



AECOM
1111 Third Avenue, Suite 1600
Seattle WA, 98101

Project: Portland Harbor Pre-Remedial Design
Project Number: 60566335
Project Manager: Karen Mixon

Reported:
04-Jan-2019 12:58

ANALYTICAL REPORT FOR SAMPLES

Sample ID	Laboratory ID	Matrix	Date Sampled	Date Received
PDI-WS-T02-1811	18L0044-01	Water	30-Nov-2018 15:06	04-Dec-2018 13:50
PDI-WS-T06-1811	18L0044-02	Water	30-Nov-2018 16:26	04-Dec-2018 13:50
PDI-WS-T04-1812	18L0044-03	Water	01-Dec-2018 13:10	04-Dec-2018 13:50



AECOM
1111 Third Avenue, Suite 1600
Seattle WA, 98101

Project: Portland Harbor Pre-Remedial Design
Project Number: 60566335
Project Manager: Karen Mixon

Reported:
04-Jan-2019 12:58

Work Order Case Narrative

Sample receipt

Samples as listed on the preceding page were received December 4, 2018 under ARI work order 18L0044. For details regarding sample receipt, please refer to the Cooler Receipt Form.

Pentachlorophenol - EPA Method SW8041A

The samples were extracted and analyzed within the recommended holding times.

Initial and continuing calibrations were within method requirements.

The surrogate percent recoveries were within control limits.

The method blank was clean at the reporting limits.

The LCS percent recoveries were within control limits.

A matrix spike and matrix spike duplicate were prepared in conjunction with sample PDI-WS-T02-1811. The matrix spike/matrix spike duplicate percent recoveries and RPD were within QC limits.



Analytical Resources, Incorporated
Analytical Chemists and Consultants

Cooler Receipt Form

ARI Client: Alcon

COC No(s): _____ (NA)

Assigned ARI Job No: 18L0044

Project Name: Portland Harbor

Delivered by: Fed-Ex UPS Courier Hand Delivered Other: _____

Tracking No: 784161773045 NA

Preliminary Examination Phase:

Were intact, properly signed and dated custody seals attached to the outside of to cooler? YES NO

Were custody papers included with the cooler? YES NO

Were custody papers properly filled out (ink, signed, etc.) YES NO

Temperature of Cooler(s) (°C) (recommended 2.0-6.0 °C for chemistry) 2.6

Time: 1350

If cooler temperature is out of compliance fill out form 00070F Temp Gun ID#: D002565

Cooler Accepted by: SeF Date: 12-4-18 Time: 1350

Complete custody forms and attach all shipping documents

Log-In Phase:

Was a temperature blank included in the cooler? YES NO

What kind of packing material was used? ... Bubble Wrap Wet Ice Gel Packs Baggies Foam Block Paper Other: _____

Was sufficient ice used (if appropriate)? NA YES NO

Were all bottles sealed in individual plastic bags? YES NO

Did all bottles arrive in good condition (unbroken)? YES NO

Were all bottle labels complete and legible? YES NO

Did the number of containers listed on COC match with the number of containers received? YES NO

Did all bottle labels and tags agree with custody papers? YES NO

Were all bottles used correct for the requested analyses? YES NO

Do any of the analyses (bottles) require preservation? (attach preservation sheet, excluding VOCs)... NA YES NO

Were all VOC vials free of air bubbles? NA YES NO

Was sufficient amount of sample sent in each bottle? YES NO

Date VOC Trip Blank was made at ARI... NA

Was Sample Split by ARI: NA YES Date/Time: _____ Equipment: _____ Split by: _____

Samples Logged by: SeF Date: 12-4-18 Time: 1428

**** Notify Project Manager of discrepancies or concerns ****

Sample ID on Bottle	Sample ID on COC	Sample ID on Bottle	Sample ID on COC

Additional Notes, Discrepancies, & Resolutions:

By: _____ Date: _____

Small Air Bubbles ~ 2mm 	Peabubbles 2-4 mm 	LARGE Air Bubbles > 4 mm 	Small → "sm" (< 2 mm) Peabubbles → "pb" (2 to < 4 mm) Large → "lg" (4 to < 6 mm) Headspace → "hs" (> 6 mm)
---------------------------------------	---------------------------------	--	---

0016F
3/2/10

Cooler Receipt Form

Revision 014



AECOM
1111 Third Avenue, Suite 1600
Seattle WA, 98101

Project: Portland Harbor Pre-Remedial Design
Project Number: 60566335
Project Manager: Karen Mixon

Reported:
04-Jan-2019 12:58

PDI-WS-T02-1811
18L0044-01 (Water)

Phenols

Method: EPA 8041A
Instrument: ECD8 Analyst: VTS

Sampled: 11/30/2018 15:06

Analyzed: 12/11/2018 20:56

Sample Preparation: Preparation Method: EPA 3510C SepF
Preparation Batch: BGL0131 Sample Size: 500 mL
Prepared: 06-Dec-2018 Final Volume: 5 mL

Analyte	CAS Number	Dilution	Detection Limit	Reporting Limit	Result	Units	Notes
Pentachlorophenol	87-86-5	1	0.014	0.025	ND	ug/L	U
Surrogate: 2,4,6-Tribromophenol				10-181 %	84.5	%	
Surrogate: 2,4,6-Tribromophenol [2C]				10-181 %	73.8	%	



AECOM
1111 Third Avenue, Suite 1600
Seattle WA, 98101

Project: Portland Harbor Pre-Remedial Design
Project Number: 60566335
Project Manager: Karen Mixon

Reported:
04-Jan-2019 12:58

PDI-WS-T06-1811
18L0044-02 (Water)

Phenols

Method: EPA 8041A
Instrument: ECD8 Analyst: yz

Sampled: 11/30/2018 16:26

Analyzed: 12/11/2018 21:13

Sample Preparation: Preparation Method: EPA 3510C SepF
Preparation Batch: BGL0131 Sample Size: 500 mL
Prepared: 06-Dec-2018 Final Volume: 5 mL

Analyte	CAS Number	Dilution	Detection Limit	Reporting Limit	Result	Units	Notes
Pentachlorophenol	87-86-5	1	0.014	0.025	ND	ug/L	U
Surrogate: 2,4,6-Tribromophenol				10-181 %	85.9	%	
Surrogate: 2,4,6-Tribromophenol [2C]				10-181 %	72.1	%	



AECOM
1111 Third Avenue, Suite 1600
Seattle WA, 98101

Project: Portland Harbor Pre-Remedial Design
Project Number: 60566335
Project Manager: Karen Mixon

Reported:
04-Jan-2019 12:58

PDI-WS-T04-1812
18L0044-03 (Water)

Phenols

Method: EPA 8041A
Instrument: ECD8 Analyst: yz

Sampled: 12/01/2018 13:10

Analyzed: 12/11/2018 21:31

Sample Preparation: Preparation Method: EPA 3510C SepF
Preparation Batch: BGL0131 Sample Size: 500 mL
Prepared: 06-Dec-2018 Final Volume: 5 mL

Analyte	CAS Number	Dilution	Detection Limit	Reporting Limit	Result	Units	Notes
Pentachlorophenol	87-86-5	1	0.014	0.025	ND	ug/L	U
Surrogate: 2,4,6-Tribromophenol				10-181 %	85.2	%	
Surrogate: 2,4,6-Tribromophenol [2C]				10-181 %	71.4	%	



AECOM
1111 Third Avenue, Suite 1600
Seattle WA, 98101

Project: Portland Harbor Pre-Remedial Design
Project Number: 60566335
Project Manager: Karen Mixon

Reported:
04-Jan-2019 12:58

Phenols - Quality Control

Batch BGL0131 - EPA 3510C SepF

Instrument: ECD8 Analyst: yz

QC Sample/Analyte	Result	Detection Limit	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Blank (BGL0131-BLK1)					Prepared: 06-Dec-2018 Analyzed: 11-Dec-2018 19:44						
Pentachlorophenol	ND	0.014	0.025	ug/L							U
Surrogate: 2,4,6-Tribromophenol	0.214			ug/L	0.250		85.5	10-181			
Surrogate: 2,4,6-Tribromophenol [2C]	0.180			ug/L	0.250		72.0	10-181			
LCS (BGL0131-BS1)					Prepared: 06-Dec-2018 Analyzed: 11-Dec-2018 20:02						
Pentachlorophenol	0.162	0.014	0.025	ug/L	0.250		65.0	36-159			
Surrogate: 2,4,6-Tribromophenol	0.226			ug/L	0.250		90.3	10-181			
Surrogate: 2,4,6-Tribromophenol [2C]	0.187			ug/L	0.250		74.6	10-181			
Matrix Spike (BGL0131-MS1)					Source: 18L0044-01 Prepared: 06-Dec-2018 Analyzed: 11-Dec-2018 20:20						
Pentachlorophenol	0.220	0.014	0.025	ug/L	0.250	ND	88.0	36-159			
Surrogate: 2,4,6-Tribromophenol	0.233			ug/L	0.250		93.4	10-181			
Surrogate: 2,4,6-Tribromophenol [2C]	0.194			ug/L	0.250		77.7	10-181			
Recovery limits for target analytes in MS/MSD QC samples are advisory only.											
Matrix Spike Dup (BGL0131-MSD1)					Source: 18L0044-01 Prepared: 06-Dec-2018 Analyzed: 11-Dec-2018 20:38						
Pentachlorophenol	0.193	0.014	0.025	ug/L	0.250	ND	77.2	36-159	12.80	30	
Surrogate: 2,4,6-Tribromophenol	0.215			ug/L	0.250		86.1	10-181			
Surrogate: 2,4,6-Tribromophenol [2C]	0.184			ug/L	0.250		73.7	10-181			

Recovery limits for target analytes in MS/MSD QC samples are advisory only.



AECOM
1111 Third Avenue, Suite 1600
Seattle WA, 98101

Project: Portland Harbor Pre-Remedial Design
Project Number: 60566335
Project Manager: Karen Mixon

Reported:
04-Jan-2019 12:58

Certified Analyses included in this Report

Analyte

Certifications

Code	Description	Number	Expires
ADEC	Alaska Dept of Environmental Conservation	17-015	02/07/2019
CALAP	California Department of Public Health CAELAP	2748	06/30/2019
DoD-ELAP	DoD-Environmental Laboratory Accreditation Program	66169	01/01/2021
NELAP	ORELAP - Oregon Laboratory Accreditation Program	WA100006-011	05/12/2019
WADOE	WA Dept of Ecology	C558	06/30/2019
WA-DW	Ecology - Drinking Water	C558	06/30/2019



AECOM
1111 Third Avenue, Suite 1600
Seattle WA, 98101

Project: Portland Harbor Pre-Remedial Design
Project Number: 60566335
Project Manager: Karen Mixon

Reported:
04-Jan-2019 12:58

Notes and Definitions

J	Estimated concentration value detected below the reporting limit.
U	This analyte is not detected above the reporting limit (RL) or if noted, not detected above the limit of detection (LOD).
DET	Analyte DETECTED
ND	Analyte NOT DETECTED at or above the reporting limit
NR	Not Reported
dry	Sample results reported on a dry weight basis
RPD	Relative Percent Difference
[2C]	Indicates this result was quantified on the second column on a dual column analysis.