

Data Validation Report

Project:	Portland Harbor Pre-Remedial Design Investigation and Baseline Sampling	
Laboratory:	Test America, Knoxville, Tennessee	
Service Request:	580-78853-3	
Analyses/Method:	Chlorinated Biphenyls by HRGC/HRMS / E1668A	
Validation Level:	Stage 2A	
AECOM Project Number:	60566335.2.12	
Prepared by:	Ann Marie Kropovitch/AECOM	Completed on: 09/21/2018
Reviewed by:	George Kisluk/AECOM	File Name: 580-78853-3 DVR

SUMMARY

The samples listed below were collected by AECOM in Portland Harbor in Portland, OR on July 13, 2018.

Sample ID	Matrix/Sample Type
PDI-RB-VV-180713	Water Quality

Data validation activities were conducted with reference to:

- *EPA Method 1668A: Chlorinated Biphenyl Congeners in Water, Soil, Sediment, Biosolids, and Tissue by HRGC/HRMS (USEPA, August 2003),*
- *USEPA Contract Laboratory Program National Functional Guidelines for High Resolution Superfund Methods Data Review (April 2016),*
- *Quality Assurance Project Plan, Portland Harbor Pre-Remedial Design Investigation and Baseline Sampling, Portland Harbor Superfund Site (March 2018), and the*
- Laboratory quality control (QC) limits.

The National Functional Guidelines were modified to accommodate the non-CLP methodologies. In the absence of method-specific information, laboratory QC limits, project-specific requirements and/or AECOM professional judgment were used as appropriate.

REVIEW ELEMENTS

The data were evaluated based on the following parameters (where applicable to the method):

- ✓ Data completeness [chain-of-custody (COC)/sample integrity]
- ✓ Holding times and sample preservation
- ✗ Laboratory blanks/equipment blanks
- NA Matrix spike (MS) and/or matrix spike duplicate (MSD) results
- ✓ Ongoing precision and recovery results
- NA Field duplicate results
- ✓ Labeled compounds and labeled clean-up standard recoveries

X Sample results/reporting issues

The symbol (✓) indicates that no validation qualifiers were applied based on this parameter. An NA indicates that the parameter was not included as part of this data set or was not applicable to this validation and therefore not reviewed. The symbol (X) indicates that a QC nonconformance resulted in the qualification of data. Any QC nonconformance that resulted in the qualification of data is discussed below. In addition, nonconformances or other issues that were noted during validation, but did not result in qualification of data, may be discussed for informational purposes only.

The data appear valid as qualified and may be used for decision making purposes. Select data points were qualified as estimated or negated due to nonconformances of certain QC criteria (see discussion below). Qualified sample results are presented in Table 1.

RESULTS

Data Completeness (COC)/Sample Integrity

The data package was reviewed and found to meet acceptance criteria for completeness:

- The COCs were reviewed for completeness of information relevant to the samples and requested analyses, and for signatures indicating transfer of sample custody.
- The laboratory sample login sheet(s) were reviewed for issues potentially affecting sample integrity, including the condition of sample containers upon receipt at the laboratory.
- Completeness of analyses was verified by comparing the reported results to the COC requests.

The sample identification of PDI-RB-VV-20180713 listed on the COC was changed to PDI-RB-VV-180713 at AECOM's request.

Holding Times and Sample Preservation

Sample preservation and preparation/analysis holding times were reviewed for conformance with method criteria. All method QC acceptance criteria were met.

Laboratory Blanks

Method and equipment rinsate blank results are evaluated as to whether there are contaminants detected above the estimated detection limit (EDL).

Target compounds were detected in the laboratory method blanks and the equipment blank associated with the samples in this data set. The equipment blank contamination, after laboratory method blank actions were applied, is summarized below for informational purposes only. Equipment blank contamination was not used to qualify field samples.

Blank ID	Compound	Result	EDL	Units
PDI-RB-VV-180713	PCB-1	0.0034	0.00026	ng/L
	PCB-110	0.0060	0.000071	ng/L
	PCB-115	0.0060	0.000071	ng/L
	PCB-132	0.0015	0.000093	ng/L
	PCB-174	0.0015	0.00028	ng/L

	PCB-18	0.015	0.0026	ng/L
	PCB-183	0.0023	0.00026	ng/L
	PCB-185	0.0023	0.00026	ng/L
	PCB-20	0.0064	0.0018	ng/L
	PCB-28	0.0064	0.0018	ng/L
	PCB-30	0.015	0.0026	ng/L
	PCB-52	0.0069	0.00042	ng/L
	PCB-56	0.0021	0.00031	ng/L
	PCB-60	0.0014	0.00031	ng/L
	PCB-64	0.0031	0.00028	ng/L
	PCB-83	0.0018	0.00010	ng/L
	PCB-95	0.0014	0.00011	ng/L
	PCB-99	0.0018	0.00010	ng/L

The NFG guidance stipulates that a conservative approach should be taken with regards to qualification of PCB congeners due to the toxicity of these compounds and the reporting of false negative results should be avoided. Therefore, in order to avoid the reporting of false negative results professional judgment was used to qualify the data in the following manner. As allowed in the NFG, a blank action limit (BAL) was determined as 5 times the method blank result:

- When the sample results were < the method blank result, the sample result was qualified as nondetect (U) at the sample result.
- When the sample result was \geq the method blank result and \leq the BAL, the sample result was qualified as estimated and potentially biased high (J+).
- When the sample result was > the BAL, sample result was not qualified.

Qualified sample results are summarized in Table 1.

MS/MSD Results

A MS/MSD was not submitted with this sample delivery group (SDG).

Ongoing Precision and Recovery

The OPR %Rs and/or RPDs were reviewed for conformance with the method QC acceptance criteria. All method QC acceptance criteria were met.

Field Duplicate Results

A field duplicate was not submitted with this SDG.

Labeled Compounds and Labeled Clean-up Standard Recoveries

The labeled compounds and labeled clean-up standard %Rs were reviewed for conformance with the QC acceptance criteria. No QC outliers were noted during the sample review.

Sample Results/Reporting Issues

All sample results detected at concentrations less than the lowest calibration standard (or PQL) but greater than the EDL are qualified by the laboratory as estimated (J). This "J" qualifier is retained during data validation.

The laboratory qualified the sample results with a "q" to indicate that the ion abundance ratio was outside of the QC acceptance limits; the result should be considered as an Estimated Maximum Possible Concentration (EMPC). These results were qualified as estimated and tentatively identified (JN). Qualified sample results are summarized in Table 1.

It should be noted that the "JN" qualifier was retained rather than replacement with the conventional overall "J", "J+", and "J-" qualifiers in instances where sample results were qualified for multiple quality control nonconformances.

Percent Solids Content

The percent solids data were reviewed since the amount of moisture in a solid sample may have an impact on data representativeness. Due to the extremely low solubility of PCB congeners in water, these analytes should be contained in the solid phase. Consequently, the NFG guidance does not stipulate a percent solids criterion. If applicable, EPA Regional guidance is used when assessing percent solids content. In the absence of EPA Regional guidance, AECOM uses 30% solids (from the NFG semivolatile guidance) as a benchmark to evaluate the percent solids content and professional judgment is used to determine the necessity to qualify data. Data were not qualified on the basis of percent solids content.

QUALIFICATION ACTIONS

Sample results qualified as a result of validation actions are summarized in Table 1. All actions are described above.

ATTACHMENTS

Attachment A: Qualifier Codes and Explanations

Attachment B: Reason Codes and Explanations

Table 1 - Data Validation Summary of Qualified Data

Sample ID	Matrix	Compound	Result	RDL	EDL	Units	Validation Qualifiers	Validation Reason
PDI-RB-VV-180713	WQ	PCB-101		0.0041	0.000086	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-11		0.024	0.015	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-110	0.0060	0.000071	0.000071	ng/L	JN	bl,k
PDI-RB-VV-180713	WQ	PCB-113		0.0041	0.000086	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-115	0.0060	0.000071	0.000071	ng/L	JN	bl,k
PDI-RB-VV-180713	WQ	PCB-129		0.0042	0.000073	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-132	0.0015	0.000093	0.000093	ng/L	JN	k
PDI-RB-VV-180713	WQ	PCB-138		0.0042	0.000073	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-147		0.0031	0.000091	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-149		0.0031	0.000091	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-153		0.0038	0.000063	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-160		0.0042	0.000073	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-163		0.0042	0.000073	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-168		0.0038	0.000063	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-174	0.0015	0.00028	0.00028	ng/L	JN	k
PDI-RB-VV-180713	WQ	PCB-18	0.015	0.0026	0.0026	ng/L	JN	k
PDI-RB-VV-180713	WQ	PCB-183	0.0023	0.00026	0.00026	ng/L	JN	k
PDI-RB-VV-180713	WQ	PCB-185	0.0023	0.00026	0.00026	ng/L	JN	k
PDI-RB-VV-180713	WQ	PCB-194		0.0013	0.00022	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-20	0.0064	0.0018	0.0018	ng/L	JN	bl,k
PDI-RB-VV-180713	WQ	PCB-28	0.0064	0.0018	0.0018	ng/L	JN	bl,k
PDI-RB-VV-180713	WQ	PCB-3		0.0040	0.00033	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-30	0.015	0.0026	0.0026	ng/L	JN	k
PDI-RB-VV-180713	WQ	PCB-44		0.0037	0.00037	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-47		0.0037	0.00037	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-52	0.0069	0.00042	0.00042	ng/L	JN	bl,k
PDI-RB-VV-180713	WQ	PCB-56	0.0021	0.00031	0.00031	ng/L	JN	k
PDI-RB-VV-180713	WQ	PCB-60	0.0014	0.00031	0.00031	ng/L	JN	k
PDI-RB-VV-180713	WQ	PCB-61		0.0043	0.00029	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-65		0.0037	0.00037	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-66		0.0012	0.00029	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-70		0.0043	0.00029	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-74		0.0043	0.00029	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-76		0.0043	0.00029	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-90		0.0041	0.000086	ng/L	U	bl
PDI-RB-VV-180713	WQ	PCB-95	0.0014	0.00011	0.00011	ng/L	JN	k

Attachment A**Qualifier Codes and Explanations**

Qualifier	Explanation
J	The analyte was positively identified; the associated numerical value is the approximate concentration of the analyte in the sample.
J-	The analyte was positively identified; the associated numerical value is the approximate concentration of the analyte in the sample with a potential low bias.
J+	The analyte was positively identified; the associated numerical value is the approximate concentration of the analyte in the sample with a potential high bias.
JN	The analyte was tentatively identified; the associated numerical value is the approximate concentration of the analyte in the sample.
UJ	The analyte was not detected above the reported sample quantitation limit. However, the reported quantitation limit is approximate and may or may not represent the actual limit of quantitation necessary to accurately and precisely measure the analyte in the sample.
U	The analyte was analyzed for, but was not detected above the reported sample quantitation limit.
R	The sample results are rejected due to serious deficiencies in the ability to analyze the sample and meet quality control criteria. The presence or absence of the analyte cannot be verified.

Attachment B

Reason Codes and Explanations

Reason Code	Explanation
be	Equipment blank contamination
bf	Field blank contamination
bl	Laboratory blank contamination
c	Calibration issue
cl	Clean-up standard recovery
d	Reporting limit raised due to chromatographic interference
fd	Field duplicate RPDs
h	Holding times
i	Internal standard areas
k	Estimated Maximum Possible Concentration (EMPC)
l	LCS or OPR recoveries
lc	Labeled compound recovery
ld	Laboratory duplicate RPDs
lp	Laboratory control sample/laboratory control sample duplicate RPDs
m	Matrix spike recovery
ma	Multiple analyses. Sample analyzed more than once, a value from another analysis should be used.
md	Matrix spike/matrix spike duplicate RPDs
nb	Negative laboratory blank contamination
p	Chemical preservation issue
r	Dual column RPD
q	Quantitation issue
s	Surrogate recovery
su	Ion suppression
t	Temperature preservation issue
x	Percent solids
y	Serial dilution results
z	ICS results