



1435 Norjohn Court, Unit 1, Burlington, ON, Canada L7L 0E6

SVOC DATA PACKAGE

Client Project Information

Project ID: 60566335

Project Description: Portland Harbor Pre- Remedial Design Investigation & Baseline Sampling

Contact: Amy Dahl

ALSE Project Information

Project ID: AECOM100

Contact: Whitney Davis

Submission ID(s): L2125039

Final Package Review by:

A handwritten signature in black ink that reads "R.A. Myer".

Date Reviewed: 25-Jul-18



1435 Norjohn Court, Unit 1, Burlington, ON, Canada L7L 0E6

SVOC DATA PACKAGE

SECTION 1: PROJECT NARRATIVE

ALSE Project Information

Project ID: AECOM100

Contact: Whitney Davis

Submission ID(s): L2125039

Client Project Information

Project ID: 60566335

Project Description: Portland Harbor Pre- Remedial Design Investigation & Baseline Sampling

Contact: Amy Dahl

Analytical Method: Chlorinated Pesticides by EPA 1699 (modified)

ALS Sample ID	Client Sample Descriptions	Matrix	Date Sampled	Date Received	Date Extracted	Date Analyzed
L2125039-1	NIST 1944	Sediment	n/a	n/a	2018-Jul-06	2018-Jul-20
L2125039-2	NIST 1944 Duplicate	Sediment	n/a	n/a	2018-Jul-06	2018-Jul-20
WG2815918-1	Laboratory Method Blank	QC	n/a	n/a	2018-Jul-06	2018-Jul-20
WG2815918-2	Laboratory Control Sample	QC	n/a	n/a	2018-Jul-06	2018-Jul-20

Comments and Notes:

a) Sample Integrity:

The NIST reference material has been refrigerated since receipt. It has not exceeded its expiry date of 31 March 2027.

b) Extraction and Cleanup:

The NIST material was extracted overnight by soxhlet extraction technique using dichloromethane as the extracting solvent. The extract was cleaned by silica gel column chromatography prior to instrumental analysis.

c) Instrumental Analysis:

Extraction standard recovery limits are listed herein have been derived from those recommended in EPA method 1699. The lower acceptance limit of 47% for 4,4'-DDE-13C12 has been found not to represent observed recoveries and this lower limit will soon be revised based upon historical data. On this data set, the 1st of the duplicates gave a recovery slightly below the listed acceptance limit. Due to isotope dilution technique, this apparent exceedence is not expected to have a significant impact on data quality. Moisture content was assumed to be 1.3% as esimated in NIST 1944 CoA.

No other criteria failures or exceedances.

I certify that this data package is in compliance with the terms and condition of the contract , both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this data package (hardcopy and/or electronic version) has been authorized by the Laboratory Manager or his designee, as verified by the following signature.

Reviewer and Title

2018-Jul-25

Date

SVOC DATA PACKAGE

SECTION 2: DATA SUMMARY REPORT



1435 Norjohn Court, Unit 1, Burlington, ON, Canada L7L 0E6
Phone: 905-331-3111, FAX: 905-331-4567

Certificate of Analysis

ALS Project Contact: Whitney Davis
ALS Project ID: AECOM100
ALS WO#: L2125039
Date of Report: 25-Jul-18

Client Name: AECOM
Client Address: 1111 Third Avenue, Suite 1600
Seattle, WA 98101
United States
Client Contact: Amy Dahl
Client Project ID: 60566335

COMMENTS: Chlorinated Pesticides by EPA 1699 (modified)

Samples were analyzed by isotope dilution technique.

A high bias to results relative to the NIST reference values is expected due to the isotope dilution technique employed.

Isotope dilution technique - where the samples are spiked with C-13 labelled targets prior to extraction - provides an inherent correction for losses during extraction and clean-up, a correction that is absent by other standard methods for environmental analysis.

Certified by:

A handwritten signature in black ink, appearing to read "R. A. McLeod", is written over a horizontal line.

Ron McLeod, PhD
Director, Air Toxics and Special Chemistries, Life Sciences

Results in this certificate relate only to the samples as submitted to the laboratory.

This report shall not be reproduced, except in full, without the written permission of ALS Canada Ltd.

ALS Life sciences

Sample Analysis summary Report

Sample Name	NIST 1944	% of Ref Value	NIST 1944 Duplicate	% of Ref Value	Ref Values
ALS Sample ID	L2125039-1		L2125039-2		
Sample Size	2.01		2.02		
Sample size units	g		g		
Percent Solids	98.7%		98.7%		
Sample Matrix	Sediment		Sediment		
Sampling Date	n/a		n/a		
Extraction Date	6-Jul-18		6-Jul-18		
Target Analytes	ng/g	%	ng/g	%	ng/g
2,4'-DDE	23.0	121	23.6	124	19.0
4,4'-DDE	83.4	97	80.9	94	86
2,4'-DDD	77.8	205	57.8	152	38
4,4'-DDD	139	129	117	108	108
4,4'-DDT	189	111	199	117	170
Extraction Standards	% Rec		% Rec		
4,4'-DDE, 13C12-	44		62		
4,4'-DDD, 13C12-	37		44		
4,4'-DDT, 13C12-	22		26		

ALS Life sciences

Quality Control Summary Report

Sample Name	Method Blank	Laboratory Control Sample
ALS Sample ID	WG2815918-1	WG2815918-2
Sample Size	2.00	1.00
Sample size units	g	n/a
Percent Solids	n/a	n/a
Sample Matrix	QC	QC
Sampling Date	n/a	n/a
Extraction Date	6-Jul-18	6-Jul-18
Target Analytes	ng/g	% Rec
2,4'-DDE	<0.0049	101
4,4'-DDE	0.0350	99
2,4'-DDD	<0.015	107
4,4'-DDD	<0.010	96
4,4'-DDT	<0.013	96
Extraction Standards	% Rec	% Rec
4,4'-DDE, 13C12-	84	82
4,4'-DDD, 13C12-	82	89
4,4'-DDT, 13C12-	84	88

ALS Life sciences

Sample Analysis Report

Sample Name NIST 1944
 ALS Sample ID L2125039-1
 Analysis Method EPA 1699 (mod)
 Analysis Type Sample
 Sample Matrix Sediment

Sampling Date n/a
 Extraction Date 6-Jul-18
 Sample Size 2.01 g
 Percent Solids 98.7%
 Split Ratio 1

Approved:
R. Bakthiari
 --e-signature--
 20-Jul-2018

Run Information Run 1
 Filename 6-180719B44
 Run Date 20-Jul-18 14:55
 Final Volume 1020 uL
 Dilution Factor 1
 Analysis Units ng/g
 Instrument - Column HRMS-6 HP5MSUSR160544H

Target Analytes	Ret. Time	Conc. ng/g	EDL ng/g	Flags	EMPC ng/g	LQL
2,4'-DDE	20.98	23.0	0.051			1.0
4,4'-DDE	21.91	83.4	0.033			1.0
2,4'-DDD	22.15	77.8	0.052			1.0
4,4'-DDD	23.11	139	0.038			1.0
4,4'-DDT	24.17	189	0.054			1.0
Extraction Standards ng						
4,4'-DDE, 13C12-	125	21.90	44	47-160		
4,4'-DDD, 13C12-	125	23.11	37	5-150		
4,4'-DDT, 13C12-	125	24.17	22	5-120		

EDL Indicates the Estimated Detection Limit, based on the measured background noise for this target in this sample.
 LQL Lower Quantification Limit, based on the lowest calibration level corrected for sample size, splits and dilutions.
 M Indicates that a peak has been manually integrated.
 U Indicates that this compound was not detected above the EDL.

 J indicates that a target analyte was detected below the calibrated range.
 R Indicates that the ion abundance ratio for this compound did not meet the acceptance criterion.
 B Indicates that this target was detected in the blank at greater than 10% of the sample concentration.

 EMPC Estimated Maximum Possible Concentration – elevated detection limit due to interference or positive id criterion failure

ALS Life sciences

Sample Analysis Report

Sample Name NIST 1944 (Duplicate)
 ALS Sample ID L2125039-2
 Analysis Method EPA 1699 (mod)
 Analysis Type Sample
 Sample Matrix Sediment

Sampling Date n/a
 Extraction Date 6-Jul-18
 Sample Size 2.02 g
 Percent Solids 98.7%
 Split Ratio 1

Approved:
R. Bakthiari
 --e-signature--
 20-Jul-2018

Run Information **Run 1**
 Filename 6-180719B45
 Run Date 20-Jul-18 15:29
 Final Volume 1020 uL
 Dilution Factor 1
 Analysis Units ng/g
 Instrument - Column HRMS-6 HP5MSUSR160544H

Target Analytes	Ret. Time	Conc. ng/g	EDL ng/g	Flags	EMPC ng/g	LQL
2,4'-DDE	20.99	23.6	0.011		1.0	
4,4'-DDE	21.91	80.9	0.035		1.0	
2,4'-DDD	22.16	57.8	0.055		1.0	
4,4'-DDD	23.12	117	0.070		1.0	
4,4'-DDT	24.18	199	0.076		1.0	
Extraction Standards ng						
4,4'-DDE, 13C12-	125	21.91	62	47-160		
4,4'-DDD, 13C12-	125	23.11	44	5-150		
4,4'-DDT, 13C12-	125	24.17	26	5-120		

EDL Indicates the Estimated Detection Limit, based on the measured background noise for this target in this sample.
 LQL Lower Quantification Limit, based on the lowest calibration level corrected for sample size, splits and dilutions.
 M Indicates that a peak has been manually integrated.
 U Indicates that this compound was not detected above the EDL.

 J indicates that a target analyte was detected below the calibrated range.

 B Indicates that this target was detected in the blank at greater than 10% of the sample concentration.

 EMPC Estimated Maximum Possible Concentration – elevated detection limit due to interference or positive id criterion failure

SVOC DATA PACKAGE

SECTION 3: METHOD SUMMARY

OC Pesticide METHOD SUMMARY
Method EPA 1699

Introduction:

This summary is to provide ALSE Burlington OC pesticide method details in order to provide persons reviewing or validating this data package sufficient information to re-construct the sample calculation, data verification and review. It incorporates the analysis of organochlorine pesticides via EPA method 1699. Deviations from this reference method are documented in ALS Standard Operating Procedures (available upon request) and in this Method Summary.

Any deviations to what is listed herein or in the ALS Standard Operating Procedures would be listed in the project narrative.

To avoid the confusion and conflicting nomenclature within the performance based methods, we have defined the labeled standards in terms relating to the time of addition to the sample or extract. Therefore;

- Laboratory Surrogate (when provided/requested by the client) are added prior to sample extraction
- The Field or Sampling Standards (where used) are added prior to field sampling
- The Extraction Standards are added prior to extraction
- The GPC Recovery Standard is added (when used) prior to Gel Permeation Chromatographic cleanup
- The Clean-up Standards (where used) are added prior to extract clean-up
- The Injection Standards are added prior to extract injection.

Additional method information, such as Instrumental Descriptors, is documented in ALS Standard Operating Procedures and available upon request.

Calibration Standard Levels:

Seven levels of standard are available for calibration as listed in Table 1. These targets give a wide range of responses on the analytical instruments, thus it is expected that for any given target, either the lowest standard level(s) or the highest standard level(s) may be excluded due to poor response, poor linearity, or detector saturation. With seven levels of standard, it is expected that at least 5 points can be used for calibration for each target.

Table 1: Calibration Standards (conc in ng/mL)

	CS1	CS2	CS3	CS4	CS5	CS6	CS7
Natives							
Hexachlorobutadiene	2	7.5	20	50	150	400	1200
1,2,4,5-Tetrachlorobenzene	2	7.5	20	50	150	400	1200
1,2,3,4-Tetrachlorobenzene	2	7.5	20	50	150	400	1200
Pentachlorobenzene	2	7.5	20	50	150	400	1200
Hexachlorobenzene	2	7.5	20	50	150	400	1200
3,4,5,6-Tetrachloroveratrole	2	7.5	20	50	150	400	1200
Pentachloroanisole	2	7.5	20	50	150	400	1200
alpha-BHC	2	7.5	20	50	150	400	1200
beta-BHC	2	7.5	20	50	150	400	1200
gamma-BHC	2	7.5	20	50	150	400	1200
delta-BHC	2	7.5	20	50	150	400	1200
Pentachloronitrobenzene	2	7.5	20	50	150	400	1200
Heptachlor	2	7.5	20	50	150	400	1200
Aldrin	2	7.5	20	50	150	400	1200
4,4'-DDNU	2	7.5	20	50	150	400	1200
Dacthal	2	7.5	20	50	150	400	1200
Chlorpyrifos	10	37.5	100	250	750	2000	6000
Octachlorostyrene	2	7.5	20	50	150	400	1200
Heptachlor Epoxide B	2	7.5	20	50	150	400	1200
Heptachlor Epoxide A	2	7.5	20	50	150	400	1200
Oxychlorane	2	7.5	20	50	150	400	1200
4,4'-DDMU	2	7.5	20	50	150	400	1200
trans-Chlordane	2	7.5	20	50	150	400	1200
cis-Chlordane	2	7.5	20	50	150	400	1200
trans-Nonachlor	2	7.5	20	50	150	400	1200
Dieldrin	2	7.5	20	50	150	400	1200
Endrin	2	7.5	20	50	150	400	1200
cis-Nonachlor	2	7.5	20	50	150	400	1200
Endosulfan I	2	7.5	20	50	150	400	1200
Endosulfan II	2	7.5	20	50	150	400	1200
Endosulfan Sulfate	2	7.5	20	50	150	400	1200
24'-DDE	2	7.5	20	50	150	400	1200
44'-DDE	2	7.5	20	50	150	400	1200
24'-DDD	2	7.5	20	50	150	400	1200
44'-DDD	2	7.5	20	50	150	400	1200
24'-DDT	2	7.5	20	50	150	400	1200
44'-DDT	2	7.5	20	50	150	400	1200
Endrin Aldehyde	2	7.5	20	50	150	400	1200
Endrin Ketone	2	7.5	20	50	150	400	1200
Methoxychlor	2	7.5	20	50	150	400	1200
Dicofol	20	75	200	500	1500	4000	12000
Mirex	2	7.5	20	50	150	400	1200
Parlar-26	2	7.5	20	50	150	400	1200
Parlar-50	2	7.5	20	50	150	400	1200
Parlar-62	2	7.5	20	50	150	400	1200

Laboratory Surrogate	1,3-Dibromobenzene	20	20	20	20	20	20	20
	Endrin Ketone	2	7.5	20	50	150	400	1200
Field Surrogate	1,3,5-Tribromobenzene	20	20	20	20	20	20	20
	1,2,4,5-Tetrabromobenzene	20	20	20	20	20	20	20
	delta-BHC	2	7.5	20	50	150	400	1200
GPC Recovery Standard	13C12-PCB-133	100	100	100	100	100	100	100
Extraction Standard	13C6-Pentachlorobenzene	250	250	250	250	250	250	250
	13C6-Hexachlorobenzene	250	250	250	250	250	250	250
	13C6-alpha-BHC	250	250	250	250	250	250	250
	d6-gamma-BHC	250	250	250	250	250	250	250
	13C10-Heptachlor	250	250	250	250	250	250	250
	13C10-Oxychlorane	250	250	250	250	250	250	250
	13C10-trans-Nonachlor	250	250	250	250	250	250	250
	13C12-Dieldrin	250	250	250	250	250	250	250
	13C12-Endrin	250	250	250	250	250	250	250
	13C9-Endosulfan-II	250	250	250	250	250	250	250
	13C12-44'-DDE	250	250	250	250	250	250	250
	13C12-44'-DDD	250	250	250	250	250	250	250
	13C12-44'-DDT	250	250	250	250	250	250	250
	d6-Methoxychlor	250	250	250	250	250	250	250
13C10-Mirex	250	250	250	250	250	250	250	
Injection Standard	13C12-PCB-9	100	100	100	100	100	100	100
	13C12-PCB-52	100	100	100	100	100	100	100
	13C12-PCB-101	100	100	100	100	100	100	100

Calibration and Quality Control Limits

The calibration and QC Sample control limits are presented in Table 2 below. For the lowest standard used for initial calibration, and for each calibration verification CS3, the signal to noise ratio for each ion for both labelled and non-labelled analytes must be greater than or equal to 10:1

		Calibration		Samples and QC Samples	
		Initial Cal. %RSD	Cal. Ver. %Exp	LCS % Rec	Samples % Rec
Natives	Hexachlorobutadiene	35	70-130	5-200	
	1,2,4,5-Tetrachlorobenzene	35	70-130	5-200	
	1,2,3,4-Tetrachlorobenzene	35	70-130	5-200	
	Pentachlorobenzene	20	70-130	5-200	
	Hexachlorobenzene	20	75-125	10-150	
	3,4,5,6-Tetrachloroveratrole	35	70-130	20-200	
	Pentachloroanisole	35	70-130	20-200	
	alpha-BHC	20	75-125	50-120	
	beta-BHC	35	75-125	50-120	
	gamma-BHC	20	75-125	50-120	
	delta-BHC	35	75-125	50-120	
	Pentachloronitrobenzene	35	70-130	20-200	
	Heptachlor	20	75-125	50-120	
	Aldrin	35	75-125	50-120	
	4,4'-DDNU	35	75-125	20-160	
	Dacthal	35	50-150	20-200	
	Chlorpyrifos	35	75-125	19-163	
	Octachlorostyrene	35	70-130	50-175	
	Heptachlor Epoxide B	35	70-130	20-200	
	Heptachlor Epoxide A	35	75-125	50-120	
	Oxychlordane	20	75-125	50-120	
	4,4'-DDMU	35	75-125	20-160	
	trans-Chlordane	35	75-125	50-120	
	cis-Chlordane	35	75-125	50-120	
	trans-Nonachlor	20	75-125	50-120	
	Dieldrin	20	75-125	50-120	
	Endrin	20	75-125	50-120	
	cis-Nonachlor	35	75-125	50-120	
	Endosulfan I	35	75-125	50-120	
	Endosulfan II	20	75-125	5-200	
	Endosulfan Sulfate	35	75-125	50-200	
	24'-DDE	35	75-125	24-123	
	44'-DDE	20	75-125	50-120	
	24'-DDD	35	75-125	50-120	
	44'-DDD	20	75-125	42-120	
	24'-DDT	35	75-125	50-120	
	44'-DDT	20	75-125	50-120	
	Endrin Aldehyde	35	70-130	20-200	
	Endrin Ketone	35	75-125	50-134	
	Methoxychlor	20	75-125	50-120	
	Dicofol	35	50-150	20-200	
	Mirex	20	75-125	50-120	
	Parlar-26	35	70-130	20-200	
	Parlar-50	35	70-130	20-200	
	Parlar-62	35	70-130	20-200	

Laboratory Surrogate	1,3-Dibromobenzene	35	50-150	50-150	40-120
	Endrin Ketone	35	50-150	50-150	40-150
Field Surrogate	1,3,5-Tribromobenzene	35	50-150	50-150	60-120
	1,2,4,5-Tetrabromobenzene	35	50-150	50-150	60-120
	delta-BHC	35	50-150	50-150	60-120
GPC Recovery Standard	13C12-PCB-133	35	50-150	50-150	50-120
Extraction Standard	13C6-Pentachlorobenzene	35	70-130	5-120	5-120
	13C6-Hexachlorobenzene	35	70-130	5-120	5-120
	13C6-alpha-BHC	35	70-130	13-138	16-129
	d6-gamma-BHC	35	70-130	5-124	11-120
	13C10-Heptachlor	35	70-130	5-128	5-120
	13C10-Oxychlorane	35	70-130	5-144	23-135
	13C10-trans-Nonachlor	35	70-130	17-154	36-139
	13C12-Dieldrin	35	70-130	19-161	40-151
	13C12-Endrin	35	70-130	20-157	35-155
	13C9-Endosulfan-II	35	70-130	5-144	15-148
	13C12-44'-DDE	35	70-130	26-169	47-160
	13C12-44'-DDD	35	70-130	13-200	5-150
	13C12-44'-DDT	35	70-130	13-200	5-120
	d6-Methoxychlor	35	70-130	8-200	5-120
	13C10-Mirex	35	70-130	5-138	5-120

Additional Continuing Calibration Details:

After initial calibration is established, a CS4 standard is injected as a Continuing Calibration Verification (CCV) at the beginning of every 12 hour shift in which samples are analyzed. If the following performance criteria are met, analysis of samples may proceed:

- Ion abundance ratios are within their respective theoretical limits (see Table 3)
- All targets have a s/n ratio of at least 10:1
- The RT of each analyte is within 15 seconds of that in the initial calibration
- Endin and DDT breakdown is less than 20% (see Section 5.2.4.2)
- The %Diff is within the CCV limits (see Table 2)

If these performance criteria are not met, GC maintenance is performed or the system is adjusted and a new CCV is injected, or a new initial calibration is run.

Mid-run Calibration Verification:

While the EPA 1699 does not require a post-run calibration verification standard to be run, it is recognized that responses and/or relative responses of some targets may change significantly during HRMS analysis due to sample related contamination of GC or MS components. This problem is compounded by chemical dissimilarities between some targets and their quantification reference standards in the case of internal standard quantification. Enhanced quantification and a measure of confidence in sample results obtained during an analytical shift can be attained by injecting a CS4 calibration verification (VER) standard in the middle of, and at the end of a 12-hour run, and quantifying samples against the average of bracketing calibration standards where improved results would be achieved.

a) Mid-Run VER:

If this analysis meets the performance criteria for a pre-run CCV, then all of the samples preceding the mid-run VER can be quantified vs. the initial calibration, and analysis can proceed. If the mid-run VER does not meet pre-run CCV criteria, the preceding samples can be quantified vs. bracketing calibration runs (using the pre-run CCV and mid-run VER as a two-point calibration) and analysis can proceed, provided that the following criteria are met:

- Ion abundance ratios are within their respective theoretical limits (see Table 1) or within 15% of the ratios in the pre-run CCV
- All targets have a s/n ratio of at least 10:1
- The RT of each analyte is within 15 seconds of that in the initial calibration
- Endin and DDT breakdown is less than 20%
- The %RPD of the mid-run VER vs. the pre-run CCV meets the CCV %Diff limits (See Table 2)

If the mid-run VER does not meet the above criteria either, analysis cannot continue without corrective action (samples analyzed after the mid-run VER in an automated sequence must be re-analyzed). The samples preceding the failing mid-run VER may be flagged and reported, but must be assessed for impact on data quality:

- If a failing native target is present in any of the preceding samples above the Method Detection Limit (or above the client's lower required Detection Limit, if known), that sample must be re-analyzed for that target.
- If a failing native target's Estimated Detection Limit is above the Method Detection Limit (or above the client's lower required Detection Limit, if known) due to deterioration of system performance, that sample must be re-analyzed for that target.

a) Post-Run VER:

If this analysis meets the performance criteria for a pre-run CCV, then all of the samples preceding the post-run VER can be quantified vs. the initial calibration. If the post-run VER does not meet pre-run CCV criteria, the preceding samples can be quantified vs. bracketing calibration runs (using the post-run VER and mid-run VER as a two-point calibration) provided that the following criteria are met:

- Ion abundance ratios are within their respective theoretical limits (see Table 1) or within 15% of the ratios in the mid-run CCV
- All targets have a s/n ratio of at least 10:1
- The RT of each analyte is within 15 seconds of that in the initial calibration
- Endin and DDT breakdown is less than 20%
- The %RPD of the post-run VER vs. the mid-run VER meets the CCV %Diff limits (See Table 2)

If the post-run VER does not meet the above criteria either, the samples preceding the failing post-run VER may be flagged and reported, but must be assessed for impact on data quality:

- If a failing native target is present in any of the preceding samples above the Method Detection Limit (or above the client's lower required Detection Limit, if known), that sample must be re-analyzed for that target.
- If a failing native target's Estimated Detection Limit is above the Method Detection Limit (or above the client's lower required Detection Limit, if known) due to deterioration of system performance, that sample must be re-analyzed for that target.

Reporting Limits:

Unless indicated in the otherwise, native target data is reported down to 2.5:1 signal to noise for each isomer grouping for each extract injection. This is consistent to SW846 8290 defined protocols (i.e. EDL or Estimated Detection Limit) and is commonly applied throughout the industry to any and all performance based HRMS methods.

Method Blank:

The method blank levels must be below the response to the lowest calibration standard used for initial calibration.

MS/MSD (where required):

The % relative difference between the MS and MSD spike recoveries should be less than or equal to 20%.

Instrument/Run Performance Criteria:

a) Chromatographic Performance

For the DB-5 column, 44'-DDT and 24'-DDT (or the labelled analogues) must be uniquely resolved to a valley height of less than 60% of the shorter of the two peaks.

b) DDT and Endrin Breakdown

A custom standard (HROCP-GC_BD#1) is injected to measure the breakdown of endrin and DDT during the run. This standard must be injected at the beginning and end of each 12 hour shift, and it is also recommended that it be injected along with the mid-run CCV where used. This standard contains 13C12-4,4'-DDT, 13C12-endrin, and native endrin, endrin aldehyde and endrin ketone.

- For measurement of DDT breakdown, measure the concentration for 13C12-44'-DDE, 13C12-44'-DDD and 13C12-44'-DDT (the labelled DDT is part of the standard, and the labelled DDE and DDD are breakdown products). Calculate breakdown using the following formula:

13C12-44'-DDT % Breakdown =

$$\frac{(\text{concentration of } 13\text{C}12\text{-}44'\text{-DDD} + \text{concentration of } 13\text{C}12\text{-}44'\text{-DDE}) \times 100\%}{\text{concentration of } 13\text{C}12\text{-}44'\text{-DDT}}$$

labelled DDT = part of standard; labelled DDE and DDD = breakdown products

- Additionally, measurement of endrin breakdown can be performed. For measurement of endrin breakdown, measure the concentration of endrin, endrin aldehyde, and endrin ketone (these natives are quantified by isotope dilution vs. the 13C12-endrin). Calculate breakdown using the following formula:

Endrin % Breakdown =

$$\frac{(\text{concentration of endrin aldehyde} + \text{concentration of endrin ketone}) \times 100\%}{\text{concentration of endrin}}$$

If the breakdown of endrin and/or DDT exceeds 20% in a standard, the targets are decomposing on the inlet or column, and remedial action must be taken (inlet maintenance and trimming of the analytical column) before any valid sample data can be produced. If the breakdown of DDT or endrin in a sample exceeds 20% and there is that native in the sample above the MDL, that sample will have to be reanalyzed for that target (further cleanup or dilution of that sample is recommended before reanalysis).

Breakdown exceedences can be ignored under the following circumstances:

- Where the endrin breakdown fails but DDT breakdown passes and where DDT and/or its metabolites are the only targets.
- Where the DDT breakdown fails but endrin breakdown passes and where endrin and/or its metabolites are the only targets.
- For the determination of other pesticide targets (i.e. non-DDT and non-Endrin and metabolite targets) which have a corresponding labelled extraction/internal standard of exactly the same isomer.

c) Mass Resolution:

At the beginning of and just following the end of each 12 hour run sequence, the instrument must be checked to demonstrate a resolution of 10,000 for each quantification window.

The maximum time between scans within a descriptor is 1 second.

Lock mass deviations to the average response must be less than or equal 20%.

Laboratory Duplicates:

The % relative difference between duplicates should be less than or equal to 25% but only where the response is greater than the low calibration standard.

Analyte Identification Criteria:

Ion Ratio Criteria

For all compounds, a pair of ions with a specific isotopic ratio are being monitored. To have a confirmed positive response to a native or labelled OCP, that ratio must be within the theoretical limits in Table 1, or within 15% of the observed values on the most recent CS4 analysis.

Signal to Noise Criteria

The signal to noise ratio for each quantification and confirmation ion for labelled and non-labelled analytes must be greater than or equal to 10:1 for the initial calibration CS1 and for each calibration verification CS4. For positive identification of a native target in a sample, both ions must have a s/n ratio exceeding 2.5:1.

Matched RT on Peak Maxima

The retention time (RT) of the peak maxima for each pair of quantification ions must be no more than 2 seconds (i.e. 2 scans) difference.

Expected Retention Time (RT)

The peak must be at the expected RT

- within -1/+3 seconds of the labelled standard for natives with their own ¹³C labelled standard
- within +/- 0.008 RRT units of the RRT in the most recent CS4 analysis for targets with their own ²H labelled standard
- within +/- 0.010 RRT units of the RRT in the most recent CS4 analysis for targets without their own labelled standard

As per EPA 1699 Sections 16.5-16.6, it is possible that not all of the positive ID criteria are met. If a pesticide is deemed to be present in this case by the experienced spectroscopist, the result may be flagged as "this result is unconfirmed and must not be used for permitting or regulatory compliance purposes". If the ion abundance ratio criteria are not met, the result must also include an "R" flag.

Table 3: Monitored Masses, Ion Abundance Ratios, and Quantitation/RT References

Entry	Native Standard	Quantification Method	Quantification vs. Entry #:	Quantitation Ion	Confirmation Ion	Theoretical Ion Abundance ratio	Ion Abundance Ratio Tolerance
1	Hexachlorobutadiene	rel_int	52	259.8102	261.8072	1.25	0.25
2	1,2,4,5-Tetrachlorobenzene	rel_int	52	215.8881	217.8852	2.08	0.25
3	1,2,3,4-Tetrachlorobenzene	rel_int	52	215.8881	217.8852	2.08	0.25
4	Pentachlorobenzene	rel_int	52	249.8491	251.8462	1.56	0.25
5	Hexachlorobenzene	rel_int	53	283.8102	285.8072	1.23	0.25
6	3,4,5,6-Tetrachloroveratrole	rel_int	54	275.9092	277.9063	2.08	0.25
7	Pentachloroanisole	rel_int	54	279.8597	281.8568	1.56	0.25
8	alpha-BHC	rel_int	54	218.9116	220.9086	2.1	0.25
9	beta-BHC	rel_int	55	218.9116	220.9086	2.1	0.25
10	gamma-BHC	rel_int	55	218.9116	220.9086	2.1	0.25
11	delta-BHC	rel_int	55	218.9116	220.9086	2.1	0.25
12	Pentachloronitrobenzene	rel_int	56	294.8342	296.8313	1.56	0.25
13	Heptachlor	rel_int	56	271.8102	273.8072	1.25	0.25
14	Aldrin	rel_int	57	262.857	264.854	1.56	0.25
15	4,4'-DDNU	rel_int	57	248.016	250.013	1.56	0.25
16	Dacthal	rel_int	57	331.8991	333.8961	2.08	0.25
17	Chlorpyrifos	rel_int	57	313.9574	315.9545	1.44	0.25
18	Octachlorostyrene	rel_int	57	342.779	344.7761	1.04	0.25
19	Heptachlor Epoxide B	rel_int	57	352.844	354.841	1.2	0.25
20	Heptachlor Epoxide A	rel_int	57	352.844	354.841	1.2	0.25
21	Oxychlorane	rel_int	57	386.805	388.802	1.02	0.25
22	4,4'-DDMU	rel_int	62	247.0081	249.0052	1.56	0.25
23	trans-Chlordane	rel_int	58	262.8571	264.8541	1.56	0.25
24	cis-Chlordane	rel_int	58	262.8571	264.8541	1.56	0.25
25	trans-Nonachlor	rel_int	58	262.8571	264.8541	1.56	0.25
26	Dieldrin	rel_int	59	262.8571	264.8541	1.56	0.25
27	Endrin	rel_int	60	262.8571	264.8541	1.56	0.25
28	cis-Nonachlor	rel_int	58	262.8571	264.8541	1.56	0.25
29	Endosulfan I	rel_int	61	276.8726	278.8697	1.56	0.25
30	Endosulfan II	rel_int	61	276.8726	278.8697	1.56	0.25
31	Endosulfan Sulfate	rel_int	61	276.8726	278.8697	1.56	0.25
32	24'-DDE	rel_int	62	246.0003	247.9974	1.56	0.25
33	44'-DDE	rel_int	62	246.0003	247.9974	1.56	0.25
34	24'-DDD	rel_int	63	235.0082	237.0053	1.56	0.25
35	44'-DDD	rel_int	63	235.0082	237.0053	1.56	0.25
36	24'-DDT	rel_int	63	235.0082	237.0053	1.56	0.25
37	44'-DDT	rel_int	64	235.0082	237.0053	1.56	0.25
38	Endrin Aldehyde	rel_int	60	247.8521	249.8492	0.64	0.35
39	Endrin Ketone	rel_int	60	247.8521	249.8492	0.64	0.35
40	Methoxychlor	rel_int	65	227.1072	228.1106	6.2	0.35
41	Dicofol	rel_int	65	251.003	253.0001	1.56	0.35
42	Mirex	rel_int	66	271.8103	273.8073	1.3	0.25
43	Parlar-26	rel_int	66	304.9039	306.901	1.56	0.35
44	Parlar-50	rel_int	66	338.8649	340.862	1.25	0.35
45	Parlar-62	rel_int	66	338.8649	340.862	1.25	0.35

Laboratory Surrogate							
46	1,3-Dibromobenzene	rel_int	67	233.868	235.8659	0.51	0.25
47	Endrin Ketone	rel_int	69	316.904	318.901	1.56	0.35
Field Surrogate							
48	1,3,5-Tribromobenzene	rel_int	67	313.7764	315.7744	1.03	0.25
49	1,2,4,5-Tetrabromobenzene	rel_int	68	312.7686	314.7666	1.03	0.25
50	delta-BHC	rel_int	69	218.9116	220.9086	2.1	0.25
GPC Recovery Standard							
51	13C12-PCB-133	rel_int	69	299.947	301.944	0.78	0.25
Extraction Standard							
52	13C6-Pentachlorobenzene	rel_int	67	255.8693	257.8663	1.56	0.25
53	13C6-Hexachlorobenzene	rel_int	67	289.8303	291.8273	1.23	0.25
54	13C6-alpha-BHC	rel_int	67	224.9317	226.9287	2.1	0.25
55	d6-gamma-BHC	rel_int	67	223.943	225.94	2.1	0.25
56	13C10-Heptachlor	rel_int	68	276.8269	278.824	1.25	0.25
57	13C10-Oxychlorane	rel_int	68	396.8385	398.8355	1.02	0.25
58	13C10-trans-Nonachlor	rel_int	69	269.8804	271.8775	1.56	0.25
59	13C12-Dieldrin	rel_int	69	269.8804	271.8775	1.56	0.25
60	13C12-Endrin	rel_int	69	269.8804	271.8775	1.56	0.25
61	13C9-Endosulfan-II	rel_int	69	284.8995	286.8965	1.56	0.25
62	13C12-44'-DDE	rel_int	69	258.0405	260.0376	1.56	0.25
63	13C12-44'-DDD	rel_int	69	247.0483	249.0454	1.56	0.25
64	13C12-44'-DDT	rel_int	69	247.0483	249.0454	1.56	0.25
65	d6-Methoxychlor	rel_int	69	233.145	234.1484	6.2	0.35
66	13C10-Mirex	rel_int	69	276.8269	278.824	1.3	0.35
Injection Standard							
67	13C12-PCB-9	abs_int	100	234.0406	236.0376	1.52	0.15
68	13C12-PCB-52	abs_int	100	301.9625	303.9597	0.77	0.15
69	13C12-PCB-101	abs_int	100	337.9206	339.9176	1.56	0.15

Data Calculations:

a) Analyte Concentrations:

The relative response factor of each target relative to the standard against which it is to be calculated is determined using the area responses of both quantification ions via equation 9.1.

In cases where a native target is calculated against an exact labelled analogue, the quantification will be considered to be by isotope dilution. In other cases, the quantification will be considered to be by internal standard.

$$\text{RRF} = \frac{(A1_t + A2_t) C_s}{(A1_s + A2_s) C_t} \quad \text{Equ. 9.1}$$

Where,

$A1_t + A2_t$ = The areas of the two quantification ions for the target analyte

$A1_s + A2_s$ = The areas of the two quantification ions for the labelled compound against which the target analyte will be calculated.

C_t = The concentration in the calibration standard of the target analyte.

C_s = The concentration in the calibration standard of the labelled compound against which the target will be calculated.

For all analytes to be quantified and from the initial calibration series of standard injections, a table of RRFs is prepared. The relative standard deviation (%RSD, or the coefficient of variance) is checked to confirm that appropriate method criteria has been met as listed in Table 3. The average of the five or six levels of for each analyte, RRF_{av} is applied for quantification of samples according to Equations 9.2 and 9.3 below.

$$\text{Amount in sample (ng)} = \frac{(A1_n + A2_n) Q_l}{(A1_l + A2_l) (\text{RRF}_{av})} \quad \text{Equ. 9.2}$$

$$\text{Concentration in sample (ng/g or ng/L)} = \frac{(A1_n + A2_n) Q_l}{(A1_l + A2_l) (\text{RRF}_{av}) (W_s)} \quad \text{Equ. 9.3}$$

Where,

Q_l = The amount (pg) of labelled compound added to the sample

W_s = The weight (g) or volume (l) of sample

b) Extraction, Clean-up, and Sampling Standard Recovery Calculation:

The extraction, clean-up, and sampling standard recoveries are determined by Equation 9.4 below.

$$\% \text{ Recovery} = \frac{\text{Amount in sample}}{\text{Amount added to sample}} \times 100 \quad \text{Equ. 9.4}$$

c) Estimated Detection Limit

$$\text{EDL} = \frac{2.5 \times H_x \times Q_{es}}{H_{es} \times W \times \text{RFF}_{av}} \quad \text{Equ. 9.5}$$

Where,

EDL = estimated detection limit for native targets

H_x = sum of the height of the noise level for each quantification ions for the unlabeled target

H_{es} = Sum of the heights of responses of both quantification ions for the labelled extraction standard.

W = weight of volume of sample

RRF_{av} = average relative response factor

Q_{es} = Amount of extraction standard added

Chromatogram Annotation Codes

All manually integrated peaks are expanded and reprinted with the following annotations:

* Analyst Initials AA
 * Date YYMMDD
 * integration code CC

The Syntax is: Example:
 AAYYMMDDCC SK111220MB

Code	Mnemonic	Description
MB	Manual Baseline	The peak was manually integrated because the initial baseline was determined incorrectly by the software
MS	Manual Split	The peak was manually integrated because the peak was incorrectly or not split by the software
MJ/MC	Manual Join/Manual Combine	The peak was manually integrated because the peak was split by the software and the peak should be integrated as a single peak
MA	Manual Add	The peak was manually integrated because the signal:noise ratio was judged to be >2.5
MD	Manual Delete	The peak was excluded because the signal:noise ratio was judged to be <2.5
MX	Manual Exclude	The peak was excluded due to an interference
MT	Manual Time	The peak retention time was manually chosen

The following explanatory annotation codes may appear on the chromatograms of peaks that have been reviewed:

Code	Mnemonic	Description
+	Detected Peak	A peak was detected at this mass and retention time that was above 2.5:1 signal to noise
<	Below Detection Limit	The signal at this mass and retention time was below 2.5:1 signal to noise
EMPC	Estimated Maximum Possible Concentration	The signal at this mass and retention time is an interference such that the target compound could not be confirmed
X-RT	Not Detected due to Retention Time non-conformance	The signal at this retention time could not be used to positively identify the target compound because of retention time non-conformance (apex of quantification and confirmation ions do not maximize within the same two seconds, or the retention time of the peak does not fall within the expected range with respect to its labeled analogue)
X-LOC	Not Detected due to interference from a higher level of chlorination	The signal at this retention time is attributable to a fragment from a co-eluting compound at a higher level of chlorination, and cannot be used to positively identify the target. The result is expressed as an Estimated Maximum Possible Concentration (EMPC)
X-DPE	Not Detected due to diphenyl ether interference	The signal at this retention time is attributable to interference from a chlorinated diphenyl ether, and cannot be used to positively identify the target. The result is expressed as an Estimated Maximum Possible Concentration (EMPC)
X-IF	Not Detected due to interference	The signal at this retention time is attributable to a co-eluting interference, and cannot be used to positively identify the target. The result is expressed as an Estimated Maximum Possible Concentration (EMPC)

SVOC DATA PACKAGE

SECTION 4: CALIBRATION DATA

Including:

for Multi-Point Calibration(s)

- Multi-Point Calibration Tables
- Individual Quantitation Reports

for Continuing Calibration(s)

- Individual Quantitation Reports

ALS Life sciences

Calibration Summary Report

Calibration Level	Filename	Run Date
CS-1	6-180719B37	20-Jul-2018 10:51
CS-2	6-180719B46	20-Jul-2018 16:02
CS-3	6-180719B37	20-Jul-2018 10:51
CS-4	6-180719B46	20-Jul-2018 16:02
CS-5		
CS-6		
CS-7		

Approved:	<i>R. Bakthiari</i>
	--e-signature--
	20-Jul-2018

Target Analytes	Relative Response Factors							Mean	% RSD
	CS-1	CS-2	CS-3	CS-4	CS-5	CS-6	CS-7		
2,4'-DDE	1.423	1.479	1.423	1.479				1.451	3%
4,4'-DDE	1.142	1.135	1.142	1.135				1.139	0%
2,4'-DDD	1.055	0.957	1.055	0.957				1.006	7%
4,4'-DDD	1.079	1.180	1.079	1.180				1.130	6%
2,4'-DDT	0.965	0.832	0.965	0.832				0.899	10%
4,4'-DDT	0.899	0.917	0.899	0.917				0.908	1%
Extraction Standards									
4,4'-DDE, 13C12-	1.926	1.649	1.926	1.649				1.788	11%
4,4'-DDD, 13C12-	1.819	1.230	1.819	1.230				1.525	27%
4,4'-DDT, 13C12-	1.583	0.833	1.583	0.833				1.208	44%

ALS Life sciences

Calibration Report

ALS Sample ID **H6-18-CCV-0661**
 Analysis Method EPA 1699 (mod)
 Analysis Type Calibration

Filename	Inst #	Column	Run Date
6-180719B37	HRMS-6	HP5MSUSR160544H	20-Jul-2018 10:51

Approved: *R. Bakthiari*
 --e-signature--
 20-Jul-2018

Target Analytes	Ret. Time	Ion Ratio	Concentration ng/mL	Response	RRF
2,4'-DDE	20.95	1.56	50.00	3.25E+06	1.423
4,4'-DDE	21.89	1.54	50.00	2.61E+06	1.142
2,4'-DDD	22.12	1.63	50.00	2.41E+06	1.055
4,4'-DDD	23.1	1.61	50.00	2.33E+06	1.079
2,4'-DDT	23.19	1.64	50.00	2.08E+06	0.965
4,4'-DDT	24.17	1.56	50.00	1.69E+06	0.899
Extraction Standards					
4,4'-DDE, 13C12-	21.88	1.55	250.00	1.14E+07	1.926
4,4'-DDD, 13C12-	23.08	1.56	250.00	1.08E+07	1.819
4,4'-DDT, 13C12-	24.15	1.54	250.00	9.39E+06	1.583

ALS Life sciences

Calibration Report

ALS Sample ID **H6-18-CCV-0662**
 Analysis Method EPA 1699 (mod)
 Analysis Type Calibration

Filename	Inst #	Column	Run Date
6-180719B46	HRMS-6	HP5MSUSR160544H	20-Jul-2018 16:02

Approved: *R. Bakthiari*
 --e-signature--
 20-Jul-2018

Target Analytes	Ret. Time	Ion Ratio	Concentration ng/mL	Response	RRF
2,4'-DDE	20.95	1.56	50.00	3.86E+06	1.479
4,4'-DDE	21.88	1.56	50.00	2.96E+06	1.135
2,4'-DDD	22.12	1.63	50.00	2.50E+06	0.957
4,4'-DDD	23.1	1.61	50.00	2.30E+06	1.180
2,4'-DDT	23.18	1.62	50.00	1.62E+06	0.832
4,4'-DDT	24.15	1.56	50.00	1.21E+06	0.917
Extraction Standards					
4,4'-DDE, 13C12-	21.88	1.54	250.00	1.30E+07	1.649
4,4'-DDD, 13C12-	23.08	1.57	250.00	9.73E+06	1.230
4,4'-DDT, 13C12-	24.15	1.55	250.00	6.59E+06	0.833

ALS Life sciences

Calibration Report

ALS Sample ID **H6-18-CCV-0661**
 Analysis Method EPA 1699 (mod)
 Analysis Type Calibration

Filename 6-180719B37 Inst # HRMS-6 Column HP5MSUSR160544H Run Date 20-Jul-2018 10:51

Approved: *R. Bakhtiari*
 --e-signature--
 20-Jul-2018

Target Analytes	Ret. Time	Ion Ratio	Concentration ng/mL	Response	RRF
2,4'-DDE	20.95	1.56	50.00	3.25E+06	1.423
4,4'-DDE	21.89	1.54	50.00	2.61E+06	1.142
2,4'-DDD	22.12	1.63	50.00	2.41E+06	1.055
4,4'-DDD	23.1	1.61	50.00	2.33E+06	1.079
2,4'-DDT	23.19	1.64	50.00	2.08E+06	0.965
4,4'-DDT	24.17	1.56	50.00	1.69E+06	0.899
Extraction Standards					
4,4'-DDE, 13C12-	21.88	1.55	250.00	1.14E+07	1.926
4,4'-DDD, 13C12-	23.08	1.56	250.00	1.08E+07	1.819
4,4'-DDT, 13C12-	24.15	1.54	250.00	9.39E+06	1.583

ALS Life sciences

Calibration Report

ALS Sample ID **H6-18-CCV-0662**
 Analysis Method EPA 1699 (mod)
 Analysis Type Calibration

Filename 6-180719B46 Inst # HRMS-6 Column HP5MSUSR160544H Run Date 20-Jul-2018 16:02

Approved: *R. Bakhtiari*
 --e-signature--
 20-Jul-2018

Target Analytes	Ret. Time	Ion Ratio	Concentration ng/mL	Response	RRF
2,4'-DDE	20.95	1.56	50.00	3.86E+06	1.479
4,4'-DDE	21.88	1.56	50.00	2.96E+06	1.135
2,4'-DDD	22.12	1.63	50.00	2.50E+06	0.957
4,4'-DDD	23.1	1.61	50.00	2.30E+06	1.180
2,4'-DDT	23.18	1.62	50.00	1.62E+06	0.832
4,4'-DDT	24.15	1.56	50.00	1.21E+06	0.917
Extraction Standards					
4,4'-DDE, 13C12-	21.88	1.54	250.00	1.30E+07	1.649
4,4'-DDD, 13C12-	23.08	1.57	250.00	9.73E+06	1.230
4,4'-DDT, 13C12-	24.15	1.55	250.00	6.59E+06	0.833

SVOC DATA PACKAGE

SECTION 5: QC SAMPLE DATA

Including:

- Laboratory Method Blank Analysis Reports
- Laboratory Control Sample Analysis Reports
- Matrix Spike Analysis Reports
- Other QC Sample Analysis Reports (where applicable)

ALS Life sciences

Laboratory Method Blank Analysis Report

Sample Name	Method Blank	Sampling Date	n/a		
ALS Sample ID	WG2815918-1	Extraction Date	6-Jul-18		
Analysis Method	EPA 1699 (mod)	Sample Size	2	g	
Analysis Type	Blank	Percent Solids	n/a		
Sample Matrix	QC	Split Ratio	1		

Approved:
R. Bakthiari
--e-signature--
20-Jul-2018

Run Information	Run 1
Filename	6-180719B43
Run Date	20-Jul-18 14:22
Final Volume	1020 uL
Dilution Factor	1
Analysis Units	ng/g
Instrument - Column	HRMS-6 HP5MSUSR160544H

Target Analytes	Ret. Time	Conc. ng/g	EDL ng/g	Flags	EMPC ng/g	LQL
2,4'-DDE	NotFnd	<0.0049	0.0049	U		1.0
4,4'-DDE	21.89	0.0350	0.0086	J		1.0
2,4'-DDD	NotFnd	<0.015	0.015	U		1.0
4,4'-DDD	NotFnd	<0.010	0.010	U		1.0
4,4'-DDT	NotFnd	<0.013	0.013	U		1.0
Extraction Standards						
4,4'-DDE, 13C12-	125	21.88	84		47-160	
4,4'-DDD, 13C12-	125	23.08	82		5-150	
4,4'-DDT, 13C12-	125	24.15	84		5-120	

EDL Indicates the Estimated Detection Limit, based on the measured background noise for this target in this sample.

LQL Lower Quantification Limit, based on the lowest calibration level corrected for sample size, splits and dilutions.

M Indicates that a peak has been manually integrated.

U Indicates that this compound was not detected above the EDL.

J indicates that a target analyte was detected below the calibrated range.

R Indicates that the ion abundance ratio for this compound did not meet the acceptance criterion.

EMPC Estimated Maximum Possible Concentration – elevated detection limit due to interference or positive id criterion failure

ALS Life sciences

Laboratory Control Sample Analysis Report

Sample Name	Laboratory Control Sample	Sampling Date	n/a		
ALS Sample ID	WG2815918-2	Extraction Date	6-Jul-18		
Analysis Method	EPA 1699 (mod)	Sample Size	1	n/a	
Analysis Type	LCS	Percent Solids	n/a		
Sample Matrix	QC	Split Ratio	1		

Approved:
R. Bakthiari
--e-signature--
20-Jul-2018

Run Information	Run 1
Filename	6-180719B40
Run Date	20-Jul-18 12:43
Final Volume	1020 uL
Dilution Factor	1
Analysis Units	%
Instrument - Column	HRMS-6 HP5MSUSR160544H

Target Analytes	ng	Ret.		Limits	
		Time	% Rec		Flags
2,4'-DDE	25	20.95	101	24-123	
4,4'-DDE	25	21.88	99	50-120	
2,4'-DDD	25	22.12	107	50-120	
4,4'-DDD	25	23.10	96	42-120	
4,4'-DDT	25	24.17	96	50-120	
Extraction Standards					
	ng				
4,4'-DDE, 13C12-	125	21.88	82	26-169	
4,4'-DDD, 13C12-	125	23.08	89	13-200	
4,4'-DDT, 13C12-	125	24.15	88	13-200	

SVOC DATA PACKAGE

SECTION 6: INTERNAL RECORDS

Including:

- Prep Logs
- Independent calculation checks
- Others as listed below:

Procedure:

Extraction:

- Subsample size = 2g
- Mix with sodium sulphate (Use enough sodium sulphate and mixing to create a free-flowing mixture.)
- Spike Extraction and Native Standards.
- Soxhlet extract in DCM.
- Roto-vap down to ~2ml
- 1mL Pass through OCP Silicarb clean up
- Sili Carb Column - Load sample with 3x1mL hexane rinses
 - F1 = 25 mL of Hexane
 - F2 =250mL of 1:1 DCM:Hexane
- Reduce F2 to 1mL, vortex and transfer to robovial
- Spike injection standard and submit FV=1020ul

Approval of Deviation from Standard Method

- Procedure does not deviate from Standard Method. **(Batch Writer):** _____
- Procedure does deviate from Standard Method. **Approved (Supervisor/Manager):** _____

WG:		Prep Analyst:			
Analysis:		Date:			
	Very Good	meets Method Req	Some Outliers	Very Poor	Comments / was spreadsheet sent for review? Why?
MB					
LCS					
DUP					
ES rec					

ALS Life sciences

Sample Calculation Report

CS3 RRF Check

Approved:	<i>R. Bakthiari</i> --e-signature-- 20-Jul-2018
-----------	---

$$\begin{aligned}
 \text{RRF} &= \frac{\text{Response of 4,4'-DDE}}{\text{Response of 13C12-4,4'DDE}} \times \frac{\text{Concentration of 13C2-4,4'DDE}}{\text{Concentration of DDE}} \\
 \text{RRF} &= \frac{2609863.40}{11429690.50} \times \frac{250.00}{50} = 1.14 \quad \text{Value from TargetLyn} \times 1.14
 \end{aligned}$$

Calculation of 4,4'-DDE amount in L2125039-1

$$\begin{aligned}
 \text{ng} &= \frac{\text{Response of 4,4'-DDE}}{\text{Response of 13C12-4,4'DDE}} \times \frac{\text{ng of 13C12-4,4'-DDE spiked}}{\text{Mean RRF} * \text{Sample Size}} \\
 \text{ng/g} &= \frac{2832138.5}{1855507.9} \times \frac{125}{1.14 * 2.01} = 83.374 \quad 83.4
 \end{aligned}$$

Calculation of 13C12-4,4'-DDE Recovery in L2125039-1

$$\begin{aligned}
 \% \text{ Recovery} &= \frac{\text{Response of 13C12-4,4'-DDE}}{\text{Response of 13C12-PCB-101}} \times \frac{\text{ng of 13C12-PCB-101} * 100}{\text{Mean RRF} * \text{Amount Spiked}} \\
 \% \text{ Recovery} &= \frac{1855507.9}{933448.2} \times \frac{50 * 100}{1.79 * 125} = 44 \quad 44 \%
 \end{aligned}$$