

Data Validation Report

Project:	Portland Harbor
Laboratory:	Alpha Analytical Laboratory
Environmental Test Record (ETR):	1510016
Analyses/Method:	Polycyclic Aromatic Hydrocarbons (PAH), n-Alkanes and Total Petroleum Hydrocarbons (TPH), and Total Organic Carbon (TOC)

Summary

Twenty sediment samples were collected in Portland Harbor, Oregon on October 19, 2015 and October 20, 2015. Samples were analyzed for polycyclic aromatic hydrocarbons (PAH) by EPA Method 8270D modified by selected ion monitoring mode (SIM), n-alkanes and total petroleum hydrocarbons (TPH) by EPA Method 8015D, and total organic carbon (TOC) by EPA Method 9060A by Alpha Analytical Laboratory located in Mansfield, Massachusetts. The laboratory provided Level 4 data packages containing samples results and associated quality assurance (QA) and quality control (QC) data, preparation logs, and raw instrument output. The following sediment samples are associated with the laboratory ETR 1510016.

Sample ID	Lab ID	Matrix
PH15-20-C	1510016-01	Sediment
PH15-20-D	1510016-02	Sediment
PH15-21-C	1510016-03	Sediment
PH15-21-D	1510016-04	Sediment
PH15-22-A	1510016-05	Sediment
PH15-22-C	1510016-06	Sediment
PH15-22-D	1510016-07	Sediment
PH15-23-C	1510016-08	Sediment
PH15-23-D	1510016-09	Sediment
PH15-24-A	1510016-10	Sediment
PH15-24-C	1510016-11	Sediment
PH15-24-D	1510016-12	Sediment
PH15-25-C	1510016-13	Sediment
PH15-25-C-FD	1510016-14	Sediment
PH15-25-D	1510016-15	Sediment
PH15-26-A	1510016-16	Sediment
PH15-26-C	1510016-17	Sediment
PH15-26-D	1510016-18	Sediment
PH15-27-A	1510016-19X	Sediment
PH15-27-D	1510016-20	Sediment

The data have been independently validated using *USEPA Contact Laboratory Program National Functional Guidelines for Organic Superfund Methods Data Review* EPA-540-R-2017-002, dated January 2017. Validation includes reconstruction of the analytical data to verify that data are traceable and



sufficiently complete in order for a qualified individual other than the originator to perform reconstruction of the data. The validation included the following checks:

- Sample Receipt/Transcription error check
- Sample preservation
- Sample holding times
- Tune Summary
- Initial calibration
- Continuing calibration verification (CCV)
- Laboratory blank contamination
- Equipment blank contamination
- Surrogate spike recoveries
- Internal Standard recoveries
- Matrix spike/Matrix spike duplicate (MS/MSD) recoveries, relative percent difference (RPD)
- Standard Reference Material Sediment accuracy check
- Laboratory control sample (LCS), LCS Duplicate (LCSD) recoveries, RPD values
- Calculation checks
- Contract Required Quantitation Limit (CRQL)
- Field duplicate results
- Laboratory duplicate results
- Overall assessment of the data

Data validation is based on the QC criteria documented in *Portland Harbor Supplemental Sediment Study, Portland Oregon Quality Assurance Project Plan (QAPP)*,¹ dated October 14, 2015, and the *Portland Harbor Pre-Remedial Design Investigation and Baseline Sampling Quality Assurance Project Plan (QAPP)*,² dated March 23, 2018. Data qualifiers assigned to results reported in this sample set are included in Table 1. Reason codes and explanations for qualified data are provided in Table 2.

Sample Receipt

Chain of custody documentation were reviewed for completeness of information relevant to the samples and requested analysis. Sample IDs and sample collection dates from the chain of custody records were matched to the reported data. No discrepancies noted.

All coolers were received within $4 \pm 2^\circ\text{C}$.

ORGANIC ANALYSES

Holding Time and Sample Preservation

All samples were extracted and analyzed within holding times.

GC/MS Instrument Performance Check – Acceptable

¹ NewFields. (2015). Portland Harbor Supplemental Sediment Study, Portland Oregon Quality Assurance Project Plan (QAPP). October 14, 2015.

² AECOM and Geosyntec. 2018. Portland Harbor Pre-Remedial Design Investigation and Baseline Sampling Portland Harbor Superfund Site, Quality Assurance Project Plan. March 23, 2018,



Initial Calibration and Continuing Calibration Verifications – Acceptable

Blanks – Acceptable except as noted below:

Method Blank: The method blank SS110415B01 is associated with samples: PH15-20-C, PH15-20-D, PH15-21-C, PH15-21-D, PH15-22-A, PH15-22-C, PH15-22-D, PH15-23-C, PH15-23-D, PH15-24-A, PH15-24-C, PH15-24-D, PH15-25-C, PH15-25-C-FD, PH15-25-D, PH15-26-A, PH15-26-C, PH15-26-D, and PH15-27-D. The method blank SS111615B02 is associated with sample PH15-27-A.

The method blank SS110415B01 met the QC acceptance criteria for PAH. PAH were detected in the method blank below the reporting limit. However, with the exception of the analytes below, the associated sample results were either non-detect or were greater than 10X the blank concentration. Samples containing the below listed analytes at concentrations below the reporting limit were qualified as not detected, and were flagged “U” at the reporting limit based on the method blank result.

PAH Compounds	Result	Unit	Lab Qualifier
Naphthalene	0.0915	µg/Kg	J
Fluorene	0.0410	µg/Kg	J
Dibenzothiophene	0.0194	µg/Kg	J
C1-Dibenzothiophenes	0.0609	µg/Kg	J
4-Methyldibenzothiophene	0.0167	µg/Kg	J
1-Methyldibenzothiophene	0.0171	µg/Kg	J

The method blank SS111615B02 met the QC acceptance criteria for PAH. PAH were detected in the method blank below the reporting limit. However, with the exception of the analytes below, the associated sample results were either non-detect or were greater than 10X the blank concentration. Sample PH15-27-A containing the below listed analytes at concentrations below the reporting limit were qualified as not detected, and were flagged “U” at the reporting limit based on the method blank result.

PAH Compounds	Result	Unit	Lab Qualifier
Naphthalene	0.388	µg/Kg	J
Acenaphthene	0.311	µg/Kg	J
Fluorene	0.401	µg/Kg	J
Dibenzothiophene	0.232	µg/Kg	J
3-Methylphenanthrene	0.187	µg/Kg	J
2-Methylphenanthrene	0.139	µg/Kg	J

The method blank SS110415B01 met the QC acceptance criteria for n-alkanes and TPH. n-Alkanes were detected in the method blank below the reporting limit. However, with the exception of the analytes below, the associated sample results were either non-detect or were greater than 10X the blank concentration. Samples containing the below listed analytes at concentrations below the reporting limit were qualified as not detected, and were flagged “U” at the reporting limit based on the method blank result.

n-Alkanes and TPH Compounds	Result	Unit	Lab Qualifier
n-Octadecane (C18)	0.0300	mg/Kg	CJ
n-Pentacosane (C25)	0.0323	mg/Kg	CJ



The method blank SS111615B02 met the QC acceptance criteria for n-alkanes and TPH. n-Alkanes were detected in the method blank below the reporting limit. However, with the exception of the analytes below, the associated sample results were either non-detect or were greater than 10X the blank concentration. Sample PH15-27-A containing the below listed analytes at concentrations below the reporting limit was qualified as not detected, and were flagged “U” at the reporting limit based on the method blank result.

n-Alkanes and TPH Compounds	Result	Unit	Lab Qualifier
n-Decane (C10)	0.00800	mg/Kg	J
n-Pentadecane (C15)	0.0152	mg/Kg	J
n-Octadecane (C18)	0.192	mg/Kg	CJ
n-Eicosane (C20)	0.00280	mg/Kg	J
n-Pentacosane (C25)	0.198	mg/Kg	CJ
n-Hexacosane (C26)	0.00360	mg/Kg	J
n-Heptacosane (C27)	0.00360	mg/Kg	J
n-Octacosane (C28)	0.00880	mg/Kg	J
n-Hentriacontane (C31)	0.0140	mg/Kg	J

Rinsate Blank: Two rinsate blanks were collected on October 21, 2015 and October 22, 2015 (PH15-01-RB and PH15-02-RB, respectively [ETR 1510012]) and are associated with the samples in this ETR.

- PH15-01-RB is associated with: PH15-23-C and PH15-23-D.
- PH15-02-RB is associated with: PH15-20-C, PH15-20-D, PH15-21-C, PH15-21-D, PH15-22-A, PH15-22-C, PH15-22-D, PH15-24-A, PH15-24-C, PH15-24-D, PH15-25-C, PH15-25-C-FD, PH15-25-D, PH15-26-A, PH15-26-C, PH15-26-D, PH15-27-A and PH15-27-D.

Detections of target compounds in rinsate blanks were evaluated relative to sediment method detection limits (MDL). No target analytes were found in rinsate blanks at relative concentrations at, or above, the sediment MDL. No data were qualified based on the rinsate blank results.

Surrogate Spikes – Acceptable.

Internal Standard Areas – Acceptable.

Laboratory Control Samples – Acceptable except as noted below:

Analytes	LCS (%)	LCSD (%)	QC Limit (%)	RPD (%)	QC Limit (%)
Benzo[b]fluoranthene	ok	127	50 - 125	ok	30

The results for Benzo[b]fluoranthene were qualified as estimated and “J” flagged on the low LCS/LCSD recoveries.

Matrix Spike/Spike Duplicate – Acceptable except as noted below:

The following percent recoveries were outside QC limits:



Sample ID	Analyte	MS (%)	MSD (%)	QC Limit (%)	RPD (%)	QC Limit (%)
PH15-27-A	Benzo[b]fluoranthene	ok	132	50 - 125	ok	30
	Benzo[a]pyrene	ok	127	50 - 125	ok	30
	Indeno[1,2,3-cd]pyrene	ok	141	50 - 125	ok	30
	Dibenz[ah]anthracene/Dibenz[ac]anthracene	ok	135	50 - 125	ok	30

The results for benzo[b]fluoranthene, benzo[a]pyrene, Indeno[1,2,3-cd]pyrene, and dibenz[ah]anthracene/dibenz[ac]anthracene in the native sample were qualified as estimated and flagged “J” based on these MS/MSD results.

Standard Reference Material – Acceptable.

Field Duplicate– Acceptable except as noted below:

A field duplicate was submitted for PH15-25-C and was identified as PH15-25-C-FD. The results for the field duplicates were comparable except as noted below.

Sample ID	Field Duplicate ID	Analyte	RPD (%)	QC Limit (%)
PH15-25-C	PH15-25-C -FD	Phenanthrene	58	50
		3-Methylphenanthrene	53	50
		2-Methylphenanthrene	76	50

The results for the analytes listed above were qualified as estimated and flagged “J” based on elevated field duplicates.

Laboratory Duplicate– Acceptable except as noted below:

Sample ID	Analytes	RPD (%)	QC Limit (%)
PH15-27-D	C1-Benzo(b)thiophenes	48	30
	C4-Benzo(b)thiophenes	41	30
	C1-Naphthalenes	31	30
	Acenaphthylene	52	30
	Acenaphthene	40	30
	C3-Fluorenes	48	30
	C2-Phenanthrenes/Anthracenes	33	30
	C3-Phenanthrenes/Anthracenes	42	30
	C4-Phenanthrenes/Anthracenes	39	30
	C2-Dibenzothiophenes	35	30
	C3-Dibenzothiophenes	40	30
	C4-Dibenzothiophenes	32	30
	Benzo(b)fluorene	40	30
	Pyrene	32	30
	C1-Fluoranthenes/Pyrenes	41	30
	C2-Fluoranthenes/Pyrenes	44	30
C3-Fluoranthenes/Pyrenes	45	30	



Sample ID	Analytes	RPD (%)	QC Limit (%)
	C4-Fluoranthenes/Pyrenes	54	30
	Naphthobenzothiophenes	42	30
	C1-Naphthobenzothiophenes	39	30
	C2-Naphthobenzothiophenes	41	30
	C3-Naphthobenzothiophenes	46	30
	Benz[a]anthracene	40	30
	Chrysene/Triphenylene	33	30
	C1-Chrysenes	42	30
	C2-Chrysenes	44	30
	Benzo[b]fluoranthene	43	30
	Benzo[j]fluoranthene/Benzo[k]fluoranthene	41	30
	Benzo[a]fluoranthene	48	30
	Benzo[e]pyrene	41	30
	Benzo[a]pyrene	44	30
	Perylene	40	30
	Indeno[1,2,3-cd]pyrene	45	30
	Dibenz[ah]anthracene/Dibenz[ac]anthracene	42	30
	Benzo[g,h,i]perylene	43	30
	1-Methylnaphthalene	49	30

The results for the analytes listed above were qualified as estimated and flagged “J” based on elevated laboratory duplicates.

Target Compound Identifications– Acceptable.

Compound Quantitation and CRQLs – Acceptable.

CONVENTIONAL ANALYSES

Holding Time and Sample Preservation – Acceptable.

Initial Calibration and Continuing Calibration Verifications – Acceptable.

Blanks– Acceptable.

Matrix Spike/Spike Duplicate – Acceptable.

Standard Reference Material – Acceptable.

Field Duplicate– Acceptable.

Laboratory Duplicate– Acceptable.

Compound Quantitation and CRQLs – Acceptable




OVERALL ASSESSMENT OF DATA

The data reported in this laboratory ETR is considered usable for meeting the project objectives.

The completeness is calculated by the number of usable data points divided by the total number of data points generated, multiplied by 100. The completeness for the laboratory ETR is 100%.

Validation performed by and Date:

George Desreuisseau, Mike Mitchel and Kerylynn Krahforst, December 2018.



Staff Scientists - NewFields

Table 1. QA/QC Summary Review

Sdg	SoilSampID	Lab_ID	AnalMeth	Analyte	Result	Lab_Flag	Units	NFG Result	NFG Qualifier	validator_reason_code
1510016	PH15-24-C	1510016-11	EPA 8270D	Fluorene	0.381	JB	µg/Kg	0.7251	U	bl
1510016	PH15-24-D	1510016-12	EPA 8270D	Fluorene	0.327	JB	µg/Kg	0.7913	U	bl
1510016	PH15-26-D	1510016-18	EPA 8270D	Fluorene	0.315	JB	µg/Kg	0.7632	U	bl
1510016	PH15-27-A	1510016-19X	EPA 8270D	2-Methylphenanthrene	0.707	JB	µg/Kg	4.2721	U	bl
1510016	PH15-27-A	1510016-19X	EPA 8270D	Fluorene	1.84	JB	µg/Kg	4.2721	U	bl
1510016	PH15-27-A	1510016-19X	EPA 8270D	Dibenzothiophene	1.41	JB	µg/Kg	4.2721	U	bl
1510016	PH15-20-C	1510016-01	EPA 8270D	1-Methyldibenzothiophene	0.0932	JB	µg/Kg	0.745	U	bl
1510016	PH15-20-C	1510016-01	EPA 8270D	Naphthalene	0.52	JB	µg/Kg	0.745	U	bl
1510016	PH15-20-D	1510016-02	EPA 8270D	4-Methyldibenzothiophene	0.116	JB	µg/Kg	0.7968	U	bl
1510016	PH15-20-D	1510016-02	EPA 8270D	Naphthalene	0.291	JB	µg/Kg	0.7968	U	bl
1510016	PH15-20-D	1510016-02	EPA 8270D	C1-Dibenzothiophenes	0.256	JB	µg/Kg	0.7968	U	bl
1510016	PH15-20-D	1510016-02	EPA 8270D	1-Methyldibenzothiophene	0.0496	JB	µg/Kg	0.7968	U	bl
1510016	PH15-24-C	1510016-11	EPA 8270D	4-Methyldibenzothiophene	0.154	JB	µg/Kg	0.7251	U	bl
1510016	PH15-24-C	1510016-11	EPA 8270D	C1-Dibenzothiophenes	0.388	JB	µg/Kg	0.7251	U	bl
1510016	PH15-24-C	1510016-11	EPA 8270D	1-Methyldibenzothiophene	0.0798	JB	µg/Kg	0.7251	U	bl
1510016	PH15-24-D	1510016-12	EPA 8270D	1-Methyldibenzothiophene	0.08	JB	µg/Kg	0.7913	U	bl
1510016	PH15-24-D	1510016-12	EPA 8270D	C1-Dibenzothiophenes	0.382	JB	µg/Kg	0.7913	U	bl
1510016	PH15-26-D	1510016-18	EPA 8270D	Naphthalene	0.472	JB	µg/Kg	0.7632	U	bl
1510016	PH15-26-D	1510016-18	EPA 8270D	1-Methyldibenzothiophene	0.0152	JB	µg/Kg	0.7632	U	bl
1510016	PH15-26-D	1510016-18	EPA 8270D	4-Methyldibenzothiophene	0.0299	JB	µg/Kg	0.7632	U	bl
1510016	PH15-26-D	1510016-18	EPA 8270D	C1-Dibenzothiophenes	0.121	JB	µg/Kg	0.7632	U	bl
1510016	PH15-27-A	1510016-19X	EPA 8270D	Naphthalene	2.1	JB	µg/Kg	4.2721	U	bl
1510016	PH15-27-A	1510016-19X	EPA 8270D	3-Methylphenanthrene	0.71	JB	µg/Kg	4.2721	U	bl
1510016	PH15-27-A	1510016-19X	EPA 8270D	Acenaphthene	3.08	JB	µg/Kg	4.2721	U	bl
1510016	PH15-25-C	1510016-13	EPA 8270D	Phenanthrene	37.8		µg/Kg		J	fd
1510016	PH15-25-C	1510016-13	EPA 8270D	3-Methylphenanthrene	9.87		µg/Kg		J	fd
1510016	PH15-25-C	1510016-13	EPA 8270D	2-Methylphenanthrene	9.03		µg/Kg		J	fd
1510016	PH15-25-C-FD	1510016-14	EPA 8270D	2-Methylphenanthrene	4.05		µg/Kg		J	fd
1510016	PH15-25-C-FD	1510016-14	EPA 8270D	3-Methylphenanthrene	5.71		µg/Kg		J	fd
1510016	PH15-25-C-FD	1510016-14	EPA 8270D	Phenanthrene	20.8		µg/Kg		J	fd
1510016	PH15-27-D	1510016-20	EPA 8270D	C4-Dibenzothiophenes	2.8		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	Dibenz[ah]anthracene/Dibenz[ac]anthracene	6.01		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	Chrysene/Triphenylene	38.9		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	C3-Phenanthrenes/Anthracenes	13.2		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	C4-Fluoranthenes/Pyrenes	4.99		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	C4-Benzo(b)thiophenes	1.68		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	Naphthobenzothiophenes	11.5		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	C4-Phenanthrenes/Anthracenes	5.63		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	Perylene	18.7		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	Pyrene	71.5		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	C3-Naphthobenzothiophenes	2.92		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	Acenaphthylene	8.4		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	Indeno[1,2,3-cd]pyrene	31.1		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	C3-Fluorenes	8.34		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	1-Methylnaphthalene	1.36		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	Acenaphthene	9.52		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	Benz[a]anthracene	33.8		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	Benzo[a]fluoranthene	8.9		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	Benzo[a]pyrene	51.9		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	Benzo[b]fluoranthene	31.2		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	Benzo[e]pyrene	29.8		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	Benzo[g,h,i]perylene	36.6		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	Benzo[j]fluoranthene/Benzo[k]fluoranthene	29.6		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	C2-Naphthobenzothiophenes	3.97		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	C3-Fluoranthenes/Pyrenes	6.99		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	C3-Dibenzothiophenes	5.88		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	C2-Phenanthrenes/Anthracenes	15.9		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	C2-Fluoranthenes/Pyrenes	14.7		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	C2-Dibenzothiophenes	5.77		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	C2-Chrysenes	9.95		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	C1-Naphthobenzothiophenes	5.62		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	C1-Naphthalenes	1.3		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	C1-Fluoranthenes/Pyrenes	32.6		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	C1-Chrysenes	18.4		µg/Kg		J	ld
1510016	PH15-27-D	1510016-20	EPA 8270D	C1-Benzo(b)thiophenes	1.71		µg/Kg		J	ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	C2-Phenanthrenes/Anthracenes	11.4	⌘	µg/Kg		J	ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	Naphthobenzothiophenes	7.49	⌘	µg/Kg		J	ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	C4-Benzo(b)thiophenes	1.11	⌘	µg/Kg		J	ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	C3-Fluoranthenes/Pyrenes	4.44	⌘	µg/Kg		J	ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	C3-Fluorenes	5.09	⌘	µg/Kg		J	ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	C3-Naphthobenzothiophenes	1.82	⌘	µg/Kg		J	ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	C3-Phenanthrenes/Anthracenes	8.57	⌘	µg/Kg		J	ld

Sdg	SoilSampID	Lab_ID	AnalMeth	Analyte	Result	Lab_Flag	Units	NFG		validator_ reason_code
								Result	Qualifier	
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	C3-Dibenzothiophenes	3.92	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	C4-Dibenzothiophenes	2.03	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	C4-Fluoranthenes/Pyrenes	2.86	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	C4-Phenanthrenes/Anthracenes	3.79	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	Chrysene/Triphenylene	27.8	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	Indeno[1,2,3-cd]pyrene	19.7	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	Perylene	12.5	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	C2-Naphthobenzothiophenes	2.62	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	Benz[a]anthracene	22.6	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	Dibenz[ah]anthracene/Dibenz[ac]anthracene	3.94	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	Benzo[j]fluoranthene/Benzo[k]fluoranthene	19.4	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	Pyrene	52	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	1-Methylnaphthalene	0.825	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	Acenaphthene	6.37	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	Acenaphthylene	4.94	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	Benzo[a]fluoranthene	5.47	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	Benzo[b]fluoranthene	20.3	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	Benzo[a]pyrene	33.4	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	Benzo[g,h,i]perylene	23.6	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	C2-Fluoranthenes/Pyrenes	9.34	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	C1-Benzo(b)thiophenes	1.05	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	C1-Chrysenes	12	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	C1-Fluoranthenes/Pyrenes	21.5	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	C1-Naphthalenes	0.953	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	C1-Naphthobenzothiophenes	3.79	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	C2-Chrysenes	6.36	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	C2-Dibenzothiophenes	4.04	⌘	µg/Kg	J		ld
1510016	PH15-27-D-DUP	1510016-20D	EPA 8270D	Benzo[e]pyrene	19.7	⌘	µg/Kg	J		ld
1510016	PH15-27-A	1510016-19X	EPA 8270D	Indeno[1,2,3-cd]pyrene	25.3		µg/Kg	J		m
1510016	PH15-27-A	1510016-19X	EPA 8270D	Benzo[a]pyrene	40.8		µg/Kg	J		m
1510016	PH15-27-A	1510016-19X	EPA 8270D	Dibenz[ah]anthracene/Dibenz[ac]anthracene	4.65		µg/Kg	J		m
1510016	PH15-27-A	1510016-19X	EPA 8270D	Benzo[b]fluoranthene	24.5		µg/Kg	J		m, l
1510016	PH15-20-C	1510016-01	EPA 8015M	n-Octadecane (C18)	0.03	CJB	mg/Kg	0.0745	U	bl
1510016	PH15-20-C	1510016-01	EPA 8015M	n-Pentacosane (C25)	0.039	CJB	mg/Kg	0.0745	U	bl
1510016	PH15-20-D	1510016-02	EPA 8015M	n-Octadecane (C18)	0.0385	CJB	mg/Kg	0.0797	U	bl
1510016	PH15-21-C	1510016-03	EPA 8015M	n-Octadecane (C18)	0.0349	CJB	mg/Kg	0.0735	U	bl
1510016	PH15-21-C	1510016-03	EPA 8015M	n-Pentacosane (C25)	0.0435	CJB	mg/Kg	0.0735	U	bl
1510016	PH15-21-D	1510016-04	EPA 8015M	n-Pentacosane (C25)	0.051	CJB	mg/Kg	0.0791	U	bl
1510016	PH15-21-D	1510016-04	EPA 8015M	n-Octadecane (C18)	0.033	CJB	mg/Kg	0.0791	U	bl
1510016	PH15-22-C	1510016-06	EPA 8015M	n-Pentacosane (C25)	0.0668	CJB	mg/Kg	0.0735	U	bl
1510016	PH15-22-C	1510016-06	EPA 8015M	n-Octadecane (C18)	0.0343	CJB	mg/Kg	0.0735	U	bl
1510016	PH15-22-D	1510016-07	EPA 8015M	n-Octadecane (C18)	0.0447	CJB	mg/Kg	0.0926	U	bl
1510016	PH15-23-C	1510016-08	EPA 8015M	n-Octadecane (C18)	0.0334	CJB	mg/Kg	0.0725	U	bl
1510016	PH15-23-C	1510016-08	EPA 8015M	n-Pentacosane (C25)	0.0469	CJB	mg/Kg	0.0725	U	bl
1510016	PH15-23-D	1510016-09	EPA 8015M	n-Pentacosane (C25)	0.0491	CJB	mg/Kg	0.086	U	bl
1510016	PH15-23-D	1510016-09	EPA 8015M	n-Octadecane (C18)	0.0386	CJB	mg/Kg	0.086	U	bl
1510016	PH15-24-A	1510016-10	EPA 8015M	n-Octadecane (C18)	0.0341	CJB	mg/Kg	0.0722	U	bl
1510016	PH15-24-A	1510016-10	EPA 8015M	n-Pentacosane (C25)	0.0575	CJB	mg/Kg	0.0722	U	bl
1510016	PH15-24-C	1510016-11	EPA 8015M	n-Pentacosane (C25)	0.0391	CJB	mg/Kg	0.0725	U	bl
1510016	PH15-24-C	1510016-11	EPA 8015M	n-Octadecane (C18)	0.0292	CJB	mg/Kg	0.0725	U	bl
1510016	PH15-24-D	1510016-12	EPA 8015M	n-Pentacosane (C25)	0.0453	CJB	mg/Kg	0.0791	U	bl
1510016	PH15-24-D	1510016-12	EPA 8015M	n-Octadecane (C18)	0.0362	CJB	mg/Kg	0.0791	U	bl
1510016	PH15-25-C	1510016-13	EPA 8015M	n-Pentacosane (C25)	0.0466	CJB	mg/Kg	0.0717	U	bl
1510016	PH15-25-C	1510016-13	EPA 8015M	n-Octadecane (C18)	0.0337	CJB	mg/Kg	0.0717	U	bl
1510016	PH15-25-C-FD	1510016-14	EPA 8015M	n-Pentacosane (C25)	0.0442	CJB	mg/Kg	0.0713	U	bl
1510016	PH15-25-C-FD	1510016-14	EPA 8015M	n-Octadecane (C18)	0.0343	CJB	mg/Kg	0.0713	U	bl
1510016	PH15-25-D	1510016-15	EPA 8015M	n-Pentacosane (C25)	0.0499	CJB	mg/Kg	0.0853	U	bl
1510016	PH15-25-D	1510016-15	EPA 8015M	n-Octadecane (C18)	0.0374	CJB	mg/Kg	0.0853	U	bl
1510016	PH15-26-A	1510016-16	EPA 8015M	n-Octadecane (C18)	0.0318	CJB	mg/Kg	0.0725	U	bl
1510016	PH15-26-A	1510016-16	EPA 8015M	n-Pentacosane (C25)	0.0455	CJB	mg/Kg	0.0725	U	bl
1510016	PH15-26-C	1510016-17	EPA 8015M	n-Octadecane (C18)	0.031	CJB	mg/Kg	0.0722	U	bl
1510016	PH15-26-C	1510016-17	EPA 8015M	n-Pentacosane (C25)	0.045	CJB	mg/Kg	0.0722	U	bl
1510016	PH15-26-D	1510016-18	EPA 8015M	n-Octadecane (C18)	0.0347	JB	mg/Kg	0.0763	U	bl
1510016	PH15-26-D	1510016-18	EPA 8015M	n-Pentacosane (C25)	0.0495	CJB	mg/Kg	0.0763	U	bl
1510016	PH15-27-A	1510016-19X	EPA 8015M	n-Pentadecane (C15)	0.0799	JB	mg/Kg	0.4272	U	bl
1510016	PH15-27-A	1510016-19X	EPA 8015M	n-Heptacosane (C27)	0.0222	JB	mg/Kg	0.4272	U	bl
1510016	PH15-27-A	1510016-19X	EPA 8015M	n-Decane (C10)	0.00812	JB	mg/Kg	0.4272	U	bl
1510016	PH15-27-A	1510016-19X	EPA 8015M	n-Hexacosane (C26)	0.00598	JB	mg/Kg	0.4272	U	bl
1510016	PH15-27-A	1510016-19X	EPA 8015M	n-Octacosane (C28)	0.00983	JB	mg/Kg	0.4272	U	bl
1510016	PH15-27-A	1510016-19X	EPA 8015M	n-Pentacosane (C25)	0.22	JB	mg/Kg	0.4272	U	bl
1510016	PH15-27-A	1510016-19X	EPA 8015M	n-Hentriacontane (C31)	0.0244	JB	mg/Kg	0.4272	U	bl
1510016	PH15-27-D	1510016-20	EPA 8015M	n-Octadecane (C18)	0.0342	CJB	mg/Kg	0.0812	U	bl
1510016	PH15-27-D	1510016-20	EPA 8015M	n-Pentacosane (C25)	0.0514	CJB	mg/Kg	0.0812	U	bl

Sdg	SoilSampID	Lab_ID	AnalMeth	Analyte	Result	Lab_Flag	Units	NFG NFG Result Qualifier	validator_ reason_code
1510016	PH15-27-A	1510016-19X	EPA 8015M	n-Eicosane (C20)	0.00726	JB	mg/Kg	0.4272 U	bl

Table 2. Reason Codes and Explanations

Reason Code	Explanation
bf	Field blank contamination
bl	Laboratory blank contamination
C	Calibration issue
el	Clean-up standard recovery
d	Reporting limit raised due to chromatographic interference
fd	Field duplicate RPDs
h	Holding Times
i	Internal standard areas
k	Estimated Maximum Possible Concentration (EMPC)
l	LCS or OPR recoveries
le	Labeled compound recovery
ld	Laboratory duplicate RPDs
lp	Laboratory control sample laboratory control sample duplicate RPDs
m	Matrix spike recovery
md	Matrix spike/matrix spike duplicate RPDs
nb	Negative laboratory blank contamination
p	Chemical preservation issue
r	Dual column RPD
q	Quantitation issue
s	Surrogate recovery
su	Ion suppression
t	Temperature preservation issue
x	Percent solids
y	Serial dilution results
z	ICS results