

SGS

AXYS

2045 Mills Road West

TEL: (250) 655-5800

Sidney, BC, Canada V8L5X2

TOLL-FREE: 1-888-373-0881

SGS AXYS Client No.: 4972

Client Address: AECOM
1111 Third Avenue, Suite 1600
Seattle, WA, US, 98101

The SGS AXYS contact for these data is Sean Campbell.

BATCH SUMMARY

Batch ID:	WG65583	Date:	21-Nov-2018
Analysis Type:	Dioxin/Furan	Matrix Type:	Filter
BATCH MAKEUP			
Contract: 4972 Samples: L29923-1 PDI-RB-XF-180820 L29967-1 PDI-WS-T04-1808 L29967-2 PDI-WS-T02-1808 L29967-3 PDI-WS-T05-1808 L29967-4 PDI-WS-T07-1808 L29967-5 PDI-WS-T03-1808 L29967-6 PDI-WS-T01-1808 L29967-7 PDI-WS-T06-1808		Blank: WG65583-101	
		Reference or Spike: WG65583-102	
		Duplicate: WG65583-103	
Comments: 1. Data are considered final. 2. Data are not blank corrected. Blank data should be taken into consideration when evaluating sample data. 3. Blank data should be evaluated against specifications using the same blank sample size as the size of the client samples. 4. A disturbance of the mass ion used to monitor instrument performance (lock-mass) greater than the method specification was observed in sample 'PDI-WS-T05-1808' (AXYS ID: L29967-3) near the retention time corresponding to 1,2,3,7,8,9-HxCDD. This analyte has been flagged with a 'G'.			

Copyright SGS AXYS Analytical Services Ltd
February 2017

FQA-006 Rev. 4. 20-Sep-2013

SGS AXYS METHOD MLA-017 Rev 20

Form 1A
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-RB-XF-180820
Sample Collection:
20-Aug-2018 15:43

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.:	4972	Project No.	PORTLAND HARBOR PDI AND BASELINE WATER
Matrix:	FILTER	Lab Sample I.D.:	L29923-1 i2
Sample Receipt Date:	22-Aug-2018	Sample Size:	0.2 sample
Extraction Date:	10-Oct-2018	Initial Calibration Date:	24-Oct-2018
Analysis Date:	29-Oct-2018 Time: 12:35:59	Instrument ID:	HR GC/MS
Extract Volume (uL):	20	GC Column ID:	DB5
Injection Volume (uL):	1.0	Sample Data Filename:	DX8M_127A S: 6
Dilution Factor:	N/A	Blank Data Filename:	DX8M_127A S: 5
Concentration Units:	pg/sample	Cal. Ver. Data Filename:	DX8M_127A S: 1

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG ¹	CONCENTRATION FOUND	REPORTING LIMIT (RL) ²	ION ABUND. RATIO ³	RRT ³
2,3,7,8-TCDD	U		4.49 (Q)		
1,2,3,7,8-PECDD ⁴	U		4.49 (Q)		
1,2,3,4,7,8-HXCDD	U		4.49 (Q)		
1,2,3,6,7,8-HXCDD	U		4.49 (Q)		
1,2,3,7,8,9-HXCDD	U		4.49 (Q)		
1,2,3,4,6,7,8-HPCDD	U		4.49 (Q)		
OCDD	J	18.9	4.49 (Q)	0.77	1.000
2,3,7,8-TCDF	U		4.49 (Q)		
1,2,3,7,8-PECDF	U		4.49 (Q)		
2,3,4,7,8-PECDF	U		4.49 (Q)		
1,2,3,4,7,8-HXCDF	U		4.49 (Q)		
1,2,3,6,7,8-HXCDF	U		4.49 (Q)		
1,2,3,7,8,9-HXCDF	U		4.49 (Q)		
2,3,4,6,7,8-HXCDF	U		4.49 (Q)		
1,2,3,4,6,7,8-HPCDF	U		4.49 (Q)		
1,2,3,4,7,8,9-HPCDF	U		4.49 (Q)		
OCDF	U		4.49 (Q)		

- (1) Where applicable, custom lab flags have been used on this report; U = not detected at RL; J = concentration less than lowest calibration equivalent.
(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.
(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613.
(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ Ting Chen _____

For Axy Internal Use Only [XSL Template: Form1A.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_L29923-1_Form1A_DX8M_127AS6_SJ2456918.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 2
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-RB-XF-180820
Sample Collection:
20-Aug-2018 15:43

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.:	4972	Project No.	PORTLAND HARBOR PDI AND BASELINE WATER
Matrix:	FILTER	Lab Sample I.D.:	L29923-1 i2
Sample Receipt Date:	22-Aug-2018	Sample Size:	0.2 sample
Extraction Date:	10-Oct-2018	Initial Calibration Date:	24-Oct-2018
Analysis Date:	29-Oct-2018 Time: 12:35:59	Instrument ID:	HR GC/MS
Extract Volume (uL):	20	GC Column ID:	DB5
Injection Volume (uL):	1.0	Sample Data Filename:	DX8M_127A S: 6
Dilution Factor:	N/A	Blank Data Filename:	DX8M_127A S: 5
Concentration Units:	pg absolute	Cal. Ver. Data Filename:	DX8M_127A S: 1

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

LABELED COMPOUND	LAB FLAG ¹	SPIKE CONC.	CONC. FOUND	R(%) ²	ION ABUND. RATIO ³	RRT ³
13C-2,3,7,8-TCDD		4000	3420	85.5	0.78	1.013
13C-1,2,3,7,8-PECDD ⁴		4000	2440	61.0	0.63	1.382
13C-1,2,3,4,7,8-HXCDD		4000	2400	60.1	1.34	0.988
13C-1,2,3,6,7,8-HXCDD		4000	2740	68.5	1.19	0.990
13C-1,2,3,4,6,7,8-HPCDD		4000	2340	58.4	1.07	1.094
13C-OCDD		8000	3900	48.7	0.92	1.178
13C-2,3,7,8-TCDF		4000	3110	77.7	0.78	0.966
13C-1,2,3,7,8-PECDF		4000	2570	64.2	1.57	1.283
13C-2,3,4,7,8-PECDF		4000	2310	57.7	1.56	1.351
13C-1,2,3,4,7,8-HXCDF		4000	2130	53.3	0.51	0.954
13C-1,2,3,6,7,8-HXCDF		4000	2190	54.9	0.52	0.958
13C-1,2,3,7,8,9-HXCDF		4000	3100	77.5	0.52	1.004
13C-2,3,4,6,7,8-HXCDF		4000	2250	56.4	0.51	0.982
13C-1,2,3,4,6,7,8-HPCDF		4000	2000	50.0	0.45	1.062
13C-1,2,3,4,7,8,9-HPCDF		4000	2190	54.8	0.45	1.104

CLEANUP STANDARD

37CL-2,3,7,8-TCDD		400	318	79.5		1.001
-------------------	--	-----	-----	------	--	-------

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for percent recovery (R) are specified in Section 9.3.3, Method 1613.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613. NOTE: There is no ion abundance ratio for 37Cl4-2,3,7,8-TCDD

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

For Axys Internal Use Only [XSL Template: Form2.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_L29923-1_Form2_DX8M_127AS6_SJ2456918.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 1A
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T04-1808
Sample Collection:
23-Aug-2018 10:08

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.:	4972	Project No.	PORTLAND HARBOR PDI AND BASELINE WATER
Matrix:	FILTER	Lab Sample I.D.:	L29967-1 i2
Sample Receipt Date:	29-Aug-2018	Sample Size:	0.2 sample
Extraction Date:	10-Oct-2018	Initial Calibration Date:	24-Oct-2018
Analysis Date:	29-Oct-2018 Time: 13:31:07	Instrument ID:	HR GC/MS
Extract Volume (uL):	20	GC Column ID:	DB5
Injection Volume (uL):	1.0	Sample Data Filename:	DX8M_127A S: 7
Dilution Factor:	N/A	Blank Data Filename:	DX8M_127A S: 5
Concentration Units:	pg/sample	Cal. Ver. Data Filename:	DX8M_127A S: 1

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG ¹	CONCENTRATION FOUND	REPORTING LIMIT (RL) ²	ION ABUND. RATIO ³	RRT ³
2,3,7,8-TCDD	U		4.33 (Q)		
1,2,3,7,8-PECDD ⁴	U		4.33 (Q)		
1,2,3,4,7,8-HXCDD	K J	6.29	4.33 (Q)	0.80	1.000
1,2,3,6,7,8-HXCDD	K J	23.5	4.33 (Q)	1.43	1.001
1,2,3,7,8,9-HXCDD	J	17.6	4.33 (Q)	1.05	1.010
1,2,3,4,6,7,8-HPCDD		510	4.33 (Q)	1.06	1.000
OCDD		4380	4.33 (Q)	0.90	1.000
2,3,7,8-TCDF		45.8	4.33 (Q)	0.80	1.001
1,2,3,7,8-PECDF	J	38.8	4.33 (Q)	1.42	1.000
2,3,4,7,8-PECDF	J	17.9	4.33 (Q)	1.35	1.001
1,2,3,4,7,8-HXCDF	J	80.8	4.33 (Q)	1.24	1.001
1,2,3,6,7,8-HXCDF	J	21.2	4.33 (Q)	1.28	1.000
1,2,3,7,8,9-HXCDF	U		4.33 (Q)		
2,3,4,6,7,8-HXCDF	J	6.46	4.33 (Q)	1.27	1.001
1,2,3,4,6,7,8-HPCDF		150	4.33 (Q)	1.03	1.000
1,2,3,4,7,8,9-HPCDF	J	20.1	4.33 (Q)	1.20	1.000
OCDF		362	4.33 (Q)	0.87	1.002

(1) Where applicable, custom lab flags have been used on this report; U = not detected at RL; K = peak detected but did not meet quantification criteria, result reported represents the estimated maximum possible concentration; J = concentration less than lowest calibration equivalent.

(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613.

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ Ting Chen _____

For Axys Internal Use Only [XSL Template: Form1A.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_L29967-1_Form1A_DX8M_127AS7_SJ2456919.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 2
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T04-1808
Sample Collection:
23-Aug-2018 10:08

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.: 4972

Matrix: FILTER

Sample Receipt Date: 29-Aug-2018

Extraction Date: 10-Oct-2018

Analysis Date: 29-Oct-2018 Time: 13:31:07

Extract Volume (uL): 20

Injection Volume (uL): 1.0

Dilution Factor: N/A

Concentration Units: pg absolute

Project No.

Lab Sample I.D.:

Sample Size:

Initial Calibration Date:

Instrument ID:

GC Column ID:

Sample Data Filename:

Blank Data Filename:

Cal. Ver. Data Filename:

PORTLAND HARBOR PDI AND
BASELINE WATER

L29967-1 i2

0.2 sample

24-Oct-2018

HR GC/MS

DB5

DX8M_127A S: 7

DX8M_127A S: 5

DX8M_127A S: 1

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

LABELED COMPOUND	LAB FLAG ¹	SPIKE CONC.	CONC. FOUND	R(%) ²	ION ABUND. RATIO ³	RRT ³
13C-2,3,7,8-TCDD		4000	3450	86.3	0.77	1.013
13C-1,2,3,7,8-PECDD ⁴		4000	3580	89.5	0.62	1.382
13C-1,2,3,4,7,8-HXCDD		4000	3430	85.8	1.31	0.987
13C-1,2,3,6,7,8-HXCDD		4000	3870	96.8	1.26	0.990
13C-1,2,3,4,6,7,8-HPCDD		4000	3260	81.6	1.06	1.094
13C-OCDD		8000	5160	64.6	0.89	1.179
13C-2,3,7,8-TCDF		4000	3360	84.0	0.78	0.966
13C-1,2,3,7,8-PECDF		4000	3510	87.9	1.57	1.283
13C-2,3,4,7,8-PECDF		4000	3350	83.9	1.59	1.351
13C-1,2,3,4,7,8-HXCDF		4000	3270	81.7	0.51	0.954
13C-1,2,3,6,7,8-HXCDF		4000	3320	82.9	0.52	0.958
13C-1,2,3,7,8,9-HXCDF		4000	3190	79.6	0.52	1.004
13C-2,3,4,6,7,8-HXCDF		4000	3230	80.7	0.52	0.981
13C-1,2,3,4,6,7,8-HPCDF		4000	3050	76.3	0.45	1.062
13C-1,2,3,4,7,8,9-HPCDF		4000	3100	77.5	0.45	1.104

CLEANUP STANDARD

37CL-2,3,7,8-TCDD		400	323	80.9		1.001
-------------------	--	-----	-----	------	--	-------

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for percent recovery (R) are specified in Section 9.3.3, Method 1613.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613. NOTE: There is no ion abundance ratio for 37Cl4-2,3,7,8-TCDD

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

For Axys Internal Use Only [XSL Template: Form2.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_L29967-1_Form2_DX8M_127AS7_SJ2456919.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 1A
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T04-1808
Sample Collection:
23-Aug-2018 10:08

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.: 4972

Matrix: FILTER

Sample Receipt Date: 29-Aug-2018

Extraction Date: 10-Oct-2018

Analysis Date: 23-Oct-2018 Time: 23:39:23

Extract Volume (uL): 20

Injection Volume (uL): 2.0

Dilution Factor: N/A

Concentration Units: pg/sample

Project No.

Lab Sample I.D.:

Sample Size:

Initial Calibration Date:

Instrument ID:

GC Column ID:

Sample Data Filename:

Blank Data Filename:

Cal. Ver. Data Filename:

PORTLAND HARBOR PDI AND
BASELINE WATER
L29967-1

0.2 sample

10-Jul-2018

HR GC/MS

DB225

DB83_183 S: 7

DX8M_127A S: 5

DB83_183 S: 2

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG ¹	CONCENTRATION FOUND	REPORTING LIMIT (RL) ²	ION ABUND. RATIO ³	RRT ³
1,2,3,7,8,9-HXCDD	K J	16.2	4.33 (Q)	1.89	1.064
2,3,7,8-TCDF		36.7	4.33 (Q)	0.76	1.001

(1) Where applicable, custom lab flags have been used on this report; K = peak detected but did not meet quantification criteria, result reported represents the estimated maximum possible concentration; J = concentration less than lowest calibration equivalent.

(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

For Axys Internal Use Only [XSL Template: Form1A.xsl; Created: 21-Nov-2018 17:04:30; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB225_L29967-1_Form1A_DB83_183S7_SJ2455695.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 1A
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T02-1808
Sample Collection:
24-Aug-2018 12:22

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.:	4972	Project No.	PORTLAND HARBOR PDI AND BASELINE WATER
Matrix:	FILTER	Lab Sample I.D.:	L29967-2 i2
Sample Receipt Date:	29-Aug-2018	Sample Size:	0.2 sample
Extraction Date:	10-Oct-2018	Initial Calibration Date:	24-Oct-2018
Analysis Date:	29-Oct-2018 Time: 14:26:15	Instrument ID:	HR GC/MS
Extract Volume (uL):	20	GC Column ID:	DB5
Injection Volume (uL):	1.0	Sample Data Filename:	DX8M_127A S: 8
Dilution Factor:	N/A	Blank Data Filename:	DX8M_127A S: 5
Concentration Units:	pg/sample	Cal. Ver. Data Filename:	DX8M_127A S: 1

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG ¹	CONCENTRATION FOUND	REPORTING LIMIT (RL) ²	ION ABUND. RATIO ³	RRT ³
2,3,7,8-TCDD	U		4.34 (Q)		
1,2,3,7,8-PECDD ⁴	K J	5.96	4.34 (Q)	0.40	1.001
1,2,3,4,7,8-HXCDD	J	7.98	4.34 (Q)	1.38	1.000
1,2,3,6,7,8-HXCDD	J	21.9	4.34 (Q)	1.25	1.000
1,2,3,7,8,9-HXCDD	J	14.7	4.34 (Q)	1.26	1.010
1,2,3,4,6,7,8-HPCDD		481	4.34 (Q)	1.09	1.000
OCDD		3870	4.34 (Q)	0.88	1.000
2,3,7,8-TCDF	J	16.1	4.34 (Q)	0.72	1.001
1,2,3,7,8-PECDF	J	12.8	4.34 (Q)	1.71	1.001
2,3,4,7,8-PECDF	J	6.96	4.34 (Q)	1.51	1.001
1,2,3,4,7,8-HXCDF	J	25.1	4.34 (Q)	1.17	1.000
1,2,3,6,7,8-HXCDF	J	10.3	4.34 (Q)	1.14	1.000
1,2,3,7,8,9-HXCDF	U		4.34 (Q)		
2,3,4,6,7,8-HXCDF	J	5.23	4.34 (Q)	1.14	1.001
1,2,3,4,6,7,8-HPCDF		114	4.34 (Q)	1.05	1.000
1,2,3,4,7,8,9-HPCDF	J	10.2	4.34 (Q)	0.90	1.000
OCDF		312	4.34 (Q)	0.83	1.002

(1) Where applicable, custom lab flags have been used on this report; U = not detected at RL; K = peak detected but did not meet quantification criteria, result reported represents the estimated maximum possible concentration; J = concentration less than lowest calibration equivalent.

(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613.

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

For Axys Internal Use Only [XSL Template: Form1A.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_L29967-2_Form1A_DX8M_127AS8_SJ2456920.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 2
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T02-1808
Sample Collection:
24-Aug-2018 12:22

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.:	4972	Project No.	PORTLAND HARBOR PDI AND BASELINE WATER
Matrix:	FILTER	Lab Sample I.D.:	L29967-2 i2
Sample Receipt Date:	29-Aug-2018	Sample Size:	0.2 sample
Extraction Date:	10-Oct-2018	Initial Calibration Date:	24-Oct-2018
Analysis Date:	29-Oct-2018 Time: 14:26:15	Instrument ID:	HR GC/MS
Extract Volume (uL):	20	GC Column ID:	DB5
Injection Volume (uL):	1.0	Sample Data Filename:	DX8M_127A S: 8
Dilution Factor:	N/A	Blank Data Filename:	DX8M_127A S: 5
Concentration Units:	pg absolute	Cal. Ver. Data Filename:	DX8M_127A S: 1

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

LABELED COMPOUND	LAB FLAG ¹	SPIKE CONC.	CONC. FOUND	R(%) ²	ION ABUND. RATIO ³	RRT ³
13C-2,3,7,8-TCDD		4000	3410	85.3	0.78	1.013
13C-1,2,3,7,8-PECDD ⁴		4000	3550	88.8	0.63	1.382
13C-1,2,3,4,7,8-HXCDD		4000	3600	90.0	1.26	0.987
13C-1,2,3,6,7,8-HXCDD		4000	3860	96.4	1.24	0.990
13C-1,2,3,4,6,7,8-HPCDD		4000	3410	85.2	1.06	1.094
13C-OCDD		8000	5820	72.8	0.90	1.178
13C-2,3,7,8-TCDF		4000	3340	83.5	0.78	0.966
13C-1,2,3,7,8-PECDF		4000	3470	86.8	1.58	1.283
13C-2,3,4,7,8-PECDF		4000	3330	83.2	1.56	1.351
13C-1,2,3,4,7,8-HXCDF		4000	3420	85.5	0.51	0.954
13C-1,2,3,6,7,8-HXCDF		4000	3480	87.0	0.52	0.958
13C-1,2,3,7,8,9-HXCDF		4000	3330	83.3	0.52	1.004
13C-2,3,4,6,7,8-HXCDF		4000	3350	83.7	0.52	0.981
13C-1,2,3,4,6,7,8-HPCDF		4000	3170	79.3	0.44	1.062
13C-1,2,3,4,7,8,9-HPCDF		4000	3270	81.7	0.44	1.104

CLEANUP STANDARD

37CL-2,3,7,8-TCDD		400	314	78.6		1.001
-------------------	--	-----	-----	------	--	-------

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for percent recovery (R) are specified in Section 9.3.3, Method 1613.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613. NOTE: There is no ion abundance ratio for 37Cl4-2,3,7,8-TCDD

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

For Axys Internal Use Only [XSL Template: Form2.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_L29967-2_Form2_DX8M_127AS8_SJ2456920.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 1A
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T02-1808
Sample Collection:
24-Aug-2018 12:22

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.: 4972

Matrix: FILTER

Sample Receipt Date: 29-Aug-2018

Extraction Date: 10-Oct-2018

Analysis Date: 24-Oct-2018 Time: 00:20:38

Extract Volume (uL): 20

Injection Volume (uL): 2.0

Dilution Factor: N/A

Concentration Units: pg/sample

Project No.

Lab Sample I.D.:

Sample Size:

Initial Calibration Date:

Instrument ID:

GC Column ID:

Sample Data Filename:

Blank Data Filename:

Cal. Ver. Data Filename:

PORTLAND HARBOR PDI AND
BASELINE WATER
L29967-2

0.2 sample

10-Jul-2018

HR GC/MS

DB225

DB83_183 S: 8

DX8M_127A S: 5

DB83_183 S: 2

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG ¹	CONCENTRATION FOUND	REPORTING LIMIT (RL) ²	ION ABUND. RATIO ³	RRT ³
1,2,3,7,8,9-HXCDD	K J	13.1	4.34 (Q)	0.55	1.063
2,3,7,8-TCDF	J	12.4	4.34 (Q)	0.71	1.002

(1) Where applicable, custom lab flags have been used on this report; K = peak detected but did not meet quantification criteria, result reported represents the estimated maximum possible concentration; J = concentration less than lowest calibration equivalent.

(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

For Axys Internal Use Only [XSL Template: Form1A.xsl; Created: 21-Nov-2018 17:04:30; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB225_L29967-2_Form1A_DB83_183S8_SJ2455696.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 1A
PCDD/PCDF ANALYSIS REPORT

CLIENT SAMPLE NO.
PDI-WS-T05-1808
Sample Collection:
21-Aug-2018 20:28

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.:	4972	Project No.	PORTLAND HARBOR PDI AND BASELINE WATER
Matrix:	FILTER	Lab Sample I.D.:	L29967-3 i2
Sample Receipt Date:	29-Aug-2018	Sample Size:	0.2 sample
Extraction Date:	10-Oct-2018	Initial Calibration Date:	24-Oct-2018
Analysis Date:	29-Oct-2018 Time: 15:21:29	Instrument ID:	HR GC/MS
Extract Volume (uL):	20	GC Column ID:	DB5
Injection Volume (uL):	1.0	Sample Data Filename:	DX8M_127A S: 9
Dilution Factor:	N/A	Blank Data Filename:	DX8M_127A S: 5
Concentration Units:	pg/sample	Cal. Ver. Data Filename:	DX8M_127A S: 1

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG ¹	CONCENTRATION FOUND	REPORTING LIMIT (RL) ²	ION ABUND. RATIO ³	RRT ³
2,3,7,8-TCDD	U		4.39 (Q)		
1,2,3,7,8-PECDD ⁴	U		4.39 (Q)		
1,2,3,4,7,8-HXCDD	K J	6.07	4.39 (Q)	0.76	1.000
1,2,3,6,7,8-HXCDD	J	22.0	4.39 (Q)	1.30	1.000
1,2,3,7,8,9-HXCDD	J	12.8	4.39 (Q)	1.17	1.010
1,2,3,4,6,7,8-HPCDD		385	4.39 (Q)	1.10	1.000
OCDD		3110	8.49 (S)	0.89	1.000
2,3,7,8-TCDF	J	6.66	4.39 (Q)	0.78	1.001
1,2,3,7,8-PECDF	U		4.39 (Q)		
2,3,4,7,8-PECDF	U		4.39 (Q)		
1,2,3,4,7,8-HXCDF	J	6.52	4.39 (Q)	1.09	1.000
1,2,3,6,7,8-HXCDF	U		4.39 (Q)		
1,2,3,7,8,9-HXCDF	U		4.39 (Q)		
2,3,4,6,7,8-HXCDF	U		4.39 (Q)		
1,2,3,4,6,7,8-HPCDF	J	86.3	4.39 (Q)	1.01	1.000
1,2,3,4,7,8,9-HPCDF	J	6.44	4.39 (Q)	0.94	1.000
OCDF		211	4.39 (Q)	0.85	1.002

(1) Where applicable, custom lab flags have been used on this report; U = not detected at RL; K = peak detected but did not meet quantification criteria, result reported represents the estimated maximum possible concentration; J = concentration less than lowest calibration equivalent.

(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613.

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ **Ting Chen** _____

For Axys Internal Use Only [XSL Template: Form1A.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_L29967-3_Form1A_DX8M_127AS9_SJ2456921.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 2
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T05-1808
Sample Collection:
21-Aug-2018 20:28

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.: 4972

Matrix: FILTER

Sample Receipt Date: 29-Aug-2018

Extraction Date: 10-Oct-2018

Analysis Date: 29-Oct-2018 Time: 15:21:29

Extract Volume (uL): 20

Injection Volume (uL): 1.0

Dilution Factor: N/A

Concentration Units: pg absolute

Project No. PORTLAND HARBOR PDI AND
BASELINE WATER

Lab Sample I.D.: L29967-3 i2

Sample Size: 0.2 sample

Initial Calibration Date: 24-Oct-2018

Instrument ID: HR GC/MS

GC Column ID: DB5

Sample Data Filename: DX8M_127A S: 9

Blank Data Filename: DX8M_127A S: 5

Cal. Ver. Data Filename: DX8M_127A S: 1

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

LABELED COMPOUND	LAB FLAG ¹	SPIKE CONC.	CONC. FOUND	R(%) ²	ION ABUND. RATIO ³	RRT ³
13C-2,3,7,8-TCDD		4000	3580	89.6	0.76	1.013
13C-1,2,3,7,8-PECDD ⁴		4000	3720	92.9	0.62	1.382
13C-1,2,3,4,7,8-HXCDD		4000	3490	87.3	1.29	0.987
13C-1,2,3,6,7,8-HXCDD		4000	3790	94.8	1.26	0.990
13C-1,2,3,4,6,7,8-HPCDD		4000	3690	92.3	1.04	1.094
13C-OCDD		8000	6830	85.4	0.90	1.178
13C-2,3,7,8-TCDF		4000	3600	89.9	0.78	0.966
13C-1,2,3,7,8-PECDF		4000	3630	90.7	1.58	1.283
13C-2,3,4,7,8-PECDF		4000	3460	86.4	1.57	1.350
13C-1,2,3,4,7,8-HXCDF		4000	3230	80.7	0.51	0.954
13C-1,2,3,6,7,8-HXCDF		4000	3380	84.6	0.51	0.958
13C-1,2,3,7,8,9-HXCDF		4000	3330	83.3	0.53	1.004
13C-2,3,4,6,7,8-HXCDF		4000	3260	81.6	0.51	0.981
13C-1,2,3,4,6,7,8-HPCDF		4000	3220	80.5	0.45	1.062
13C-1,2,3,4,7,8,9-HPCDF		4000	3450	86.3	0.44	1.104

CLEANUP STANDARD

37CL-2,3,7,8-TCDD		400	352	87.9		1.001
-------------------	--	-----	-----	------	--	-------

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for percent recovery (R) are specified in Section 9.3.3, Method 1613.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613. NOTE: There is no ion abundance ratio for 37Cl4-2,3,7,8-TCDD

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

For Axys Internal Use Only [XSL Template: Form2.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_L29967-3_Form2_DX8M_127AS9_SJ2456921.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 1A
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T05-1808
Sample Collection:
21-Aug-2018 20:28

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.: 4972

Matrix: FILTER

Sample Receipt Date: 29-Aug-2018

Extraction Date: 10-Oct-2018

Analysis Date: 24-Oct-2018 Time: 01:01:54

Extract Volume (uL): 20

Injection Volume (uL): 2.0

Dilution Factor: N/A

Concentration Units: pg/sample

Project No.

Lab Sample I.D.:

Sample Size:

Initial Calibration Date:

Instrument ID:

GC Column ID:

Sample Data Filename:

Blank Data Filename:

Cal. Ver. Data Filename:

PORTLAND HARBOR PDI AND
BASELINE WATER
L29967-3

0.2 sample

10-Jul-2018

HR GC/MS

DB225

DB83_183 S: 9

DX8M_127A S: 5

DB83_183 S: 2

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG ¹	CONCENTRATION FOUND	REPORTING LIMIT (RL) ²	ION ABUND. RATIO ³	RRT ³
1,2,3,7,8,9-HXCDD	K J G	8.14	4.39 (Q)	0.91	1.062
2,3,7,8-TCDF	U		4.39 (Q)		

(1) Where applicable, custom lab flags have been used on this report; U = not detected at RL; K = peak detected but did not meet quantification criteria, result reported represents the estimated maximum possible concentration; J = concentration less than lowest calibration equivalent; G = lock mass interference present.

(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

For Axy Internal Use Only [XSL Template: Form1A.xsl; Created: 21-Nov-2018 17:04:30; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB225_L29967-3_Form1A_DB83_183S9_SJ2455697.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 1A
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T07-1808
Sample Collection:
23-Aug-2018 18:20

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.:	4972	Project No.	PORTLAND HARBOR PDI AND BASELINE WATER
Matrix:	FILTER	Lab Sample I.D.:	L29967-4 i2
Sample Receipt Date:	29-Aug-2018	Sample Size:	0.2 sample
Extraction Date:	10-Oct-2018	Initial Calibration Date:	24-Oct-2018
Analysis Date:	29-Oct-2018 Time: 23:46:25	Instrument ID:	HR GC/MS
Extract Volume (uL):	20	GC Column ID:	DB5
Injection Volume (uL):	1.0	Sample Data Filename:	DX8M_127A S: 18
Dilution Factor:	N/A	Blank Data Filename:	DX8M_127A S: 5
Concentration Units:	pg/sample	Cal. Ver. Data Filename:	DX8M_127A S: 12

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG ¹	CONCENTRATION FOUND	REPORTING LIMIT (RL) ²	ION ABUND. RATIO ³	RRT ³
2,3,7,8-TCDD	U		4.25 (Q)		
1,2,3,7,8-PECDD ⁴	U		4.25 (Q)		
1,2,3,4,7,8-HXCDD	J	4.57	4.25 (Q)	1.35	1.000
1,2,3,6,7,8-HXCDD	J	11.6	4.25 (Q)	1.34	1.000
1,2,3,7,8,9-HXCDD	J	9.92	4.25 (Q)	1.41	1.010
1,2,3,4,6,7,8-HPCDD		223	4.25 (Q)	1.05	1.000
OCDD		1910	4.25 (Q)	0.88	1.000
2,3,7,8-TCDF	U		4.25 (Q)		
1,2,3,7,8-PECDF	U		4.25 (Q)		
2,3,4,7,8-PECDF	U		4.25 (Q)		
1,2,3,4,7,8-HXCDF	U		4.25 (Q)		
1,2,3,6,7,8-HXCDF	U		4.25 (Q)		
1,2,3,7,8,9-HXCDF	U		4.25 (Q)		
2,3,4,6,7,8-HXCDF	U		4.25 (Q)		
1,2,3,4,6,7,8-HPCDF	J	52.2	4.25 (Q)	1.05	1.000
1,2,3,4,7,8,9-HPCDF	U		4.25 (Q)		
OCDF	J	109	4.25 (Q)	0.90	1.002

(1) Where applicable, custom lab flags have been used on this report; U = not detected at RL; J = concentration less than lowest calibration equivalent.

(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613.

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ Ting Chen _____

For Axy Internal Use Only [XSL Template: Form1A.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_L29967-4_Form1A_DX8M_127AS18_SJ2457402.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 2
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T07-1808
Sample Collection:
23-Aug-2018 18:20

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.: 4972

Matrix: FILTER

Sample Receipt Date: 29-Aug-2018

Extraction Date: 10-Oct-2018

Analysis Date: 29-Oct-2018 Time: 23:46:25

Extract Volume (uL): 20

Injection Volume (uL): 1.0

Dilution Factor: N/A

Concentration Units: pg absolute

Project No.

Lab Sample I.D.:

Sample Size:

Initial Calibration Date:

Instrument ID:

GC Column ID:

Sample Data Filename:

Blank Data Filename:

Cal. Ver. Data Filename:

PORTLAND HARBOR PDI AND
BASELINE WATER
L29967-4 i2

0.2 sample

24-Oct-2018

HR GC/MS

DB5

DX8M_127A S: 18

DX8M_127A S: 5

DX8M_127A S: 12

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

LABELED COMPOUND	LAB FLAG ¹	SPIKE CONC.	CONC. FOUND	R(%) ²	ION ABUND. RATIO ³	RRT ³
13C-2,3,7,8-TCDD		4000	3110	77.8	0.78	1.013
13C-1,2,3,7,8-PECDD ⁴		4000	3160	79.1	0.64	1.382
13C-1,2,3,4,7,8-HXCDD		4000	3420	85.5	1.26	0.987
13C-1,2,3,6,7,8-HXCDD		4000	3630	90.9	1.25	0.990
13C-1,2,3,4,6,7,8-HPCDD		4000	3440	85.9	1.05	1.094
13C-OCDD		8000	6280	78.5	0.89	1.179
13C-2,3,7,8-TCDF		4000	3030	75.7	0.77	0.966
13C-1,2,3,7,8-PECDF		4000	3060	76.6	1.57	1.283
13C-2,3,4,7,8-PECDF		4000	2960	74.1	1.56	1.351
13C-1,2,3,4,7,8-HXCDF		4000	3220	80.5	0.52	0.954
13C-1,2,3,6,7,8-HXCDF		4000	3270	81.8	0.52	0.958
13C-1,2,3,7,8,9-HXCDF		4000	3150	78.8	0.51	1.004
13C-2,3,4,6,7,8-HXCDF		4000	3160	79.0	0.53	0.981
13C-1,2,3,4,6,7,8-HPCDF		4000	3080	77.0	0.45	1.062
13C-1,2,3,4,7,8,9-HPCDF		4000	3210	80.2	0.44	1.104

CLEANUP STANDARD

37CL-2,3,7,8-TCDD	400	299	74.7	1.001
-------------------	-----	-----	------	-------

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for percent recovery (R) are specified in Section 9.3.3, Method 1613.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613. NOTE: There is no ion abundance ratio for 37Cl4-2,3,7,8-TCDD

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

For Axys Internal Use Only [XSL Template: Form2.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_L29967-4_Form2_DX8M_127AS18_SJ2457402.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 1A
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T07-1808
Sample Collection:
23-Aug-2018 18:20

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.: 4972

Matrix: FILTER

Sample Receipt Date: 29-Aug-2018

Extraction Date: 10-Oct-2018

Analysis Date: 24-Oct-2018 Time: 01:43:07

Extract Volume (uL): 20

Injection Volume (uL): 2.0

Dilution Factor: N/A

Concentration Units: pg/sample

Project No.

Lab Sample I.D.:

Sample Size:

Initial Calibration Date:

Instrument ID:

GC Column ID:

Sample Data Filename:

Blank Data Filename:

Cal. Ver. Data Filename:

PORTLAND HARBOR PDI AND
BASELINE WATER
L29967-4

0.2 sample

10-Jul-2018

HR GC/MS

DB225

DB83_183 S: 10

DX8M_127A S: 5

DB83_183 S: 2

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG ¹	CONCENTRATION FOUND	REPORTING LIMIT (RL) ²	ION ABUND. RATIO ³	RRT ³
1,2,3,7,8,9-HXCDD	K J	13.0	4.25 (Q)	0.93	1.063

(1) Where applicable, custom lab flags have been used on this report; K = peak detected but did not meet quantification criteria, result reported represents the estimated maximum possible concentration; J = concentration less than lowest calibration equivalent.

(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

For Axs Internal Use Only [XSL Template: Form1A.xsl; Created: 21-Nov-2018 17:04:30; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB225_L29967-4_Form1A_DB83_183S10_SJ2455698.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 1A
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T03-1808
Sample Collection:
22-Aug-2018 17:55

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.:	4972	Project No.	PORTLAND HARBOR PDI AND BASELINE WATER
Matrix:	FILTER	Lab Sample I.D.:	L29967-5 i2 (A)
Sample Receipt Date:	29-Aug-2018	Sample Size:	0.2 sample
Extraction Date:	10-Oct-2018	Initial Calibration Date:	24-Oct-2018
Analysis Date:	29-Oct-2018 Time: 16:16:42	Instrument ID:	HR GC/MS
Extract Volume (uL):	20	GC Column ID:	DB5
Injection Volume (uL):	1.0	Sample Data Filename:	DX8M_127A S: 10
Dilution Factor:	N/A	Blank Data Filename:	DX8M_127A S: 5
Concentration Units:	pg/sample	Cal. Ver. Data Filename:	DX8M_127A S: 1

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG ¹	CONCENTRATION FOUND	REPORTING LIMIT (RL) ²	ION ABUND. RATIO ³	RRT ³
2,3,7,8-TCDD	U		4.35 (Q)		
1,2,3,7,8-PECDD ⁴	J	4.42	4.35 (Q)	0.70	1.000
1,2,3,4,7,8-HXCDD	J	8.63	4.35 (Q)	1.10	1.000
1,2,3,6,7,8-HXCDD	J	31.7	4.35 (Q)	1.41	1.000
1,2,3,7,8,9-HXCDD	J	21.6	4.35 (Q)	1.22	1.010
1,2,3,4,6,7,8-HPCDD		773	4.35 (Q)	1.04	1.000
OCDD		8820	4.35 (Q)	0.88	1.000
2,3,7,8-TCDF		20.0	4.35 (Q)	0.84	1.001
1,2,3,7,8-PECDF	K J	16.3	4.35 (Q)	1.21	1.000
2,3,4,7,8-PECDF	J	9.88	4.35 (Q)	1.36	1.000
1,2,3,4,7,8-HXCDF	J	39.0	4.35 (Q)	1.23	1.000
1,2,3,6,7,8-HXCDF	J	14.5	4.35 (Q)	1.25	1.000
1,2,3,7,8,9-HXCDF	U		4.35 (Q)		
2,3,4,6,7,8-HXCDF	J	6.19	4.35 (Q)	1.36	1.001
1,2,3,4,6,7,8-HPCDF		177	4.35 (Q)	1.01	1.000
1,2,3,4,7,8,9-HPCDF	J	14.2	4.35 (Q)	0.92	1.000
OCDF		483	4.35 (Q)	0.88	1.002

(1) Where applicable, custom lab flags have been used on this report; U = not detected at RL; K = peak detected but did not meet quantification criteria, result reported represents the estimated maximum possible concentration; J = concentration less than lowest calibration equivalent.

(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613.

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: Ting Chen

For Axys Internal Use Only [XSL Template: Form1A.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_L29967-5_Form1A_DX8M_127AS10_SJ2456908.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 2
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T03-1808
Sample Collection:
22-Aug-2018 17:55

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.:	4972	Project No.	PORTLAND HARBOR PDI AND BASELINE WATER
Matrix:	FILTER	Lab Sample I.D.:	L29967-5 i2 (A)
Sample Receipt Date:	29-Aug-2018	Sample Size:	0.2 sample
Extraction Date:	10-Oct-2018	Initial Calibration Date:	24-Oct-2018
Analysis Date:	29-Oct-2018 Time: 16:16:42	Instrument ID:	HR GC/MS
Extract Volume (uL):	20	GC Column ID:	DB5
Injection Volume (uL):	1.0	Sample Data Filename:	DX8M_127A S: 10
Dilution Factor:	N/A	Blank Data Filename:	DX8M_127A S: 5
Concentration Units:	pg absolute	Cal. Ver. Data Filename:	DX8M_127A S: 1

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

LABELLED COMPOUND	LAB FLAG ¹	SPIKE CONC.	CONC. FOUND	R(%) ²	ION ABUND. RATIO ³	RRT ³
13C-2,3,7,8-TCDD		4000	3230	80.8	0.79	1.013
13C-1,2,3,7,8-PECDD ⁴		4000	3270	81.8	0.63	1.382
13C-1,2,3,4,7,8-HXCDD		4000	3170	79.3	1.27	0.987
13C-1,2,3,6,7,8-HXCDD		4000	3400	85.0	1.25	0.990
13C-1,2,3,4,6,7,8-HPCDD		4000	3080	77.0	1.06	1.094
13C-OCDD		8000	4800	60.0	0.90	1.178
13C-2,3,7,8-TCDF		4000	3140	78.6	0.78	0.966
13C-1,2,3,7,8-PECDF		4000	3220	80.5	1.59	1.283
13C-2,3,4,7,8-PECDF		4000	3020	75.6	1.57	1.351
13C-1,2,3,4,7,8-HXCDF		4000	2980	74.5	0.52	0.954
13C-1,2,3,6,7,8-HXCDF		4000	3080	77.1	0.52	0.958
13C-1,2,3,7,8,9-HXCDF		4000	2900	72.5	0.52	1.004
13C-2,3,4,6,7,8-HXCDF		4000	2920	73.0	0.51	0.980
13C-1,2,3,4,6,7,8-HPCDF		4000	2710	67.7	0.45	1.062
13C-1,2,3,4,7,8,9-HPCDF		4000	2850	71.3	0.45	1.104

CLEANUP STANDARD

37CL-2,3,7,8-TCDD		400	299	74.6		1.001
-------------------	--	-----	-----	------	--	-------

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for percent recovery (R) are specified in Section 9.3.3, Method 1613.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613. NOTE: There is no ion abundance ratio for 37Cl4-2,3,7,8-TCDD

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

For Axys Internal Use Only [XSL Template: Form2.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_L29967-5_Form2_DX8M_127AS10_SJ2456908.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 1A
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T03-1808
Sample Collection:
22-Aug-2018 17:55

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.: 4972

Matrix: FILTER

Sample Receipt Date: 29-Aug-2018

Extraction Date: 10-Oct-2018

Analysis Date: 24-Oct-2018 Time: 02:24:19

Extract Volume (uL): 20

Injection Volume (uL): 2.0

Dilution Factor: N/A

Concentration Units: pg/sample

Project No.

Lab Sample I.D.:

Sample Size:

Initial Calibration Date:

Instrument ID:

GC Column ID:

Sample Data Filename:

Blank Data Filename:

Cal. Ver. Data Filename:

PORTLAND HARBOR PDI AND
BASELINE WATER
L29967-5 (A)

0.2 sample

10-Jul-2018

HR GC/MS

DB225

DB83_183 S: 11

DX8M_127A S: 5

DB83_183 S: 2

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG ¹	CONCENTRATION FOUND	REPORTING LIMIT (RL) ²	ION ABUND. RATIO ³	RRT ³
1,2,3,7,8,9-HXCDD	K J	17.5	4.35 (Q)	2.64	1.064
2,3,7,8-TCDF	K	19.2	4.35 (Q)	0.94	1.001

(1) Where applicable, custom lab flags have been used on this report; K = peak detected but did not meet quantification criteria, result reported represents the estimated maximum possible concentration; J = concentration less than lowest calibration equivalent.

(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

For Axys Internal Use Only [XSL Template: Form1A.xsl; Created: 21-Nov-2018 17:04:30; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB225_L29967-5_Form1A_DB83_183S11_SJ2455699.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 1A
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T03-1808 (Duplicate)
Sample Collection:
22-Aug-2018 17:55

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.:	4972	Project No.	PORTLAND HARBOR PDI AND BASELINE WATER
Matrix:	FILTER	Lab Sample I.D.:	WG65583-103 i2 (DUP L29967-5)
Sample Receipt Date:	29-Aug-2018	Sample Size:	0.2 sample
Extraction Date:	10-Oct-2018	Initial Calibration Date:	24-Oct-2018
Analysis Date:	29-Oct-2018 Time: 17:11:57	Instrument ID:	HR GC/MS
Extract Volume (uL):	20	GC Column ID:	DB5
Injection Volume (uL):	1.0	Sample Data Filename:	DX8M_127A S: 11
Dilution Factor:	N/A	Blank Data Filename:	DX8M_127A S: 5
Concentration Units:	pg/sample	Cal. Ver. Data Filename:	DX8M_127A S: 1

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG ¹	CONCENTRATION FOUND	REPORTING LIMIT (RL) ²	ION ABUND. RATIO ³	RRT ³
2,3,7,8-TCDD	K J	5.77	4.25 (Q)	0.51	1.001
1,2,3,7,8-PECDD ⁴	J	4.72	4.37 (S)	0.67	1.000
1,2,3,4,7,8-HXCDD	J	7.64	4.25 (Q)	1.41	1.000
1,2,3,6,7,8-HXCDD	J	33.0	4.25 (Q)	1.23	1.000
1,2,3,7,8,9-HXCDD	J	23.0	4.25 (Q)	1.07	1.010
1,2,3,4,6,7,8-HPCDD		825	4.25 (Q)	1.01	1.000
OCDD		9440	4.25 (Q)	0.88	1.000
2,3,7,8-TCDF		20.3	4.25 (Q)	0.88	1.001
1,2,3,7,8-PECDF	K J	19.9	4.91 (S)	1.29	1.002
2,3,4,7,8-PECDF	J	12.4	4.63 (S)	1.69	1.000
1,2,3,4,7,8-HXCDF	J	39.8	4.25 (Q)	1.18	1.000
1,2,3,6,7,8-HXCDF	J	14.7	4.25 (Q)	1.37	1.000
1,2,3,7,8,9-HXCDF	U		4.25 (Q)		
2,3,4,6,7,8-HXCDF	J	6.31	4.25 (Q)	1.08	1.000
1,2,3,4,6,7,8-HPCDF		193	4.25 (Q)	1.00	1.000
1,2,3,4,7,8,9-HPCDF	J	12.3	4.25 (Q)	1.19	1.000
OCDF		457	4.25 (Q)	0.84	1.002

(1) Where applicable, custom lab flags have been used on this report; U = not detected at RL; K = peak detected but did not meet quantification criteria, result reported represents the estimated maximum possible concentration; J = concentration less than lowest calibration equivalent.

(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613.

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: Ting Chen

For Axys Internal Use Only [XSL Template: Form1A.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_WG65583-103_Form1A_DX8M_127AS11_SJ2456909.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 2
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T03-1808 (Duplicate)
Sample Collection:
22-Aug-2018 17:55

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.: 4972

Matrix: FILTER

Sample Receipt Date: 29-Aug-2018

Extraction Date: 10-Oct-2018

Analysis Date: 29-Oct-2018 Time: 17:11:57

Extract Volume (uL): 20

Injection Volume (uL): 1.0

Dilution Factor: N/A

Concentration Units: pg absolute

Project No.

Lab Sample I.D.:

Sample Size:

Initial Calibration Date:

Instrument ID:

GC Column ID:

Sample Data Filename:

Blank Data Filename:

Cal. Ver. Data Filename:

PORTLAND HARBOR PDI AND
BASELINE WATER
WG65583-103 i2 (DUP L29967-5)

0.2 sample

24-Oct-2018

HR GC/MS

DB5

DX8M_127A S: 11

DX8M_127A S: 5

DX8M_127A S: 1

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

LABELED COMPOUND	LAB FLAG ¹	SPIKE CONC.	CONC. FOUND	R(%) ²	ION ABUND. RATIO ³	RRT ³
13C-2,3,7,8-TCDD		4000	3590	89.7	0.76	1.013
13C-1,2,3,7,8-PECDD ⁴		4000	3630	90.8	0.63	1.381
13C-1,2,3,4,7,8-HXCDD		4000	3500	87.6	1.26	0.987
13C-1,2,3,6,7,8-HXCDD		4000	3750	93.6	1.25	0.990
13C-1,2,3,4,6,7,8-HPCDD		4000	3480	86.9	1.02	1.094
13C-OCDD		8000	6280	78.5	0.90	1.179
13C-2,3,7,8-TCDF		4000	3500	87.6	0.79	0.966
13C-1,2,3,7,8-PECDF		4000	3650	91.1	1.61	1.282
13C-2,3,4,7,8-PECDF		4000	3340	83.5	1.57	1.349
13C-1,2,3,4,7,8-HXCDF		4000	3110	77.8	0.51	0.954
13C-1,2,3,6,7,8-HXCDF		4000	3340	83.4	0.51	0.958
13C-1,2,3,7,8,9-HXCDF		4000	3100	77.5	0.52	1.005
13C-2,3,4,6,7,8-HXCDF		4000	3180	79.4	0.52	0.981
13C-1,2,3,4,6,7,8-HPCDF		4000	2950	73.7	0.46	1.062
13C-1,2,3,4,7,8,9-HPCDF		4000	3190	79.8	0.43	1.104

CLEANUP STANDARD

37CL-2,3,7,8-TCDD	400	335	83.6	1.001
-------------------	-----	-----	------	-------

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for percent recovery (R) are specified in Section 9.3.3, Method 1613.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613. NOTE: There is no ion abundance ratio for 37Cl4-2,3,7,8-TCDD

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

For Axys Internal Use Only [XSL Template: Form2.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_WG65583-103_Form2_DX8M_127AS11_SJ2456909.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 1A
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T03-1808 (Duplicate)
Sample Collection:
22-Aug-2018 17:55

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.: 4972

Matrix: FILTER

Sample Receipt Date: 29-Aug-2018

Extraction Date: 10-Oct-2018

Analysis Date: 24-Oct-2018 Time: 03:05:32

Extract Volume (uL): 20

Injection Volume (uL): 2.0

Dilution Factor: N/A

Concentration Units: pg/sample

Project No.

Lab Sample I.D.:

Sample Size:

Initial Calibration Date:

Instrument ID:

GC Column ID:

Sample Data Filename:

Blank Data Filename:

Cal. Ver. Data Filename:

PORTLAND HARBOR PDI AND
BASELINE WATER
WG65583-103 (DUP L29967-5)

0.2 sample

10-Jul-2018

HR GC/MS

DB225

DB83_183 S: 12

DX8M_127A S: 5

DB83_183 S: 2

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG ¹	CONCENTRATION FOUND	REPORTING LIMIT (RL) ²	ION ABUND. RATIO ³	RRT ³
1,2,3,7,8,9-HXCDD	K J	8.51	4.25 (Q)	1.67	1.062
2,3,7,8-TCDF	K	17.0	4.25 (Q)	0.53	1.001

(1) Where applicable, custom lab flags have been used on this report; K = peak detected but did not meet quantification criteria, result reported represents the estimated maximum possible concentration; J = concentration less than lowest calibration equivalent.

(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

For Axys Internal Use Only [XSL Template: Form1A.xsl; Created: 21-Nov-2018 17:04:30; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB225_WG65583-103_Form1A_DB83_183S12_SJ2455700.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.: 4972

Client ID: PDI-WS-T03-1808

Project No.

PCDD/PCDF ANALYSIS REPORT
RELATIVE PERCENT DIFFERENCE

PORTLAND HARBOR PDI AND
BASELINE WATER

Concentration Units:

pg/sample

COMPOUND	L29967-5 (A)		WG65583-103		MEAN	RELATIVE PERCENT DIFFERENCE
	LAB FLAG ¹	CONC. FOUND	LAB FLAG ¹	CONC. FOUND		
2,3,7,8-TCDD	U		K J	5.77		
1,2,3,7,8-PECDD	J	4.42	J	4.72	4.57	6.47
1,2,3,4,7,8-HXCDD	J	8.63	J	7.64	8.13	12.2
1,2,3,6,7,8-HXCDD	J	31.7	J	33.0	32.4	4.22
1,2,3,7,8,9-HXCDD	J	21.6	J	23.0	22.3	6.57
1,2,3,4,6,7,8-HPCDD		773		825	799	6.42
OCDD		8820		9440	9130	6.71
2,3,7,8-TCDF		20.0		20.3	20.1	1.78
1,2,3,7,8-PECDF	K J	16.3	K J	19.9		
2,3,4,7,8-PECDF	J	9.88	J	12.4	11.1	22.5
1,2,3,4,7,8-HXCDF	J	39.0	J	39.8	39.4	2.16
1,2,3,6,7,8-HXCDF	J	14.5	J	14.7	14.6	0.911
1,2,3,7,8,9-HXCDF	U		U			
2,3,4,6,7,8-HXCDF	J	6.19	J	6.31	6.25	2.02
1,2,3,4,6,7,8-HPCDF		177		193	185	8.71
1,2,3,4,7,8,9-HPCDF	J	14.2	J	12.3	13.3	13.9
OCDF		483		457	470	5.65

(1) Where applicable, custom lab flags have been used on this report; U = not detected at RL; K = peak detected but did not meet quantification criteria, result reported represents the estimated maximum possible concentration; J = concentration less than lowest calibration equivalent.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: Ting Chen

These pages are part of a larger report that may contain information necessary for full data evaluation. Results reported relate only to the sample tested.

For Axys Internal Use Only [XSL Template: RPD.xml; Created: 21-Nov-2018 17:05:12; Application: XMLTransformer-1.16.51;
Report Filename: RPD_DIOXINS_1613-RPD_WG65583-103_L29967-5_.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 1A
PCDD/PCDF ANALYSIS REPORT

CLIENT SAMPLE NO.
PDI-WS-T01-1808
Sample Collection:
25-Aug-2018 12:02

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.:	4972	Project No.	PORTLAND HARBOR PDI AND BASELINE WATER
Matrix:	FILTER	Lab Sample I.D.:	L29967-6
Sample Receipt Date:	29-Aug-2018	Sample Size:	0.2 sample
Extraction Date:	10-Oct-2018	Initial Calibration Date:	24-Oct-2018
Analysis Date:	25-Oct-2018 Time: 09:22:33	Instrument ID:	HR GC/MS
Extract Volume (uL):	20	GC Column ID:	DB5
Injection Volume (uL):	1.0	Sample Data Filename:	DX8M_126A S: 28
Dilution Factor:	N/A	Blank Data Filename:	DX8M_127A S: 5
Concentration Units:	pg/sample	Cal. Ver. Data Filename:	DX8M_126A S: 20

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG ¹	CONCENTRATION FOUND	REPORTING LIMIT (RL) ²	ION ABUND. RATIO ³	RRT ³
2,3,7,8-TCDD	K J	4.49	4.31 (Q)	0.44	1.002
1,2,3,7,8-PECDD ⁴	U		4.31 (Q)		
1,2,3,4,7,8-HXCDD	J	6.91	4.31 (Q)	1.08	1.000
1,2,3,6,7,8-HXCDD	J	23.5	4.31 (Q)	1.25	1.000
1,2,3,7,8,9-HXCDD	J	14.6	4.31 (Q)	1.17	1.010
1,2,3,4,6,7,8-HPCDD		465	4.31 (Q)	1.01	1.000
OCDD		4770	4.31 (Q)	0.89	1.000
2,3,7,8-TCDF	J	10.7	4.31 (Q)	0.86	1.001
1,2,3,7,8-PECDF	J	8.47	4.31 (Q)	1.43	1.000
2,3,4,7,8-PECDF	K J	6.53	4.31 (Q)	1.23	1.001
1,2,3,4,7,8-HXCDF	J	13.6	4.31 (Q)	1.42	1.000
1,2,3,6,7,8-HXCDF	J	7.05	4.31 (Q)	1.23	1.001
1,2,3,7,8,9-HXCDF	U		4.31 (Q)		
2,3,4,6,7,8-HXCDF	U		4.31 (Q)		
1,2,3,4,6,7,8-HPCDF		125	4.31 (Q)	1.09	1.000
1,2,3,4,7,8,9-HPCDF	J	10.2	4.31 (Q)	1.19	1.000
OCDF		495	4.31 (Q)	0.87	1.002

(1) Where applicable, custom lab flags have been used on this report; U = not detected at RL; K = peak detected but did not meet quantification criteria, result reported represents the estimated maximum possible concentration; J = concentration less than lowest calibration equivalent.

(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613.

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: Ting Chen

For Axys Internal Use Only [XSL Template: Form1A.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_L29967-6_Form1A_DX8M_126AS28_SJ2455652.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 2
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T01-1808
Sample Collection:
25-Aug-2018 12:02

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.: 4972

Matrix: FILTER

Sample Receipt Date: 29-Aug-2018

Extraction Date: 10-Oct-2018

Analysis Date: 25-Oct-2018 Time: 09:22:33

Extract Volume (uL): 20

Injection Volume (uL): 1.0

Dilution Factor: N/A

Concentration Units: pg absolute

Project No. PORTLAND HARBOR PDI AND
BASELINE WATER

Lab Sample I.D.: L29967-6

Sample Size: 0.2 sample

Initial Calibration Date: 24-Oct-2018

Instrument ID: HR GC/MS

GC Column ID: DB5

Sample Data Filename: DX8M_126A S: 28

Blank Data Filename: DX8M_127A S: 5

Cal. Ver. Data Filename: DX8M_126A S: 20

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

LABELED COMPOUND	LAB FLAG ¹	SPIKE CONC.	CONC. FOUND	R(%) ²	ION ABUND. RATIO ³	RRT ³
13C-2,3,7,8-TCDD		4000	3320	83.1	0.79	1.012
13C-1,2,3,7,8-PECDD ⁴		4000	3750	93.7	0.63	1.381
13C-1,2,3,4,7,8-HXCDD		4000	3270	81.7	1.27	0.987
13C-1,2,3,6,7,8-HXCDD		4000	3440	85.9	1.27	0.990
13C-1,2,3,4,6,7,8-HPCDD		4000	3250	81.2	1.03	1.095
13C-OCDD		8000	5190	64.9	0.90	1.179
13C-2,3,7,8-TCDF		4000	3260	81.6	0.77	0.966
13C-1,2,3,7,8-PECDF		4000	3510	87.8	1.58	1.283
13C-2,3,4,7,8-PECDF		4000	3350	83.8	1.60	1.350
13C-1,2,3,4,7,8-HXCDF		4000	3170	79.3	0.51	0.954
13C-1,2,3,6,7,8-HXCDF		4000	3220	80.5	0.51	0.958
13C-1,2,3,7,8,9-HXCDF		4000	3060	76.4	0.51	1.004
13C-2,3,4,6,7,8-HXCDF		4000	3020	75.6	0.51	0.981
13C-1,2,3,4,6,7,8-HPCDF		4000	2910	72.7	0.45	1.062
13C-1,2,3,4,7,8,9-HPCDF		4000	3080	76.9	0.45	1.104

CLEANUP STANDARD

37CL-2,3,7,8-TCDD		400	312	78.0		1.001
-------------------	--	-----	-----	------	--	-------

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for percent recovery (R) are specified in Section 9.3.3, Method 1613.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613. NOTE: There is no ion abundance ratio for 37Cl4-2,3,7,8-TCDD

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

For Axys Internal Use Only [XSL Template: Form2.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_L29967-6_Form2_DX8M_126AS28_SJ2455652.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 1A
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T01-1808
Sample Collection:
25-Aug-2018 12:02

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.: 4972

Matrix: FILTER

Sample Receipt Date: 29-Aug-2018

Extraction Date: 10-Oct-2018

Analysis Date: 24-Oct-2018 Time: 03:46:48

Extract Volume (uL): 20

Injection Volume (uL): 2.0

Dilution Factor: N/A

Concentration Units: pg/sample

Project No.

Lab Sample I.D.:

Sample Size:

Initial Calibration Date:

Instrument ID:

GC Column ID:

Sample Data Filename:

Blank Data Filename:

Cal. Ver. Data Filename:

PORTLAND HARBOR PDI AND
BASELINE WATER
L29967-6

0.2 sample

10-Jul-2018

HR GC/MS

DB225

DB83_183 S: 13

DX8M_127A S: 5

DB83_183 S: 2

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG ¹	CONCENTRATION FOUND	REPORTING LIMIT (RL) ²	ION ABUND. RATIO ³	RRT ³
1,2,3,7,8,9-HXCDD	K J	9.86	4.31 (Q)	1.31	1.062
2,3,7,8-TCDF	J	5.38	4.31 (Q)	0.79	1.000

(1) Where applicable, custom lab flags have been used on this report; K = peak detected but did not meet quantification criteria, result reported represents the estimated maximum possible concentration; J = concentration less than lowest calibration equivalent.

(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

For Axys Internal Use Only [XSL Template: Form1A.xsl; Created: 21-Nov-2018 17:04:30; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB225_L29967-6_Form1A_DB83_183S13_SJ2455701.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 1A
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T06-1808
Sample Collection:
24-Aug-2018 16:32

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.:	4972	Project No.	PORTLAND HARBOR PDI AND BASELINE WATER
Matrix:	FILTER	Lab Sample I.D.:	L29967-7
Sample Receipt Date:	29-Aug-2018	Sample Size:	0.2 sample
Extraction Date:	10-Oct-2018	Initial Calibration Date:	24-Oct-2018
Analysis Date:	25-Oct-2018 Time: 10:17:46	Instrument ID:	HR GC/MS
Extract Volume (uL):	20	GC Column ID:	DB5
Injection Volume (uL):	1.0	Sample Data Filename:	DX8M_126A S: 29
Dilution Factor:	N/A	Blank Data Filename:	DX8M_127A S: 5
Concentration Units:	pg/sample	Cal. Ver. Data Filename:	DX8M_126A S: 20

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG ¹	CONCENTRATION FOUND	REPORTING LIMIT (RL) ²	ION ABUND. RATIO ³	RRT ³
2,3,7,8-TCDD	K J	4.75	4.30 (Q)	0.20	1.002
1,2,3,7,8-PECDD ⁴	K J	4.30	4.30 (Q)	0.81	1.002
1,2,3,4,7,8-HXCDD	K J	6.09	4.30 (Q)	1.45	1.000
1,2,3,6,7,8-HXCDD	J	16.1	4.30 (Q)	1.08	1.000
1,2,3,7,8,9-HXCDD	J	14.1	4.30 (Q)	1.36	1.010
1,2,3,4,6,7,8-HPCDD		321	4.30 (Q)	1.02	1.000
OCDD		3210	4.30 (Q)	0.88	1.000
2,3,7,8-TCDF	J	4.48	4.30 (Q)	0.67	1.002
1,2,3,7,8-PECDF	U		4.30 (Q)		
2,3,4,7,8-PECDF	U		4.30 (Q)		
1,2,3,4,7,8-HXCDF	J	6.14	4.30 (Q)	1.14	1.000
1,2,3,6,7,8-HXCDF	J	4.46	4.30 (Q)	1.38	1.000
1,2,3,7,8,9-HXCDF	U		4.30 (Q)		
2,3,4,6,7,8-HXCDF	J	4.79	4.30 (Q)	1.18	1.001
1,2,3,4,6,7,8-HPCDF	J	73.9	4.30 (Q)	1.02	1.000
1,2,3,4,7,8,9-HPCDF	J	5.99	4.30 (Q)	1.17	1.000
OCDF	J	164	4.30 (Q)	0.93	1.002

(1) Where applicable, custom lab flags have been used on this report; U = not detected at RL; K = peak detected but did not meet quantification criteria, result reported represents the estimated maximum possible concentration; J = concentration less than lowest calibration equivalent.

(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613.

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ Ting Chen _____

For Axys Internal Use Only [XSL Template: Form1A.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_L29967-7_Form1A_DX8M_126AS29_SJ2455653.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 2
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T06-1808
Sample Collection:
24-Aug-2018 16:32

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.:	4972	Project No.	PORTLAND HARBOR PDI AND BASELINE WATER
Matrix:	FILTER	Lab Sample I.D.:	L29967-7
Sample Receipt Date:	29-Aug-2018	Sample Size:	0.2 sample
Extraction Date:	10-Oct-2018	Initial Calibration Date:	24-Oct-2018
Analysis Date:	25-Oct-2018 Time: 10:17:46	Instrument ID:	HR GC/MS
Extract Volume (uL):	20	GC Column ID:	DB5
Injection Volume (uL):	1.0	Sample Data Filename:	DX8M_126A S: 29
Dilution Factor:	N/A	Blank Data Filename:	DX8M_127A S: 5
Concentration Units:	pg absolute	Cal. Ver. Data Filename:	DX8M_126A S: 20

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

LABELLED COMPOUND	LAB FLAG ¹	SPIKE CONC.	CONC. FOUND	R(%) ²	ION ABUND. RATIO ³	RRT ³
13C-2,3,7,8-TCDD		4000	3190	79.6	0.78	1.013
13C-1,2,3,7,8-PECDD ⁴		4000	3350	83.9	0.63	1.382
13C-1,2,3,4,7,8-HXCDD		4000	3300	82.5	1.27	0.987
13C-1,2,3,6,7,8-HXCDD		4000	3510	87.8	1.26	0.990
13C-1,2,3,4,6,7,8-HPCDD		4000	3330	83.3	1.05	1.095
13C-OCDD		8000	5730	71.6	0.91	1.179
13C-2,3,7,8-TCDF		4000	3160	79.0	0.78	0.966
13C-1,2,3,7,8-PECDF		4000	3410	85.2	1.59	1.283
13C-2,3,4,7,8-PECDF		4000	3190	79.8	1.58	1.350
13C-1,2,3,4,7,8-HXCDF		4000	3160	78.9	0.51	0.954
13C-1,2,3,6,7,8-HXCDF		4000	3280	81.9	0.51	0.958
13C-1,2,3,7,8,9-HXCDF		4000	3070	76.6	0.52	1.004
13C-2,3,4,6,7,8-HXCDF		4000	3090	77.3	0.52	0.981
13C-1,2,3,4,6,7,8-HPCDF		4000	3080	77.1	0.45	1.062
13C-1,2,3,4,7,8,9-HPCDF		4000	3080	77.1	0.44	1.104

CLEANUP STANDARD

37CL-2,3,7,8-TCDD		400	308	76.9		1.001
-------------------	--	-----	-----	------	--	-------

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for percent recovery (R) are specified in Section 9.3.3, Method 1613.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613. NOTE: There is no ion abundance ratio for 37Cl4-2,3,7,8-TCDD

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

For Axys Internal Use Only [XSL Template: Form2.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_L29967-7_Form2_DX8M_126AS29_SJ2455653.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 1A
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
PDI-WS-T06-1808
Sample Collection:
24-Aug-2018 16:32

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.: 4972

Matrix: FILTER

Sample Receipt Date: 29-Aug-2018

Extraction Date: 10-Oct-2018

Analysis Date: 24-Oct-2018 Time: 04:28:00

Extract Volume (uL): 20

Injection Volume (uL): 2.0

Dilution Factor: N/A

Concentration Units: pg/sample

Project No.

Lab Sample I.D.:

Sample Size:

Initial Calibration Date:

Instrument ID:

GC Column ID:

Sample Data Filename:

Blank Data Filename:

Cal. Ver. Data Filename:

PORTLAND HARBOR PDI AND
BASELINE WATER
L29967-7

0.2 sample

10-Jul-2018

HR GC/MS

DB225

DB83_183 S: 14

DX8M_127A S: 5

DB83_183 S: 2

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG ¹	CONCENTRATION FOUND	REPORTING LIMIT (RL) ²	ION ABUND. RATIO ³	RRT ³
1,2,3,7,8,9-HXCDD	K J	11.8	4.30 (Q)	1.14	1.063
2,3,7,8-TCDF	U		4.30 (Q)		

(1) Where applicable, custom lab flags have been used on this report; U = not detected at RL; K = peak detected but did not meet quantification criteria, result reported represents the estimated maximum possible concentration; J = concentration less than lowest calibration equivalent.

(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

For Axys Internal Use Only [XSL Template: Form1A.xsl; Created: 21-Nov-2018 17:04:30; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB225_L29967-7_Form1A_DB83_183S14_SJ2455702.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 1A
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
Lab Blank
Sample Collection:
N/A

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.: 4972

Matrix: FILTER

Sample Receipt Date: N/A

Extraction Date: 10-Oct-2018

Analysis Date: 29-Oct-2018 Time: 11:40:46

Extract Volume (uL): 20

Injection Volume (uL): 1.0

Dilution Factor: N/A

Concentration Units: pg/sample

Project No.

N/A

Lab Sample I.D.:

WG65583-101 i2

Sample Size:

0.2 sample

Initial Calibration Date:

24-Oct-2018

Instrument ID:

HR GC/MS

GC Column ID:

DB5

Sample Data Filename:

DX8M_127A S: 5

Blank Data Filename:

DX8M_127A S: 5

Cal. Ver. Data Filename:

DX8M_127A S: 1

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

COMPOUND	LAB FLAG ¹	CONCENTRATION FOUND	REPORTING LIMIT (RL) ²	ION ABUND. RATIO ³	RRT ³
2,3,7,8-TCDD	U		4.28 (Q)		
1,2,3,7,8-PECDD ⁴	U		4.28 (Q)		
1,2,3,4,7,8-HXCDD	U		4.28 (Q)		
1,2,3,6,7,8-HXCDD	U		4.28 (Q)		
1,2,3,7,8,9-HXCDD	U		4.28 (Q)		
1,2,3,4,6,7,8-HPCDD	U		4.28 (Q)		
OCDD	K J	7.93	4.28 (Q)	1.05	1.000
2,3,7,8-TCDF	U		4.28 (Q)		
1,2,3,7,8-PECDF	U		4.28 (Q)		
2,3,4,7,8-PECDF	U		4.28 (Q)		
1,2,3,4,7,8-HXCDF	U		4.28 (Q)		
1,2,3,6,7,8-HXCDF	U		4.28 (Q)		
1,2,3,7,8,9-HXCDF	U		4.28 (Q)		
2,3,4,6,7,8-HXCDF	U		4.28 (Q)		
1,2,3,4,6,7,8-HPCDF	U		4.28 (Q)		
1,2,3,4,7,8,9-HPCDF	U		4.28 (Q)		
OCDF	U		4.28 (Q)		

(1) Where applicable, custom lab flags have been used on this report; U = not detected at RL; K = peak detected but did not meet quantification criteria, result reported represents the estimated maximum possible concentration; J = concentration less than lowest calibration equivalent.

(2) Reporting Limit (Code): S = sample detection limit; M = method detection limit; L = lowest calibration level equivalent; Q = minimum reporting level.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613.

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

For Axy Internal Use Only [XSL Template: Form1A.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_WG65583-101_Form1A_DX8M_127AS5_SJ2456917.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 2
PCDD/PCDF ANALYSIS REPORTCLIENT SAMPLE NO.
Lab Blank
Sample Collection:
N/A

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.: 4972

Project No.

N/A

Lab Sample I.D.:

WG65583-101 i2

Matrix: FILTER

Sample Size:

0.2 sample

Sample Receipt Date: N/A

Initial Calibration Date:

24-Oct-2018

Extraction Date: 10-Oct-2018

Instrument ID:

HR GC/MS

Analysis Date: 29-Oct-2018 Time: 11:40:46

GC Column ID:

DB5

Extract Volume (uL): 20

Sample Data Filename:

DX8M_127A S: 5

Injection Volume (uL): 1.0

Blank Data Filename:

DX8M_127A S: 5

Dilution Factor: N/A

Cal. Ver. Data Filename:

DX8M_127A S: 1

Concentration Units: pg absolute

This page is part of a total report that contains information necessary for accreditation compliance.
This test is not NELAP accredited. Sample results relate only to the sample tested.

LABELED COMPOUND	LAB FLAG ¹	SPIKE CONC.	CONC. FOUND	R(%) ²	ION ABUND. RATIO ³	RRT ³
13C-2,3,7,8-TCDD		4000	3470	86.7	0.77	1.013
13C-1,2,3,7,8-PECDD ⁴		4000	3480	87.0	0.63	1.382
13C-1,2,3,4,7,8-HXCDD		4000	3400	85.0	1.26	0.987
13C-1,2,3,6,7,8-HXCDD		4000	3460	86.6	1.27	0.990
13C-1,2,3,4,6,7,8-HPCDD		4000	3340	83.5	1.03	1.094
13C-OCDD		8000	5770	72.2	0.91	1.179
13C-2,3,7,8-TCDF		4000	3360	84.1	0.79	0.966
13C-1,2,3,7,8-PECDF		4000	3490	87.2	1.56	1.283
13C-2,3,4,7,8-PECDF		4000	3320	83.0	1.56	1.351
13C-1,2,3,4,7,8-HXCDF		4000	3240	81.1	0.51	0.954
13C-1,2,3,6,7,8-HXCDF		4000	3350	83.8	0.52	0.958
13C-1,2,3,7,8,9-HXCDF		4000	3110	77.9	0.52	1.004
13C-2,3,4,6,7,8-HXCDF		4000	3170	79.3	0.52	0.980
13C-1,2,3,4,6,7,8-HPCDF		4000	3050	76.2	0.45	1.062
13C-1,2,3,4,7,8,9-HPCDF		4000	3150	78.8	0.45	1.104

CLEANUP STANDARD

37CL-2,3,7,8-TCDD		400	322	80.6		1.001
-------------------	--	-----	-----	------	--	-------

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for percent recovery (R) are specified in Section 9.3.3, Method 1613.

(3) Contract-required limits for RRTs and ion abundance ratios are specified in Tables 2 and 9, respectively, Method 1613. NOTE: There is no ion abundance ratio for 37Cl4-2,3,7,8-TCDD

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

For Axs Internal Use Only [XSL Template: Form2.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_WG65583-101_Form2_DX8M_127AS5_SJ2456917.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 8A

PCDD/PCDF ONGOING PRECISION AND RECOVERY (OPR)

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.: 4972

OPR Data Filename:

DX8M_127A S: 2

Matrix: FILTER

Lab Sample I.D.:

WG65583-102 i2

Extraction Date: 10-Oct-2018

Analysis Date:

29-Oct-2018 Time: 08:57:47

ALL CONCENTRATIONS REPORTED ON THIS FORM ARE CONCENTRATIONS IN EXTRACT, BASED ON A 20 µL EXTRACT VOLUME.

COMPOUND	LAB FLAG ¹	ION ABUND. RATIO ²	SPIKE CONC. (ng/mL)	CONC. FOUND (ng/mL)	OPR CONC. LIMITS ³ (ng/mL)	% RECOVERY
2,3,7,8-TCDD		0.79	20.0	19.1	13.4 - 31.6	95.5
1,2,3,7,8-PECDD ⁴		0.62	100	98.1	70.0 - 142	98.1
1,2,3,4,7,8-HXCDD		1.24	100	98.3	70.0 - 164	98.3
1,2,3,6,7,8-HXCDD		1.24	100	99.1	76.0 - 134	99.1
1,2,3,7,8,9-HXCDD		1.23	100	96.8	64.0 - 162	96.8
1,2,3,4,6,7,8-HPCDD		1.07	100	99.9	70.0 - 140	99.9
OCDD		0.88	200	193	156 - 288	96.5
2,3,7,8-TCDF		0.78	20.0	19.2	15.0 - 31.6	96.2
1,2,3,7,8-PECDF		1.56	100	94.5	80.0 - 134	94.5
2,3,4,7,8-PECDF		1.56	100	97.9	68.0 - 160	97.9
1,2,3,4,7,8-HXCDF		1.24	100	96.6	72.0 - 134	96.6
1,2,3,6,7,8-HXCDF		1.25	100	97.4	84.0 - 130	97.4
1,2,3,7,8,9-HXCDF		1.25	100	96.1	78.0 - 130	96.1
2,3,4,6,7,8-HXCDF		1.24	100	98.9	70.0 - 156	98.9
1,2,3,4,6,7,8-HPCDF		1.04	100	105	82.0 - 122	105
1,2,3,4,7,8,9-HPCDF		1.03	100	102	78.0 - 138	102
OCDF		0.91	200	189	126 - 340	94.7

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required Ion Abundance Ratios are specified in Table 9, Method 1613.

(3) Contract-required concentration range as determined from the percent of the test concentration in Table 6, Method 1613, under OPR.

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

These pages are part of a larger report that may contain information necessary for full data evaluation. Results reported relate only to the sample tested.

For Axys Internal Use Only [XSL Template: Form8A.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_WG65583-102_Form8A_SJ2456913.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 8B

PCDD/PCDF ONGOING PRECISION AND RECOVERY (OPR)

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Contract No.: 4972

OPR Data Filename:

DX8M_127A S: 2

Matrix: FILTER

Lab Sample I.D.:

WG65583-102 i2

Extraction Date: 10-Oct-2018

Analysis Date:

29-Oct-2018 Time: 08:57:47

ALL CONCENTRATIONS REPORTED ON THIS FORM ARE CONCENTRATIONS IN EXTRACT, BASED ON A 20 µL EXTRACT VOLUME.

LABELLED COMPOUND	LAB FLAG ¹	ION ABUND. RATIO ²	SPIKE CONC. (ng/mL)	CONC. FOUND (ng/mL)	OPR CONC. LIMITS ³ (ng/mL)	% RECOVERY
13C-2,3,7,8-TCDD		0.79	200	169	40.0-350	84.6
13C-1,2,3,7,8-PECDD ⁴		0.63	200	167	42.0-454	83.5
13C-1,2,3,4,7,8-HXCDD		1.26	200	170	42.0-386	84.9
13C-1,2,3,6,7,8-HXCDD		1.25	200	177	50.0-326	88.5
13C-1,2,3,4,6,7,8-HPCDD		1.06	200	162	52.0-332	80.9
13C-OCDD		0.89	400	294	52.0-794	73.4
13C-2,3,7,8-TCDF		0.78	200	168	44.0-304	84.2
13C-1,2,3,7,8-PECDF		1.57	200	168	42.0-384	84.0
13C-2,3,4,7,8-PECDF		1.59	200	160	26.0-656	79.8
13C-1,2,3,4,7,8-HXCDF		0.52	200	163	38.0-404	81.7
13C-1,2,3,6,7,8-HXCDF		0.51	200	168	42.0-318	84.2
13C-1,2,3,7,8,9-HXCDF		0.53	200	158	34.0-410	79.2
13C-2,3,4,6,7,8-HXCDF		0.52	200	160	44.0-352	80.1
13C-1,2,3,4,6,7,8-HPCDF		0.45	200	153	42.0-316	76.5
13C-1,2,3,4,7,8,9-HPCDF		0.45	200	157	40.0-372	78.4

CLEANUP STANDARD

37CL-2,3,7,8-TCDD			20.0	15.7	6.20-38.2	78.6
-------------------	--	--	------	------	-----------	------

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required Ion Abundance Ratios are specified in Table 9, Method 1613.

(3) Contract-required concentration limits for OPR as specified in Table 6, Method 1613. Labeled compound concentrations limits are based on required percent recovery (Section 15.5, Method 1613).

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

These pages are part of a larger report that may contain information necessary for full data evaluation. Results reported relate only to the sample tested.

For Axys Internal Use Only [XSL Template: Form8B.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_1613DB5_WG65583-102_Form8B_SJ2456913.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 3A

PCDD/PCDF INITIAL CALIBRATION RELATIVE RESPONSES

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
 V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 10-Jul-2018

Instrument ID: HR GC/MS

GC Column ID: DB225

CS0 Data Filename: DB83_121 S: 4

CS1 Data Filename: DB83_121 S: 3

CS2 Data Filename: DB83_121 S: 2

CS3 Data Filename: DB83_121 S: 7

CS4 Data Filename: DB83_121 S: 6

CS5 Data Filename: DB83_121 S: 5

CS6 Data Filename: N/A

COMPOUND	LAB FLAG ¹	RELATIVE RESPONSE (RR)						MEAN RR	CV (%RSD) ²
		CS0	CS1	CS2	CS3	CS4	CS5		
1,2,3,7,8,9-HxCDD		0.82	0.89	0.85	0.97	0.99	1.00	0.92	8.53
2,3,7,8-TCDF		0.93	0.91	0.89	1.04	1.03	1.01	0.97	7.11

(1) Where applicable, custom lab flags have been used on this report.

(2) For contract CV specifications, see Section 10.5.4, Method 1613.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ Bjorn Arvi _____

For Axs Internal Use Only [XSL Template: Form3A.xsl; Created: 21-Nov-2018 17:04:30; Application: XMLTransformer-1.16.51;
 Report Filename: 1613_DIOXINS_10-Jul-2018_DB83__Form3A_GS78423.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 3C

PCDD/PCDF INITIAL CALIBRATION ION ABUNDANCE RATIOS

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 10-Jul-2018

Instrument ID: HR GC/MS

GC Column ID: DB225

CS0 Data Filename: DB83_121 S: 4

CS1 Data Filename: DB83_121 S: 3

CS2 Data Filename: DB83_121 S: 2

CS3 Data Filename: DB83_121 S: 7

CS4 Data Filename: DB83_121 S: 6

CS5 Data Filename: DB83_121 S: 5

CS6 Data Filename: N/A

COMPOUND	LAB FLAG ¹	M/Z's FORMING RATIO ²	ION ABUNDANCE RATIO						QC LIMITS ³
			CS0	CS1	CS2	CS3	CS4	CS5	
1,2,3,7,8,9-HxCDD		M+2/M+4	1.25	1.29	1.26	1.26	1.25	1.25	1.05-1.43
2,3,7,8-TCDF		M/M+2	0.88	0.81	0.78	0.78	0.78	0.78	0.65-0.89

(1) Where applicable, custom lab flags have been used on this report.

(2) See Table 8, Method 1613, for m/z specifications.

(3) Ion Abundance Ratio Control Limits from Table 9, Method 1613.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ Bjorn Arvi _____

For Axy Internal Use Only [XSL Template: Form3C.xsl; Created: 21-Nov-2018 17:04:30; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_10-Jul-2018_DB83__Form3C_GS78423.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 5

PCDD/PCDF RT WINDOW AND ISOMER SPECIFICITY STANDARDS

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Instrument ID:	HR GC/MS	Initial Calibration Date:	10-Jul-2018
RT Window Data Filename:		Analysis Date:	Time:
DB-5 IS Data Filename:		Analysis Date:	Time:
DB-225 IS Data Filename:	DB83_121 S: 1	Analysis Date:	10-Jul-2018
		Time:	15:40:53

DB225 RT WINDOW DEFINING STANDARDS RESULT

ISOMERS	ABSOLUTE RT	ISOMERS	ABSOLUTE RT
1,3,6,8-TCDD (F)	N/A	1,3,6,8-TCDF (F)	N/A
1,2,8,9-TCDD (L)	N/A	1,2,8,9-TCDF (L)	N/A
1,2,4,7,9-PECDD (F)	N/A	1,3,4,6,8-PECDF (F)	N/A
1,2,3,8,9-PECDD (L)	N/A	1,2,3,8,9-PECDF (L)	N/A
1,2,4,6,7,9-HXCDD (F)	N/A	1,2,3,4,6,8-HXCDF (F)	N/A
1,2,3,4,6,7-HXCDD (L)	N/A	1,2,3,4,8,9-HXCDF (L)	N/A
1,2,3,4,6,7,9-HPCDD (F)	N/A	1,2,3,4,6,7,8-HPCDF (F)	N/A
1,2,3,4,6,7,8-HPCDD (L)	N/A	1,2,3,4,7,8,9-HPCDF (L)	N/A

(F) = First eluting isomer (DB-5); (L) = Last eluting isomer (DB-5)

ISOMER SPECIFICITY (IS) TEST STANDARDS RESULT

Isomers	% Valley Height Between Compared Peaks	Isomers	% Valley Height Between Compared Peaks
1,2,3,4-TCDD 1,2,7,8-TCDD	N/A	1,2,3,8-TCDD 2,3,7,8-TCDD	N/A
1,2,7,8-TCDD 1,4,7,8-TCDD	N/A	2,3,4,7-TCDF 2,3,7,8-TCDF	3
1,4,7,8-TCDD 1,2,3,7-TCDD	N/A	2,3,7,8-TCDF 1,2,3,9-TCDF	5
1,2,3,7-TCDD 1,2,3,8-TCDD	N/A		

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ **Brian Watson** _____

For Axy Internal Use Only [XSL Template: DXForm5.xsl; Created: 21-Nov-2018 17:04:30; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS__DB83_121S1_Form5_SJ2403414.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 3A

PCDD/PCDF INITIAL CALIBRATION RELATIVE RESPONSES

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

Instrument ID: HR GC/MS

GC Column ID: DB5

CS0 Data Filename: DX8M_126A S: 1

CS1 Data Filename: DX8M_126A S: 2

CS2 Data Filename: DX8M_126A S: 3

CS3 Data Filename: DX8M_126A S: 6

CS4 Data Filename: DX8M_126A S: 5

CS5 Data Filename: DX8M_126A S: 4

CS6 Data Filename: N/A

COMPOUND	LAB FLAG ¹	RELATIVE RESPONSE (RR)						MEAN RR	CV (%RSD) ²
		CS0	CS1	CS2	CS3	CS4	CS5		
2,3,7,8-TCDD		1.05	1.07	0.96	1.04	1.05	1.07	1.04	3.91
1,2,3,7,8-PECDD ³		0.97	1.00	0.96	1.03	1.03	1.04	1.01	3.58
1,2,3,4,7,8-HXCDD		0.96	0.99	0.96	1.02	1.03	1.03	1.00	3.42
1,2,3,6,7,8-HXCDD		0.89	0.92	0.90	0.96	0.97	0.96	0.93	3.98
1,2,3,7,8,9-HXCDD ⁴		0.91	0.96	0.90	0.97	0.98	0.96	0.95	3.48
1,2,3,4,6,7,8-HPCDD		0.85	0.94	0.95	1.01	1.01	1.01	0.96	6.30
OCDD		0.94	0.98	0.97	1.04	1.04	1.01	1.00	4.06
2,3,7,8-TCDF		0.97	0.91	0.88	0.96	0.96	0.95	0.94	3.63
1,2,3,7,8-PECDF		0.91	0.93	0.87	0.98	0.95	0.96	0.93	4.16
2,3,4,7,8-PECDF		0.89	0.92	0.90	0.96	0.98	0.98	0.94	4.30
1,2,3,4,7,8-HXCDF		1.09	1.11	1.09	1.19	1.18	1.16	1.14	4.13
1,2,3,6,7,8-HXCDF		1.04	1.06	1.05	1.10	1.14	1.12	1.09	3.79
1,2,3,7,8,9-HXCDF		1.01	1.05	0.98	1.04	1.06	1.05	1.03	2.98
2,3,4,6,7,8-HXCDF		1.11	1.15	1.11	1.19	1.22	1.20	1.16	4.23
1,2,3,4,6,7,8-HPCDF		1.12	1.21	1.16	1.29	1.30	1.28	1.23	6.06
1,2,3,4,7,8,9-HPCDF		1.18	1.20	1.19	1.24	1.30	1.29	1.23	4.08
OCDF ⁵		1.06	1.14	1.14	1.26	1.30	1.28	1.20	7.95

(1) Where applicable, custom lab flags have been used on this report.

(2) For contract CV specifications, see Section 10.5.4, Method 1613.

(3) Alternate confirmation and quantitation ions used for native and labeled PECDD.

(4) Response ratios are calculated relative to the labeled analogs of the other two HXCDDs (Section 17.1.2, Method 1613).

(5) Response ratios are calculated relative to the labeled analog of OCDD (Section 17.1.1, Method 1613).

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: Thong Do

For Axys Internal Use Only [XSL Template: Form3A.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_24-Oct-2018_DX8M_Form3A_GS78420.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 3B

PCDD/PCDF INITIAL CALIBRATION RELATIVE RESPONSES

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

Instrument ID: HR GC/MS

GC Column ID: DB5

CS0 Data Filename: DX8M_126A S: 1

CS1 Data Filename: DX8M_126A S: 2

CS2 Data Filename: DX8M_126A S: 3

CS3 Data Filename: DX8M_126A S: 6

CS4 Data Filename: DX8M_126A S: 5

CS5 Data Filename: DX8M_126A S: 4

CS6 Data Filename: N/A

LABELED COMPOUND	LAB FLAG ¹	RELATIVE RESPONSE (RR)						MEAN RR	CV (%RSD) ²
		CS0	CS1	CS2	CS3	CS4	CS5		
13C-2,3,7,8-TCDD		0.97	0.93	0.97	0.96	0.97	1.00	0.97	2.19
13C-1,2,3,7,8-PECDD ³		0.71	0.68	0.71	0.70	0.74	0.79	0.72	5.52
13C-1,2,3,4,7,8-HXCDD		1.01	0.94	0.98	0.98	0.99	0.99	0.98	2.38
13C-1,2,3,6,7,8-HXCDD		1.09	1.01	1.04	1.05	1.05	1.07	1.05	2.73
13C-1,2,3,4,6,7,8-HPCDD		0.87	0.80	0.82	0.78	0.86	0.85	0.83	4.24
13C-OCDD		0.74	0.67	0.67	0.67	0.74	0.79	0.71	7.28
13C-2,3,7,8-TCDF		1.54	1.49	1.53	1.51	1.52	1.55	1.52	1.54
13C-1,2,3,7,8-PECDF		1.22	1.19	1.23	1.21	1.29	1.36	1.25	5.26
13C-2,3,4,7,8-PECDF		1.21	1.18	1.22	1.20	1.28	1.36	1.24	5.32
13C-1,2,3,4,7,8-HXCDF		1.32	1.24	1.28	1.28	1.26	1.27	1.27	2.10
13C-1,2,3,6,7,8-HXCDF		1.42	1.30	1.36	1.38	1.37	1.39	1.37	2.82
13C-1,2,3,7,8,9-HXCDF		1.15	1.08	1.13	1.14	1.16	1.17	1.14	2.82
13C-2,3,4,6,7,8-HXCDF		1.31	1.20	1.25	1.23	1.23	1.23	1.24	2.81
13C-1,2,3,4,6,7,8-HPCDF		1.09	1.01	1.05	1.02	1.07	1.06	1.05	2.79
13C-1,2,3,4,7,8,9-HPCDF		0.83	0.77	0.80	0.76	0.81	0.84	0.80	3.64
CLEANUP STANDARD									
37CL-2,3,7,8-TCDD		1.28	1.01	1.06	1.04	1.03	1.12	1.09	9.20

(1) Where applicable, custom lab flags have been used on this report.

(2) For contract CV specifications, see Section 10.5.4, Method 1613.

(3) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: Thong Do

For Axs Internal Use Only [XSL Template: Form3B.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_24-Oct-2018_DX8M_Form3B_GS78420.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 3C

PCDD/PCDF INITIAL CALIBRATION ION ABUNDANCE RATIOS

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

Instrument ID: HR GC/MS

GC Column ID: DB5

CS0 Data Filename: DX8M_126A S: 1

CS1 Data Filename: DX8M_126A S: 2

CS2 Data Filename: DX8M_126A S: 3

CS3 Data Filename: DX8M_126A S: 6

CS4 Data Filename: DX8M_126A S: 5

CS5 Data Filename: DX8M_126A S: 4

CS6 Data Filename: N/A

COMPOUND	LAB FLAG ¹	M/Z's FORMING RATIO ²	ION ABUNDANCE RATIO						QC LIMITS ³
			CS0	CS1	CS2	CS3	CS4	CS5	
2,3,7,8-TCDD		M/M+2	0.85	0.76	0.84	0.77	0.78	0.78	0.65-0.89
1,2,3,7,8-PECDD ⁴		M/M+2	0.61	0.62	0.60	0.62	0.62	0.62	0.52-0.70
1,2,3,4,7,8-HXCDD		M+2/M+4	1.20	1.22	1.23	1.25	1.25	1.25	1.05-1.43
1,2,3,6,7,8-HXCDD		M+2/M+4	1.25	1.19	1.25	1.25	1.24	1.25	1.05-1.43
1,2,3,7,8,9-HXCDD		M+2/M+4	1.17	1.21	1.25	1.25	1.24	1.24	1.05-1.43
1,2,3,4,6,7,8-HPCDD		M+2/M+4	1.12	1.03	1.06	1.03	1.03	1.04	0.88-1.20
OCDD		M+2/M+4	0.91	0.90	0.88	0.90	0.88	0.88	0.76-1.02
2,3,7,8-TCDF		M/M+2	0.73	0.80	0.78	0.76	0.78	0.78	0.65-0.89
1,2,3,7,8-PECDF		M+2/M+4	1.71	1.56	1.55	1.57	1.58	1.56	1.32-1.78
2,3,4,7,8-PECDF		M+2/M+4	1.62	1.53	1.55	1.57	1.56	1.55	1.32-1.78
1,2,3,4,7,8-HXCDF		M+2/M+4	1.23	1.26	1.27	1.26	1.24	1.23	1.05-1.43
1,2,3,6,7,8-HXCDF		M+2/M+4	1.26	1.27	1.27	1.26	1.23	1.24	1.05-1.43
1,2,3,7,8,9-HXCDF		M+2/M+4	1.20	1.24	1.26	1.25	1.25	1.25	1.05-1.43
2,3,4,6,7,8-HXCDF		M+2/M+4	1.17	1.25	1.24	1.26	1.24	1.24	1.05-1.43
1,2,3,4,6,7,8-HPCDF		M+2/M+4	1.01	1.00	1.04	1.05	1.03	1.04	0.88-1.20
1,2,3,4,7,8,9-HPCDF		M+2/M+4	0.97	1.03	1.01	1.02	1.05	1.03	0.88-1.20
OCDF		M+2/M+4	0.90	0.93	0.91	0.90	0.90	0.90	0.76-1.02

(1) Where applicable, custom lab flags have been used on this report.

(2) See Table 8, Method 1613, for m/z specifications.

(3) Ion Abundance Ratio Control Limits from Table 9, Method 1613.

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ Thong Do _____

For Axy Internal Use Only [XSL Template: Form3C.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_24-Oct-2018_DX8M_Form3C_GS78420.html; Workgroup: WG65583; Design ID: 3359]

Form 3D
PCDD/PCDF INITIAL CALIBRATION ION ABUNDANCE RATIOS

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
 V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811
Initial Calibration Date: 24-Oct-2018

Instrument ID: HR GC/MS

GC Column ID: DB5

CS0 Data Filename: DX8M_126A S: 1

CS1 Data Filename: DX8M_126A S: 2

CS2 Data Filename: DX8M_126A S: 3

CS3 Data Filename: DX8M_126A S: 6

CS4 Data Filename: DX8M_126A S: 5

CS5 Data Filename: DX8M_126A S: 4

CS6 Data Filename: N/A

LABELED COMPOUND	LAB FLAG ¹	M/Z's FORMING RATIO ²	ION ABUNDANCE RATIO						QC LIMITS ³
			CS0	CS1	CS2	CS3	CS4	CS5	
13C-2,3,7,8-TCDD		M/M+2	0.77	0.78	0.76	0.76	0.78	0.77	0.65-0.89
13C-1,2,3,7,8-PECDD ⁴		M/M+2	0.63	0.63	0.63	0.64	0.63	0.63	0.52-0.70
13C-1,2,3,4,7,8-HXCDD		M+2/M+4	1.30	1.28	1.29	1.27	1.27	1.29	1.05-1.43
13C-1,2,3,6,7,8-HXCDD		M+2/M+4	1.28	1.26	1.27	1.25	1.25	1.25	1.05-1.43
13C-1,2,3,4,6,7,8-HPCDD		M+2/M+4	1.04	1.05	1.06	1.04	1.03	1.06	0.88-1.20
13C-OCDD		M+2/M+4	0.89	0.89	0.91	0.90	0.89	0.90	0.76-1.02
13C-2,3,7,8-TCDF		M/M+2	0.79	0.78	0.77	0.79	0.79	0.78	0.65-0.89
13C-1,2,3,7,8-PECDF		M+2/M+4	1.60	1.60	1.57	1.60	1.60	1.59	1.32-1.78
13C-2,3,4,7,8-PECDF		M+2/M+4	1.57	1.58	1.60	1.58	1.59	1.57	1.32-1.78
13C-1,2,3,4,7,8-HXCDF		M/M+2	0.51	0.51	0.51	0.52	0.51	0.52	0.43-0.59
13C-1,2,3,6,7,8-HXCDF		M/M+2	0.51	0.52	0.52	0.52	0.52	0.52	0.43-0.59
13C-1,2,3,7,8,9-HXCDF		M/M+2	0.52	0.53	0.51	0.51	0.52	0.52	0.43-0.59
13C-2,3,4,6,7,8-HXCDF		M/M+2	0.51	0.52	0.52	0.51	0.52	0.52	0.43-0.59
13C-1,2,3,4,6,7,8-HPCDF		M/M+2	0.44	0.44	0.45	0.44	0.45	0.45	0.37-0.51
13C-1,2,3,4,7,8,9-HPCDF		M/M+2	0.45	0.45	0.44	0.44	0.45	0.43	0.37-0.51

(1) Where applicable, custom lab flags have been used on this report.

(2) See Table 8, Method 1613, for m/z specifications.

(3) Ion Abundance Ratio Control Limits from Table 9, Method 1613.

(4) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ **Thong Do** _____

For Axy Internal Use Only [XSL Template: Form3D.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
 Report Filename: 1613_DIOXINS_24-Oct-2018_DX8M_Form3D_GS78420.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 4A
PCDD/PCDF CALIBRATION VERIFICATION

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_126A S: 20

Instrument ID: HR GC/MS

Analysis Date: 25-Oct-2018

GC Column ID: DB5

Analysis Time: 01:54:19

COMPOUND	LAB FLAG ¹	MZ's FORMING RATIO ²	ION ABUND. RATIO	QC LIMITS ³	CONC. FOUND (ng/mL)	CONC. RANGE (ng/mL) ⁴
2,3,7,8-TCDD		M/M+2	0.77	0.65-0.89	10.1	7.8 - 12.9
1,2,3,7,8-PECDD ⁵		M/M+2	0.61	0.52-0.70	50.9	39 - 65
1,2,3,4,7,8-HXCDD		M+2/M+4	1.25	1.05-1.43	50.4	39 - 64
1,2,3,6,7,8-HXCDD		M+2/M+4	1.23	1.05-1.43	52.1	39 - 64
1,2,3,7,8,9-HXCDD		M+2/M+4	1.24	1.05-1.43	51.8	41 - 61
1,2,3,4,6,7,8-HPCDD		M+2/M+4	1.06	0.88-1.20	52.4	43 - 58
OCDD		M+2/M+4	0.87	0.76-1.02	101	79 - 126
2,3,7,8-TCDF		M/M+2	0.77	0.65-0.89	10.1	8.4 - 12
1,2,3,7,8-PECDF		M+2/M+4	1.57	1.32-1.78	51.2	41 - 60
2,3,4,7,8-PECDF		M+2/M+4	1.55	1.32-1.78	51.1	41 - 61
1,2,3,4,7,8-HXCDF		M+2/M+4	1.24	1.05-1.43	52.5	45 - 56
1,2,3,6,7,8-HXCDF		M+2/M+4	1.25	1.05-1.43	51.5	44 - 57
1,2,3,7,8,9-HXCDF		M+2/M+4	1.24	1.05-1.43	50.2	45 - 56
2,3,4,6,7,8-HXCDF		M+2/M+4	1.25	1.05-1.43	51.3	44 - 57
1,2,3,4,6,7,8-HPCDF		M+2/M+4	1.03	0.88-1.20	51.5	45 - 55
1,2,3,4,7,8,9-HPCDF		M+2/M+4	1.04	0.88-1.20	50.9	43 - 58
OCDF		M+2/M+4	0.90	0.76-1.02	103	63 - 159

(1) Where applicable, custom lab flags have been used on this report.

(2) See Table 8, Method 1613, for m/z specifications.

(3) Ion Abundance Ratio Control Limits as specified in Table 9, Method 1613.

(4) Contract-required concentration range as determined from the percent of the test concentration in Table 6, Method 1613, under VER.

(5) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Thong Do_____

For Axy Internal Use Only [XSL Template: Form4A.xsl; Created: 22-Nov-2018 12:38:02; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DX8M_126AS20__Form4A_SJ2455651.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 4B
PCDD/PCDF CALIBRATION VERIFICATION

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_126A S: 20

Instrument ID: HR GC/MS

Analysis Date: 25-Oct-2018

GC Column ID: DB5

Analysis Time: 01:54:19

LAB FLAG ¹	MZ's FORMING RATIO ²	ION ABUND. RATIO	QC LIMITS ³	CONC. FOUND (ng/mL)	CONC. RANGE (ng/mL) ⁴
Labeled Compound					
13C-2,3,7,8-TCDD	M/M+2	0.77	0.65-0.89	100	82 - 121
13C-1,2,3,7,8-PECDD ⁵	M/M+2	0.63	0.52-0.70	98.7	62 - 160
13C-1,2,3,4,7,8-HXCDD	M+2/M+4	1.27	1.05-1.43	99.2	85 - 117
13C-1,2,3,6,7,8-HXCDD	M+2/M+4	1.28	1.05-1.43	97.8	85 - 118
13C-1,2,3,4,6,7,8-HPCDD	M+2/M+4	1.04	0.88-1.20	97.2	72 - 138
13C-OCDD	M+2/M+4	0.88	0.76-1.02	187	96 - 415
13C-2,3,7,8-TCDF	M/M+2	0.78	0.65-0.89	99.7	71 - 140
13C-1,2,3,7,8-PECDF	M+2/M+4	1.61	1.32-1.78	98.0	76 - 130
13C-2,3,4,7,8-PECDF	M+2/M+4	1.59	1.32-1.78	98.0	77 - 130
13C-1,2,3,4,7,8-HXCDF	M/M+2	0.51	0.43-0.59	96.5	76 - 131
13C-1,2,3,6,7,8-HXCDF	M/M+2	0.52	0.43-0.59	96.6	70 - 143
13C-1,2,3,7,8,9-HXCDF	M/M+2	0.51	0.43-0.59	98.7	74 - 135
13C-2,3,4,6,7,8-HXCDF	M/M+2	0.51	0.43-0.59	97.3	73 - 137
13C-1,2,3,4,6,7,8-HPCDF	M/M+2	0.45	0.37-0.51	97.7	78 - 129
13C-1,2,3,4,7,8,9-HPCDF	M/M+2	0.46	0.37-0.51	97.4	77 - 129
Cleanup Standard					
37CL-2,3,7,8-TCDD ⁶				9.31	7.9 - 12.7

(1) Where applicable, custom lab flags have been used on this report.

(2) See Table 8, Method 1613, for m/z specifications.

(3) Ion Abundance Ratio Control Limits as specified in Table 9, Method 1613.

(4) Contract-required concentration range as determined from the percent of the test concentration in Table 6, Method 1613, under VER.

(5) Alternate confirmation and quantitation ions used for native and labeled PECDD.

(6) No ion abundance ratio for 37Cl4-2,3,7,8-TCDD; concentration reported.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: Thong DoFor Axys Internal Use Only [XSL Template: Form4B.xsl; Created: 22-Nov-2018 12:38:02; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DX8M_126AS20__Form4B_SJ2455651.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 6A
PCDD/PCDF RELATIVE RETENTION TIMES

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_126A S: 20

Instrument ID: HR GC/MS

Analysis Date: 25-Oct-2018

GC Column ID: DB5

Analysis Time: 01:54:19

COMPOUND	LAB FLAG ¹	RETENTION TIME REFERENCE	RRT	RRT QC LIMITS ²
2,3,7,8-TCDD		13C-2,3,7,8-TCDD	1.001	0.999-1.002
1,2,3,7,8-PECDD ³		13C-1,2,3,7,8-PECDD	1.001	0.999-1.002
1,2,3,4,7,8-HXCDD		13C-1,2,3,4,7,8-HXCDD	1.000	0.999-1.001
1,2,3,6,7,8-HXCDD		13C-1,2,3,6,7,8-HXCDD	1.000	0.998-1.004
1,2,3,7,8,9-HXCDD		13C-1,2,3,6,7,8-HXCDD	1.010	1.000-1.019
1,2,3,4,6,7,8-HPCDD		13C-1,2,3,4,6,7,8-HPCDD	1.000	0.999-1.001
OCDD		13C-OCDD	1.000	0.999-1.001
2,3,7,8-TCDF		13C-2,3,7,8-TCDF	1.001	0.999-1.003
1,2,3,7,8-PECDF		13C-1,2,3,7,8-PECDF	1.001	0.999-1.002
2,3,4,7,8-PECDF		13C-2,3,4,7,8-PECDF	1.001	0.999-1.002
1,2,3,4,7,8-HXCDF		13C-1,2,3,4,7,8-HXCDF	1.000	0.999-1.001
1,2,3,6,7,8-HXCDF		13C-1,2,3,6,7,8-HXCDF	1.001	0.997-1.005
1,2,3,7,8,9-HXCDF		13C-1,2,3,7,8,9-HXCDF	1.001	0.999-1.001
2,3,4,6,7,8-HXCDF		13C-2,3,4,6,7,8-HXCDF	1.000	0.999-1.001
1,2,3,4,6,7,8-HPCDF		13C-1,2,3,4,6,7,8-HPCDF	1.000	0.999-1.001
1,2,3,4,7,8,9-HPCDF		13C-1,2,3,4,7,8,9-HPCDF	1.000	0.999-1.001
OCDF		13C-OCDD	1.002	0.999-1.008

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for Relative Retention Times (RRT) as specified in Table 2, Method 1613.

(3) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ Thong Do _____

For Axys Internal Use Only [XSL Template: Form6A.xsl; Created: 22-Nov-2018 12:38:02; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DX8M_126AS20__Form6A_SJ2455651.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 6B
PCDD/PCDF RELATIVE RETENTION TIMES

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_126A S: 20

Instrument ID: HR GC/MS

Analysis Date: 25-Oct-2018

GC Column ID: DB5

Analysis Time: 01:54:19

LAB FLAG ¹	RETENTION TIME REFERENCE	RRT	RRT QC LIMITS ²
LABELED COMPOUND			
13C-2,3,7,8-TCDD	13C-1,2,3,4-TCDD	1.013	0.976-1.043
13C-1,2,3,7,8-PECDD	13C-1,2,3,4-TCDD	1.382	1.000-1.567
13C-1,2,3,4,7,8-HXCDD	13C-1,2,3,7,8,9-HXCDD	0.987	0.977-1.000
13C-1,2,3,6,7,8-HXCDD	13C-1,2,3,7,8,9-HXCDD	0.990	0.981-1.003
13C-1,2,3,4,6,7,8-HPCDD	13C-1,2,3,7,8,9-HXCDD	1.095	1.086-1.110
13C-OCDD	13C-1,2,3,7,8,9-HXCDD	1.179	1.032-1.311
13C-2,3,7,8-TCDF	13C-1,2,3,4-TCDD	0.966	0.923-1.103
13C-1,2,3,7,8-PECDF	13C-1,2,3,4-TCDD	1.283	1.000-1.425
13C-2,3,4,7,8-PECDF	13C-1,2,3,4-TCDD	1.351	1.011-1.526
13C-1,2,3,4,7,8-HXCDF	13C-1,2,3,7,8,9-HXCDD	0.954	0.944-0.970
13C-1,2,3,6,7,8-HXCDF	13C-1,2,3,7,8,9-HXCDD	0.958	0.949-0.975
13C-1,2,3,7,8,9-HXCDF	13C-1,2,3,7,8,9-HXCDD	1.004	0.977-1.047
13C-2,3,4,6,7,8-HXCDF	13C-1,2,3,7,8,9-HXCDD	0.980	0.959-1.021
13C-1,2,3,4,6,7,8-HPCDF	13C-1,2,3,7,8,9-HXCDD	1.062	1.043-1.085
13C-1,2,3,4,7,8,9-HPCDF	13C-1,2,3,7,8,9-HXCDD	1.104	1.057-1.151
CLEANUP STANDARD			
37CL-2,3,7,8-TCDD	13C-1,2,3,4-TCDD	1.001	0.989-1.052

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for Relative Retention Times (RRT) as specified in Table 2, Method 1613.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ Thong Do _____

For Axys Internal Use Only [XSL Template: Form6B.xsl; Created: 22-Nov-2018 12:38:02; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DX8M_126AS20_Form6B_SJ2455651.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 4A
PCDD/PCDF CALIBRATION VERIFICATION

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_126A S: 32

Instrument ID: HR GC/MS

Analysis Date: 25-Oct-2018

GC Column ID: DB5

Analysis Time: 13:03:27

COMPOUND	LAB FLAG ¹	MZ's FORMING RATIO ²	ION ABUND. RATIO	QC LIMITS ³	CONC. FOUND (ng/mL)	CONC. RANGE (ng/mL) ⁴
2,3,7,8-TCDD		M/M+2	0.78	0.65-0.89	10.1	7.8 - 12.9
1,2,3,7,8-PECDD ⁵		M/M+2	0.62	0.52-0.70	50.7	39 - 65
1,2,3,4,7,8-HXCDD		M+2/M+4	1.26	1.05-1.43	51.1	39 - 64
1,2,3,6,7,8-HXCDD		M+2/M+4	1.23	1.05-1.43	51.0	39 - 64
1,2,3,7,8,9-HXCDD		M+2/M+4	1.23	1.05-1.43	52.0	41 - 61
1,2,3,4,6,7,8-HPCDD		M+2/M+4	1.03	0.88-1.20	50.2	43 - 58
OCDD		M+2/M+4	0.88	0.76-1.02	103	79 - 126
2,3,7,8-TCDF		M/M+2	0.79	0.65-0.89	10.1	8.4 - 12
1,2,3,7,8-PECDF		M+2/M+4	1.56	1.32-1.78	51.1	41 - 60
2,3,4,7,8-PECDF		M+2/M+4	1.57	1.32-1.78	50.9	41 - 61
1,2,3,4,7,8-HXCDF		M+2/M+4	1.25	1.05-1.43	51.4	45 - 56
1,2,3,6,7,8-HXCDF		M+2/M+4	1.25	1.05-1.43	51.9	44 - 57
1,2,3,7,8,9-HXCDF		M+2/M+4	1.25	1.05-1.43	49.4	45 - 56
2,3,4,6,7,8-HXCDF		M+2/M+4	1.26	1.05-1.43	51.1	44 - 57
1,2,3,4,6,7,8-HPCDF		M+2/M+4	1.03	0.88-1.20	51.6	45 - 55
1,2,3,4,7,8,9-HPCDF		M+2/M+4	1.04	0.88-1.20	50.7	43 - 58
OCDF		M+2/M+4	0.89	0.76-1.02	101	63 - 159

(1) Where applicable, custom lab flags have been used on this report.

(2) See Table 8, Method 1613, for m/z specifications.

(3) Ion Abundance Ratio Control Limits as specified in Table 9, Method 1613.

(4) Contract-required concentration range as determined from the percent of the test concentration in Table 6, Method 1613, under VER.

(5) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Thong Do_____

For Axy Internal Use Only [XSL Template: Form4A.xsl; Created: 22-Nov-2018 12:38:02; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DX8M_126AS32_Form4A_SJ2455654.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 4B
PCDD/PCDF CALIBRATION VERIFICATION

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_126A S: 32

Instrument ID: HR GC/MS

Analysis Date: 25-Oct-2018

GC Column ID: DB5

Analysis Time: 13:03:27

LABELLED COMPOUND	LAB FLAG ¹	MZ's FORMING RATIO ²	ION ABUND. RATIO	QC LIMITS ³	CONC. FOUND (ng/mL)	CONC. RANGE (ng/mL) ⁴
13C-2,3,7,8-TCDD		M/M+2	0.77	0.65-0.89	101	82 - 121
13C-1,2,3,7,8-PECDD ⁵		M/M+2	0.63	0.52-0.70	101	62 - 160
13C-1,2,3,4,7,8-HXCDD		M+2/M+4	1.27	1.05-1.43	97.7	85 - 117
13C-1,2,3,6,7,8-HXCDD		M+2/M+4	1.26	1.05-1.43	99.3	85 - 118
13C-1,2,3,4,6,7,8-HPCDD		M+2/M+4	1.05	0.88-1.20	99.2	72 - 138
13C-OCDD		M+2/M+4	0.89	0.76-1.02	195	96 - 415
13C-2,3,7,8-TCDF		M/M+2	0.78	0.65-0.89	100	71 - 140
13C-1,2,3,7,8-PECDF		M+2/M+4	1.57	1.32-1.78	98.5	76 - 130
13C-2,3,4,7,8-PECDF		M+2/M+4	1.60	1.32-1.78	99.5	77 - 130
13C-1,2,3,4,7,8-HXCDF		M/M+2	0.51	0.43-0.59	97.6	76 - 131
13C-1,2,3,6,7,8-HXCDF		M/M+2	0.52	0.43-0.59	96.8	70 - 143
13C-1,2,3,7,8,9-HXCDF		M/M+2	0.52	0.43-0.59	99.2	74 - 135
13C-2,3,4,6,7,8-HXCDF		M/M+2	0.53	0.43-0.59	96.7	73 - 137
13C-1,2,3,4,6,7,8-HPCDF		M/M+2	0.44	0.37-0.51	96.5	78 - 129
13C-1,2,3,4,7,8,9-HPCDF		M/M+2	0.44	0.37-0.51	99.6	77 - 129
CLEANUP STANDARD						
37CL-2,3,7,8-TCDD ⁶					9.72	7.9 - 12.7

(1) Where applicable, custom lab flags have been used on this report.

(2) See Table 8, Method 1613, for m/z specifications.

(3) Ion Abundance Ratio Control Limits as specified in Table 9, Method 1613.

(4) Contract-required concentration range as determined from the percent of the test concentration in Table 6, Method 1613, under VER.

(5) Alternate confirmation and quantitation ions used for native and labeled PECDD.

(6) No ion abundance ratio for 37Cl4-2,3,7,8-TCDD; concentration reported.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: Thong DoFor Axys Internal Use Only [XSL Template: Form4B.xsl; Created: 22-Nov-2018 12:38:02; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DX8M_126AS32_Form4B_SJ2455654.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 6A
PCDD/PCDF RELATIVE RETENTION TIMES

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_126A S: 32

Instrument ID: HR GC/MS

Analysis Date: 25-Oct-2018

GC Column ID: DB5

Analysis Time: 13:03:27

COMPOUND	LAB FLAG ¹	RETENTION TIME REFERENCE	RRT	RRT QC LIMITS ²
2,3,7,8-TCDD		13C-2,3,7,8-TCDD	1.001	0.999-1.002
1,2,3,7,8-PECDD ³		13C-1,2,3,7,8-PECDD	1.001	0.999-1.002
1,2,3,4,7,8-HXCDD		13C-1,2,3,4,7,8-HXCDD	1.000	0.999-1.001
1,2,3,6,7,8-HXCDD		13C-1,2,3,6,7,8-HXCDD	1.000	0.998-1.004
1,2,3,7,8,9-HXCDD		13C-1,2,3,6,7,8-HXCDD	1.010	1.000-1.019
1,2,3,4,6,7,8-HPCDD		13C-1,2,3,4,6,7,8-HPCDD	1.000	0.999-1.001
OCDD		13C-OCDD	1.000	0.999-1.001
2,3,7,8-TCDF		13C-2,3,7,8-TCDF	1.001	0.999-1.003
1,2,3,7,8-PECDF		13C-1,2,3,7,8-PECDF	1.001	0.999-1.002
2,3,4,7,8-PECDF		13C-2,3,4,7,8-PECDF	1.001	0.999-1.002
1,2,3,4,7,8-HXCDF		13C-1,2,3,4,7,8-HXCDF	1.001	0.999-1.001
1,2,3,6,7,8-HXCDF		13C-1,2,3,6,7,8-HXCDF	1.000	0.997-1.005
1,2,3,7,8,9-HXCDF		13C-1,2,3,7,8,9-HXCDF	1.000	0.999-1.001
2,3,4,6,7,8-HXCDF		13C-2,3,4,6,7,8-HXCDF	1.000	0.999-1.001
1,2,3,4,6,7,8-HPCDF		13C-1,2,3,4,6,7,8-HPCDF	1.000	0.999-1.001
1,2,3,4,7,8,9-HPCDF		13C-1,2,3,4,7,8,9-HPCDF	1.000	0.999-1.001
OCDF		13C-OCDD	1.002	0.999-1.008

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for Relative Retention Times (RRT) as specified in Table 2, Method 1613.

(3) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ Thong Do _____

For Axys Internal Use Only [XSL Template: Form6A.xsl; Created: 22-Nov-2018 12:38:02; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DX8M_126AS32_Form6A_SJ2455654.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 6B
PCDD/PCDF RELATIVE RETENTION TIMES

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
 V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_126A S: 32

Instrument ID: HR GC/MS

Analysis Date: 25-Oct-2018

GC Column ID: DB5

Analysis Time: 13:03:27

LAB FLAG ¹	RETENTION TIME REFERENCE	RRT	RRT QC LIMITS ²
LABELED COMPOUND			
13C-2,3,7,8-TCDD	13C-1,2,3,4-TCDD	1.013	0.976-1.043
13C-1,2,3,7,8-PECDD	13C-1,2,3,4-TCDD	1.382	1.000-1.567
13C-1,2,3,4,7,8-HXCDD	13C-1,2,3,7,8,9-HXCDD	0.987	0.977-1.000
13C-1,2,3,6,7,8-HXCDD	13C-1,2,3,7,8,9-HXCDD	0.990	0.981-1.003
13C-1,2,3,4,6,7,8-HPCDD	13C-1,2,3,7,8,9-HXCDD	1.094	1.086-1.110
13C-OCDD	13C-1,2,3,7,8,9-HXCDD	1.179	1.032-1.311
13C-2,3,7,8-TCDF	13C-1,2,3,4-TCDD	0.966	0.923-1.103
13C-1,2,3,7,8-PECDF	13C-1,2,3,4-TCDD	1.283	1.000-1.425
13C-2,3,4,7,8-PECDF	13C-1,2,3,4-TCDD	1.350	1.011-1.526
13C-1,2,3,4,7,8-HXCDF	13C-1,2,3,7,8,9-HXCDD	0.954	0.944-0.970
13C-1,2,3,6,7,8-HXCDF	13C-1,2,3,7,8,9-HXCDD	0.958	0.949-0.975
13C-1,2,3,7,8,9-HXCDF	13C-1,2,3,7,8,9-HXCDD	1.004	0.977-1.047
13C-2,3,4,6,7,8-HXCDF	13C-1,2,3,7,8,9-HXCDD	0.980	0.959-1.021
13C-1,2,3,4,6,7,8-HPCDF	13C-1,2,3,7,8,9-HXCDD	1.062	1.043-1.085
13C-1,2,3,4,7,8,9-HPCDF	13C-1,2,3,7,8,9-HXCDD	1.104	1.057-1.151
CLEANUP STANDARD			
37CL-2,3,7,8-TCDD	13C-1,2,3,4-TCDD	1.001	0.989-1.052

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for Relative Retention Times (RRT) as specified in Table 2, Method 1613.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ **Thong Do** _____

For Axys Internal Use Only [XSL Template: Form6B.xsl; Created: 22-Nov-2018 12:38:02; Application: XMLTransformer-1.16.51;
 Report Filename: 1613_DIOXINS_DX8M_126AS32_Form6B_SJ2455654.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 5

PCDD/PCDF RT WINDOW AND ISOMER SPECIFICITY STANDARDS

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Instrument ID:	HR GC/MS	Initial Calibration Date:	24-Oct-2018
RT Window Data Filename:	DX8M_126A S: 6	Analysis Date:	24-Oct-2018
DB-5 IS Data Filename:	DX8M_126A S: 6	Analysis Date:	24-Oct-2018
DB-225 IS Data Filename:		Analysis Date:	
		Time:	12:54:49
		Time:	12:54:49
		Time:	

DB5 RT WINDOW DEFINING STANDARDS RESULT

ISOMERS	ABSOLUTE RT	ISOMERS	ABSOLUTE RT
1,3,6,8-TCDD (F)	23:00	1,3,6,8-TCDF (F)	21:27
1,2,8,9-TCDD (L)	28:21	1,2,8,9-TCDF (L)	28:12
1,2,4,7,9-PECDD (F)	32:07	1,3,4,6,8-PECDF (F)	28:56
1,2,3,8,9-PECDD (L)	37:06	1,2,3,8,9-PECDF (L)	37:09
1,2,4,6,7,9-HXCDD (F)	40:04	1,2,3,4,6,8-HXCDF (F)	39:01
1,2,3,4,6,7-HXCDD (L)	42:43	1,2,3,4,8,9-HXCDF (L)	43:03
1,2,3,4,6,7,9-HPCDD (F)	45:50	1,2,3,4,6,7,8-HPCDF (F)	45:22
1,2,3,4,6,7,8-HPCDD (L)	46:45	1,2,3,4,7,8,9-HPCDF (L)	47:10

(F) = First eluting isomer (DB-5); (L) = Last eluting isomer (DB-5)

ISOMER SPECIFICITY (IS) TEST STANDARDS RESULT

Isomers	% Valley Height Between Compared Peaks	Isomers	% Valley Height Between Compared Peaks
1,2,3,4-TCDD 1,2,7,8-TCDD	0	1,2,3,8-TCDD 2,3,7,8-TCDD	9
1,2,7,8-TCDD 1,4,7,8-TCDD	0	2,3,4,7-TCDF 2,3,7,8-TCDF	N/A
1,4,7,8-TCDD 1,2,3,7-TCDD	0	2,3,7,8-TCDF 1,2,3,9-TCDF	N/A
1,2,3,7-TCDD 1,2,3,8-TCDD	DB-5 column; co-elute as per Figure 6 in Method		

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ Leanne Henley _____

SGS AXYS METHOD MLA-017 Rev 20

Form 5

PCDD/PCDF RT WINDOW AND ISOMER SPECIFICITY STANDARDS

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Instrument ID:	HR GC/MS	Initial Calibration Date:	24-Oct-2018
RT Window Data Filename:	DX8M_126A S: 20	Analysis Date:	25-Oct-2018
DB-5 IS Data Filename:	DX8M_126A S: 20	Analysis Date:	25-Oct-2018
DB-225 IS Data Filename:		Analysis Date:	
		Time:	01:54:19
		Time:	01:54:19
		Time:	

DB5 RT WINDOW DEFINING STANDARDS RESULT

ISOMERS	ABSOLUTE RT	ISOMERS	ABSOLUTE RT
1,3,6,8-TCDD (F)	23:00	1,3,6,8-TCDF (F)	21:27
1,2,8,9-TCDD (L)	28:21	1,2,8,9-TCDF (L)	28:12
1,2,4,7,9-PECDD (F)	32:07	1,3,4,6,8-PECDF (F)	28:56
1,2,3,8,9-PECDD (L)	37:06	1,2,3,8,9-PECDF (L)	37:09
1,2,4,6,7,9-HXCDD (F)	40:04	1,2,3,4,6,8-HXCDF (F)	39:01
1,2,3,4,6,7-HXCDD (L)	42:43	1,2,3,4,8,9-HXCDF (L)	43:03
1,2,3,4,6,7,9-HPCDD (F)	45:50	1,2,3,4,6,7,8-HPCDF (F)	45:22
1,2,3,4,6,7,8-HPCDD (L)	46:45	1,2,3,4,7,8,9-HPCDF (L)	47:10

(F) = First eluting isomer (DB-5); (L) = Last eluting isomer (DB-5)

ISOMER SPECIFICITY (IS) TEST STANDARDS RESULT

Isomers	% Valley Height Between Compared Peaks	Isomers	% Valley Height Between Compared Peaks
1,2,3,4-TCDD 1,2,7,8-TCDD	0	1,2,3,8-TCDD 2,3,7,8-TCDD	9
1,2,7,8-TCDD 1,4,7,8-TCDD	0	2,3,4,7-TCDF 2,3,7,8-TCDF	N/A
1,4,7,8-TCDD 1,2,3,7-TCDD	0	2,3,7,8-TCDF 1,2,3,9-TCDF	N/A
1,2,3,7-TCDD 1,2,3,8-TCDD	DB-5 column; co-elute as per Figure 6 in Method		

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ Thong Do _____

For Axy Internal Use Only [XSL Template: DXForm5.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS__DX8M_126AS20_Form5_SJ2455407.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 4A
PCDD/PCDF CALIBRATION VERIFICATION

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_127A S: 1

Instrument ID: HR GC/MS

Analysis Date: 29-Oct-2018

GC Column ID: DB5

Analysis Time: 07:59:05

COMPOUND	LAB FLAG ¹	MZ's FORMING RATIO ²	ION ABUND. RATIO	QC LIMITS ³	CONC. FOUND (ng/mL)	CONC. RANGE (ng/mL) ⁴
2,3,7,8-TCDD		M/M+2	0.79	0.65-0.89	9.98	7.8 - 12.9
1,2,3,7,8-PECDD ⁵		M/M+2	0.61	0.52-0.70	50.7	39 - 65
1,2,3,4,7,8-HXCDD		M+2/M+4	1.26	1.05-1.43	50.5	39 - 64
1,2,3,6,7,8-HXCDD		M+2/M+4	1.23	1.05-1.43	50.0	39 - 64
1,2,3,7,8,9-HXCDD		M+2/M+4	1.23	1.05-1.43	50.4	41 - 61
1,2,3,4,6,7,8-HPCDD		M+2/M+4	1.03	0.88-1.20	51.7	43 - 58
OCDD		M+2/M+4	0.88	0.76-1.02	101	79 - 126
2,3,7,8-TCDF		M/M+2	0.78	0.65-0.89	10.0	8.4 - 12
1,2,3,7,8-PECDF		M+2/M+4	1.57	1.32-1.78	50.9	41 - 60
2,3,4,7,8-PECDF		M+2/M+4	1.56	1.32-1.78	50.0	41 - 61
1,2,3,4,7,8-HXCDF		M+2/M+4	1.24	1.05-1.43	51.3	45 - 56
1,2,3,6,7,8-HXCDF		M+2/M+4	1.26	1.05-1.43	51.4	44 - 57
1,2,3,7,8,9-HXCDF		M+2/M+4	1.24	1.05-1.43	48.5	45 - 56
2,3,4,6,7,8-HXCDF		M+2/M+4	1.23	1.05-1.43	50.6	44 - 57
1,2,3,4,6,7,8-HPCDF		M+2/M+4	1.04	0.88-1.20	52.4	45 - 55
1,2,3,4,7,8,9-HPCDF		M+2/M+4	1.05	0.88-1.20	51.5	43 - 58
OCDF		M+2/M+4	0.90	0.76-1.02	99.6	63 - 159

(1) Where applicable, custom lab flags have been used on this report.

(2) See Table 8, Method 1613, for m/z specifications.

(3) Ion Abundance Ratio Control Limits as specified in Table 9, Method 1613.

(4) Contract-required concentration range as determined from the percent of the test concentration in Table 6, Method 1613, under VER.

(5) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Thong Do_____

For Axy Internal Use Only [XSL Template: Form4A.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DX8M_127AS1_Form4A_SJ2456911.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 4B
PCDD/PCDF CALIBRATION VERIFICATION

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_127A S: 1

Instrument ID: HR GC/MS

Analysis Date: 29-Oct-2018

GC Column ID: DB5

Analysis Time: 07:59:05

LABELLED COMPOUND	LAB FLAG ¹	MZ's FORMING RATIO ²	ION ABUND. RATIO	QC LIMITS ³	CONC. FOUND (ng/mL)	CONC. RANGE (ng/mL) ⁴
13C-2,3,7,8-TCDD		M/M+2	0.78	0.65-0.89	103	82 - 121
13C-1,2,3,7,8-PECDD ⁵		M/M+2	0.64	0.52-0.70	93.0	62 - 160
13C-1,2,3,4,7,8-HXCDD		M+2/M+4	1.27	1.05-1.43	98.5	85 - 117
13C-1,2,3,6,7,8-HXCDD		M+2/M+4	1.23	1.05-1.43	101	85 - 118
13C-1,2,3,4,6,7,8-HPCDD		M+2/M+4	1.06	0.88-1.20	97.1	72 - 138
13C-OCDD		M+2/M+4	0.91	0.76-1.02	186	96 - 415
13C-2,3,7,8-TCDF		M/M+2	0.79	0.65-0.89	103	71 - 140
13C-1,2,3,7,8-PECDF		M+2/M+4	1.58	1.32-1.78	92.1	76 - 130
13C-2,3,4,7,8-PECDF		M+2/M+4	1.58	1.32-1.78	91.9	77 - 130
13C-1,2,3,4,7,8-HXCDF		M/M+2	0.51	0.43-0.59	97.9	76 - 131
13C-1,2,3,6,7,8-HXCDF		M/M+2	0.52	0.43-0.59	96.9	70 - 143
13C-1,2,3,7,8,9-HXCDF		M/M+2	0.53	0.43-0.59	98.9	74 - 135
13C-2,3,4,6,7,8-HXCDF		M/M+2	0.52	0.43-0.59	97.1	73 - 137
13C-1,2,3,4,6,7,8-HPCDF		M/M+2	0.44	0.37-0.51	94.1	78 - 129
13C-1,2,3,4,7,8,9-HPCDF		M/M+2	0.45	0.37-0.51	92.4	77 - 129
CLEANUP STANDARD						
37CL-2,3,7,8-TCDD ⁶					9.92	7.9 - 12.7

(1) Where applicable, custom lab flags have been used on this report.

(2) See Table 8, Method 1613, for m/z specifications.

(3) Ion Abundance Ratio Control Limits as specified in Table 9, Method 1613.

(4) Contract-required concentration range as determined from the percent of the test concentration in Table 6, Method 1613, under VER.

(5) Alternate confirmation and quantitation ions used for native and labeled PECDD.

(6) No ion abundance ratio for 37Cl4-2,3,7,8-TCDD; concentration reported.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: Thong DoFor Axys Internal Use Only [XSL Template: Form4B.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DX8M_127AS1_Form4B_SJ2456911.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 6A
PCDD/PCDF RELATIVE RETENTION TIMES

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_127A S: 1

Instrument ID: HR GC/MS

Analysis Date: 29-Oct-2018

GC Column ID: DB5

Analysis Time: 07:59:05

COMPOUND	LAB FLAG ¹	RETENTION TIME REFERENCE	RRT	RRT QC LIMITS ²
2,3,7,8-TCDD		13C-2,3,7,8-TCDD	1.001	0.999-1.002
1,2,3,7,8-PECDD ³		13C-1,2,3,7,8-PECDD	1.001	0.999-1.002
1,2,3,4,7,8-HXCDD		13C-1,2,3,4,7,8-HXCDD	1.000	0.999-1.001
1,2,3,6,7,8-HXCDD		13C-1,2,3,6,7,8-HXCDD	1.000	0.998-1.004
1,2,3,7,8,9-HXCDD		13C-1,2,3,6,7,8-HXCDD	1.010	1.000-1.019
1,2,3,4,6,7,8-HPCDD		13C-1,2,3,4,6,7,8-HPCDD	1.000	0.999-1.001
OCDD		13C-OCDD	1.000	0.999-1.001
2,3,7,8-TCDF		13C-2,3,7,8-TCDF	1.001	0.999-1.003
1,2,3,7,8-PECDF		13C-1,2,3,7,8-PECDF	1.000	0.999-1.002
2,3,4,7,8-PECDF		13C-2,3,4,7,8-PECDF	1.000	0.999-1.002
1,2,3,4,7,8-HXCDF		13C-1,2,3,4,7,8-HXCDF	1.001	0.999-1.001
1,2,3,6,7,8-HXCDF		13C-1,2,3,6,7,8-HXCDF	1.000	0.997-1.005
1,2,3,7,8,9-HXCDF		13C-1,2,3,7,8,9-HXCDF	1.001	0.999-1.001
2,3,4,6,7,8-HXCDF		13C-2,3,4,6,7,8-HXCDF	1.001	0.999-1.001
1,2,3,4,6,7,8-HPCDF		13C-1,2,3,4,6,7,8-HPCDF	1.000	0.999-1.001
1,2,3,4,7,8,9-HPCDF		13C-1,2,3,4,7,8,9-HPCDF	1.000	0.999-1.001
OCDF		13C-OCDD	1.002	0.999-1.008

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for Relative Retention Times (RRT) as specified in Table 2, Method 1613.

(3) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ **Thong Do** _____

For Axys Internal Use Only [XSL Template: Form6A.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DX8M_127AS1__Form6A_SJ2456911.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 6B
PCDD/PCDF RELATIVE RETENTION TIMES

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_127A S: 1

Instrument ID: HR GC/MS

Analysis Date: 29-Oct-2018

GC Column ID: DB5

Analysis Time: 07:59:05

LAB FLAG ¹	RETENTION TIME REFERENCE	RRT	RRT QC LIMITS ²
LABELED COMPOUND			
13C-2,3,7,8-TCDD	13C-1,2,3,4-TCDD	1.013	0.976-1.043
13C-1,2,3,7,8-PECDD	13C-1,2,3,4-TCDD	1.381	1.000-1.567
13C-1,2,3,4,7,8-HXCDD	13C-1,2,3,7,8,9-HXCDD	0.987	0.977-1.000
13C-1,2,3,6,7,8-HXCDD	13C-1,2,3,7,8,9-HXCDD	0.990	0.981-1.003
13C-1,2,3,4,6,7,8-HPCDD	13C-1,2,3,7,8,9-HXCDD	1.094	1.086-1.110
13C-OCDD	13C-1,2,3,7,8,9-HXCDD	1.178	1.032-1.311
13C-2,3,7,8-TCDF	13C-1,2,3,4-TCDD	0.966	0.923-1.103
13C-1,2,3,7,8-PECDF	13C-1,2,3,4-TCDD	1.283	1.000-1.425
13C-2,3,4,7,8-PECDF	13C-1,2,3,4-TCDD	1.350	1.011-1.526
13C-1,2,3,4,7,8-HXCDF	13C-1,2,3,7,8,9-HXCDD	0.954	0.944-0.970
13C-1,2,3,6,7,8-HXCDF	13C-1,2,3,7,8,9-HXCDD	0.958	0.949-0.975
13C-1,2,3,7,8,9-HXCDF	13C-1,2,3,7,8,9-HXCDD	1.004	0.977-1.047
13C-2,3,4,6,7,8-HXCDF	13C-1,2,3,7,8,9-HXCDD	0.980	0.959-1.021
13C-1,2,3,4,6,7,8-HPCDF	13C-1,2,3,7,8,9-HXCDD	1.062	1.043-1.085
13C-1,2,3,4,7,8,9-HPCDF	13C-1,2,3,7,8,9-HXCDD	1.104	1.057-1.151
CLEANUP STANDARD			
37CL-2,3,7,8-TCDD	13C-1,2,3,4-TCDD	1.001	0.989-1.052

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for Relative Retention Times (RRT) as specified in Table 2, Method 1613.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ Thong Do _____

For Axys Internal Use Only [XSL Template: Form6B.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DX8M_127AS1_Form6B_SJ2456911.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 4A
PCDD/PCDF CALIBRATION VERIFICATION

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_127A S: 12

Instrument ID: HR GC/MS

Analysis Date: 29-Oct-2018

GC Column ID: DB5

Analysis Time: 18:07:09

COMPOUND	LAB FLAG ¹	MZ's FORMING RATIO ²	ION ABUND. RATIO	QC LIMITS ³	CONC. FOUND (ng/mL)	CONC. RANGE (ng/mL) ⁴
2,3,7,8-TCDD		M/M+2	0.79	0.65-0.89	9.85	7.8 - 12.9
1,2,3,7,8-PECDD ⁵		M/M+2	0.61	0.52-0.70	50.6	39 - 65
1,2,3,4,7,8-HXCDD		M+2/M+4	1.24	1.05-1.43	50.6	39 - 64
1,2,3,6,7,8-HXCDD		M+2/M+4	1.26	1.05-1.43	51.3	39 - 64
1,2,3,7,8,9-HXCDD		M+2/M+4	1.24	1.05-1.43	51.7	41 - 61
1,2,3,4,6,7,8-HPCDD		M+2/M+4	1.05	0.88-1.20	50.9	43 - 58
OCDD		M+2/M+4	0.88	0.76-1.02	101	79 - 126
2,3,7,8-TCDF		M/M+2	0.77	0.65-0.89	9.90	8.4 - 12
1,2,3,7,8-PECDF		M+2/M+4	1.56	1.32-1.78	50.5	41 - 60
2,3,4,7,8-PECDF		M+2/M+4	1.56	1.32-1.78	50.1	41 - 61
1,2,3,4,7,8-HXCDF		M+2/M+4	1.25	1.05-1.43	51.1	45 - 56
1,2,3,6,7,8-HXCDF		M+2/M+4	1.24	1.05-1.43	49.9	44 - 57
1,2,3,7,8,9-HXCDF		M+2/M+4	1.27	1.05-1.43	49.0	45 - 56
2,3,4,6,7,8-HXCDF		M+2/M+4	1.25	1.05-1.43	50.7	44 - 57
1,2,3,4,6,7,8-HPCDF		M+2/M+4	1.02	0.88-1.20	50.1	45 - 55
1,2,3,4,7,8,9-HPCDF		M+2/M+4	1.04	0.88-1.20	50.8	43 - 58
OCDF		M+2/M+4	0.90	0.76-1.02	97.7	63 - 159

(1) Where applicable, custom lab flags have been used on this report.

(2) See Table 8, Method 1613, for m/z specifications.

(3) Ion Abundance Ratio Control Limits as specified in Table 9, Method 1613.

(4) Contract-required concentration range as determined from the percent of the test concentration in Table 6, Method 1613, under VER.

(5) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Thong Do_____

For Axy Internal Use Only [XSL Template: Form4A.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DX8M_127AS12_Form4A_SJ2456910.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 4A
PCDD/PCDF CALIBRATION VERIFICATION

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_127A S: 12

Instrument ID: HR GC/MS

Analysis Date: 29-Oct-2018

GC Column ID: DB5

Analysis Time: 18:07:09

COMPOUND	LAB FLAG ¹	MZ's FORMING RATIO ²	ION ABUND. RATIO	QC LIMITS ³	CONC. FOUND (ng/mL)	CONC. RANGE (ng/mL) ⁴
2,3,7,8-TCDD		M/M+2	0.79	0.65-0.89	9.85	7.8 - 12.9
1,2,3,7,8-PECDD ⁵		M/M+2	0.61	0.52-0.70	50.6	39 - 65
1,2,3,4,7,8-HXCDD		M+2/M+4	1.24	1.05-1.43	50.6	39 - 64
1,2,3,6,7,8-HXCDD		M+2/M+4	1.26	1.05-1.43	51.3	39 - 64
1,2,3,7,8,9-HXCDD		M+2/M+4	1.24	1.05-1.43	51.7	41 - 61
1,2,3,4,6,7,8-HPCDD		M+2/M+4	1.05	0.88-1.20	50.9	43 - 58
OCDD		M+2/M+4	0.88	0.76-1.02	101	79 - 126
2,3,7,8-TCDF		M/M+2	0.77	0.65-0.89	9.90	8.4 - 12
1,2,3,7,8-PECDF		M+2/M+4	1.56	1.32-1.78	50.5	41 - 60
2,3,4,7,8-PECDF		M+2/M+4	1.56	1.32-1.78	50.1	41 - 61
1,2,3,4,7,8-HXCDF		M+2/M+4	1.25	1.05-1.43	51.1	45 - 56
1,2,3,6,7,8-HXCDF		M+2/M+4	1.24	1.05-1.43	49.9	44 - 57
1,2,3,7,8,9-HXCDF		M+2/M+4	1.27	1.05-1.43	49.0	45 - 56
2,3,4,6,7,8-HXCDF		M+2/M+4	1.25	1.05-1.43	50.7	44 - 57
1,2,3,4,6,7,8-HPCDF		M+2/M+4	1.02	0.88-1.20	50.1	45 - 55
1,2,3,4,7,8,9-HPCDF		M+2/M+4	1.04	0.88-1.20	50.8	43 - 58
OCDF		M+2/M+4	0.90	0.76-1.02	97.7	63 - 159

(1) Where applicable, custom lab flags have been used on this report.

(2) See Table 8, Method 1613, for m/z specifications.

(3) Ion Abundance Ratio Control Limits as specified in Table 9, Method 1613.

(4) Contract-required concentration range as determined from the percent of the test concentration in Table 6, Method 1613, under VER.

(5) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Thong Do_____

For Axy Internal Use Only [XSL Template: Form4A.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DX8M_127AS12_Form4A_SJ2457401.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 4B
PCDD/PCDF CALIBRATION VERIFICATION

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_127A S: 12

Instrument ID: HR GC/MS

Analysis Date: 29-Oct-2018

GC Column ID: DB5

Analysis Time: 18:07:09

LAB FLAG ¹	MZ's FORMING RATIO ²	ION ABUND. RATIO	QC LIMITS ³	CONC. FOUND (ng/mL)	CONC. RANGE (ng/mL) ⁴
LABELED COMPOUND					
13C-2,3,7,8-TCDD	M/M+2	0.78	0.65-0.89	103	82 - 121
13C-1,2,3,7,8-PECDD ⁵	M/M+2	0.64	0.52-0.70	102	62 - 160
13C-1,2,3,4,7,8-HXCDD	M+2/M+4	1.27	1.05-1.43	96.8	85 - 117
13C-1,2,3,6,7,8-HXCDD	M+2/M+4	1.24	1.05-1.43	96.1	85 - 118
13C-1,2,3,4,6,7,8-HPCDD	M+2/M+4	1.05	0.88-1.20	101	72 - 138
13C-OCDD	M+2/M+4	0.90	0.76-1.02	206	96 - 415
13C-2,3,7,8-TCDF	M/M+2	0.78	0.65-0.89	100	71 - 140
13C-1,2,3,7,8-PECDF	M+2/M+4	1.60	1.32-1.78	95.9	76 - 130
13C-2,3,4,7,8-PECDF	M+2/M+4	1.57	1.32-1.78	95.8	77 - 130
13C-1,2,3,4,7,8-HXCDF	M/M+2	0.51	0.43-0.59	92.6	76 - 131
13C-1,2,3,6,7,8-HXCDF	M/M+2	0.52	0.43-0.59	93.2	70 - 143
13C-1,2,3,7,8,9-HXCDF	M/M+2	0.52	0.43-0.59	96.3	74 - 135
13C-2,3,4,6,7,8-HXCDF	M/M+2	0.52	0.43-0.59	94.3	73 - 137
13C-1,2,3,4,6,7,8-HPCDF	M/M+2	0.44	0.37-0.51	95.1	78 - 129
13C-1,2,3,4,7,8,9-HPCDF	M/M+2	0.45	0.37-0.51	96.5	77 - 129
CLEANUP STANDARD					
37CL-2,3,7,8-TCDD ⁶				9.49	7.9 - 12.7

(1) Where applicable, custom lab flags have been used on this report.

(2) See Table 8, Method 1613, for m/z specifications.

(3) Ion Abundance Ratio Control Limits as specified in Table 9, Method 1613.

(4) Contract-required concentration range as determined from the percent of the test concentration in Table 6, Method 1613, under VER.

(5) Alternate confirmation and quantitation ions used for native and labeled PECDD.

(6) No ion abundance ratio for 37Cl4-2,3,7,8-TCDD; concentration reported.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: Thong DoFor Axys Internal Use Only [XSL Template: Form4B.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DX8M_127AS12_Form4B_SJ2456910.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 4B
PCDD/PCDF CALIBRATION VERIFICATION

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_127A S: 12

Instrument ID: HR GC/MS

Analysis Date: 29-Oct-2018

GC Column ID: DB5

Analysis Time: 18:07:09

LABELLED COMPOUND	LAB FLAG ¹	MZ's FORMING RATIO ²	ION ABUND. RATIO	QC LIMITS ³	CONC. FOUND (ng/mL)	CONC. RANGE (ng/mL) ⁴
13C-2,3,7,8-TCDD		M/M+2	0.78	0.65-0.89	103	82 - 121
13C-1,2,3,7,8-PECDD ⁵		M/M+2	0.64	0.52-0.70	102	62 - 160
13C-1,2,3,4,7,8-HXCDD		M+2/M+4	1.27	1.05-1.43	96.8	85 - 117
13C-1,2,3,6,7,8-HXCDD		M+2/M+4	1.24	1.05-1.43	96.1	85 - 118
13C-1,2,3,4,6,7,8-HPCDD		M+2/M+4	1.05	0.88-1.20	101	72 - 138
13C-OCDD		M+2/M+4	0.90	0.76-1.02	206	96 - 415
13C-2,3,7,8-TCDF		M/M+2	0.78	0.65-0.89	100	71 - 140
13C-1,2,3,7,8-PECDF		M+2/M+4	1.60	1.32-1.78	95.9	76 - 130
13C-2,3,4,7,8-PECDF		M+2/M+4	1.57	1.32-1.78	95.8	77 - 130
13C-1,2,3,4,7,8-HXCDF		M/M+2	0.51	0.43-0.59	92.6	76 - 131
13C-1,2,3,6,7,8-HXCDF		M/M+2	0.52	0.43-0.59	93.2	70 - 143
13C-1,2,3,7,8,9-HXCDF		M/M+2	0.52	0.43-0.59	96.3	74 - 135
13C-2,3,4,6,7,8-HXCDF		M/M+2	0.52	0.43-0.59	94.3	73 - 137
13C-1,2,3,4,6,7,8-HPCDF		M/M+2	0.44	0.37-0.51	95.1	78 - 129
13C-1,2,3,4,7,8,9-HPCDF		M/M+2	0.45	0.37-0.51	96.5	77 - 129
CLEANUP STANDARD						
37CL-2,3,7,8-TCDD ⁶					9.49	7.9 - 12.7

(1) Where applicable, custom lab flags have been used on this report.

(2) See Table 8, Method 1613, for m/z specifications.

(3) Ion Abundance Ratio Control Limits as specified in Table 9, Method 1613.

(4) Contract-required concentration range as determined from the percent of the test concentration in Table 6, Method 1613, under VER.

(5) Alternate confirmation and quantitation ions used for native and labeled PECDD.

(6) No ion abundance ratio for 37Cl4-2,3,7,8-TCDD; concentration reported.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: Thong DoFor Axys Internal Use Only [XSL Template: Form4B.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DX8M_127AS12_Form4B_SJ2457401.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 6A
PCDD/PCDF RELATIVE RETENTION TIMES

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_127A S: 12

Instrument ID: HR GC/MS

Analysis Date: 29-Oct-2018

GC Column ID: DB5

Analysis Time: 18:07:09

COMPOUND	LAB FLAG ¹	RETENTION TIME REFERENCE	RRT	RRT QC LIMITS ²
2,3,7,8-TCDD		13C-2,3,7,8-TCDD	1.001	0.999-1.002
1,2,3,7,8-PECDD ³		13C-1,2,3,7,8-PECDD	1.001	0.999-1.002
1,2,3,4,7,8-HXCDD		13C-1,2,3,4,7,8-HXCDD	1.000	0.999-1.001
1,2,3,6,7,8-HXCDD		13C-1,2,3,6,7,8-HXCDD	1.000	0.998-1.004
1,2,3,7,8,9-HXCDD		13C-1,2,3,6,7,8-HXCDD	1.010	1.000-1.019
1,2,3,4,6,7,8-HPCDD		13C-1,2,3,4,6,7,8-HPCDD	1.000	0.999-1.001
OCDD		13C-OCDD	1.000	0.999-1.001
2,3,7,8-TCDF		13C-2,3,7,8-TCDF	1.001	0.999-1.003
1,2,3,7,8-PECDF		13C-1,2,3,7,8-PECDF	1.001	0.999-1.002
2,3,4,7,8-PECDF		13C-2,3,4,7,8-PECDF	1.000	0.999-1.002
1,2,3,4,7,8-HXCDF		13C-1,2,3,4,7,8-HXCDF	1.000	0.999-1.001
1,2,3,6,7,8-HXCDF		13C-1,2,3,6,7,8-HXCDF	1.001	0.997-1.005
1,2,3,7,8,9-HXCDF		13C-1,2,3,7,8,9-HXCDF	1.000	0.999-1.001
2,3,4,6,7,8-HXCDF		13C-2,3,4,6,7,8-HXCDF	1.000	0.999-1.001
1,2,3,4,6,7,8-HPCDF		13C-1,2,3,4,6,7,8-HPCDF	1.000	0.999-1.001
1,2,3,4,7,8,9-HPCDF		13C-1,2,3,4,7,8,9-HPCDF	1.000	0.999-1.001
OCDF		13C-OCDD	1.002	0.999-1.008

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for Relative Retention Times (RRT) as specified in Table 2, Method 1613.

(3) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ Thong Do _____

For Axys Internal Use Only [XSL Template: Form6A.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DX8M_127AS12__Form6A_SJ2456910.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 6A
PCDD/PCDF RELATIVE RETENTION TIMES

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_127A S: 12

Instrument ID: HR GC/MS

Analysis Date: 29-Oct-2018

GC Column ID: DB5

Analysis Time: 18:07:09

COMPOUND	LAB FLAG ¹	RETENTION TIME REFERENCE	RRT	RRT QC LIMITS ²
2,3,7,8-TCDD		13C-2,3,7,8-TCDD	1.001	0.999-1.002
1,2,3,7,8-PECDD ³		13C-1,2,3,7,8-PECDD	1.001	0.999-1.002
1,2,3,4,7,8-HXCDD		13C-1,2,3,4,7,8-HXCDD	1.000	0.999-1.001
1,2,3,6,7,8-HXCDD		13C-1,2,3,6,7,8-HXCDD	1.000	0.998-1.004
1,2,3,7,8,9-HXCDD		13C-1,2,3,6,7,8-HXCDD	1.010	1.000-1.019
1,2,3,4,6,7,8-HPCDD		13C-1,2,3,4,6,7,8-HPCDD	1.000	0.999-1.001
OCDD		13C-OCDD	1.000	0.999-1.001
2,3,7,8-TCDF		13C-2,3,7,8-TCDF	1.001	0.999-1.003
1,2,3,7,8-PECDF		13C-1,2,3,7,8-PECDF	1.001	0.999-1.002
2,3,4,7,8-PECDF		13C-2,3,4,7,8-PECDF	1.000	0.999-1.002
1,2,3,4,7,8-HXCDF		13C-1,2,3,4,7,8-HXCDF	1.000	0.999-1.001
1,2,3,6,7,8-HXCDF		13C-1,2,3,6,7,8-HXCDF	1.001	0.997-1.005
1,2,3,7,8,9-HXCDF		13C-1,2,3,7,8,9-HXCDF	1.000	0.999-1.001
2,3,4,6,7,8-HXCDF		13C-2,3,4,6,7,8-HXCDF	1.000	0.999-1.001
1,2,3,4,6,7,8-HPCDF		13C-1,2,3,4,6,7,8-HPCDF	1.000	0.999-1.001
1,2,3,4,7,8,9-HPCDF		13C-1,2,3,4,7,8,9-HPCDF	1.000	0.999-1.001
OCDF		13C-OCDD	1.002	0.999-1.008

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for Relative Retention Times (RRT) as specified in Table 2, Method 1613.

(3) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ Thong Do _____

For Axys Internal Use Only [XSL Template: Form6A.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DX8M_127AS12__Form6A_SJ2457401.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 6B
PCDD/PCDF RELATIVE RETENTION TIMES

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
 V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_127A S: 12

Instrument ID: HR GC/MS

Analysis Date: 29-Oct-2018

GC Column ID: DB5

Analysis Time: 18:07:09

LAB FLAG ¹	RETENTION TIME REFERENCE	RRT	RRT QC LIMITS ²
LABELED COMPOUND			
13C-2,3,7,8-TCDD	13C-1,2,3,4-TCDD	1.012	0.976-1.043
13C-1,2,3,7,8-PECDD	13C-1,2,3,4-TCDD	1.381	1.000-1.567
13C-1,2,3,4,7,8-HXCDD	13C-1,2,3,7,8,9-HXCDD	0.987	0.977-1.000
13C-1,2,3,6,7,8-HXCDD	13C-1,2,3,7,8,9-HXCDD	0.990	0.981-1.003
13C-1,2,3,4,6,7,8-HPCDD	13C-1,2,3,7,8,9-HXCDD	1.094	1.086-1.110
13C-OCDD	13C-1,2,3,7,8,9-HXCDD	1.179	1.032-1.311
13C-2,3,7,8-TCDF	13C-1,2,3,4-TCDD	0.965	0.923-1.103
13C-1,2,3,7,8-PECDF	13C-1,2,3,4-TCDD	1.283	1.000-1.425
13C-2,3,4,7,8-PECDF	13C-1,2,3,4-TCDD	1.350	1.011-1.526
13C-1,2,3,4,7,8-HXCDF	13C-1,2,3,7,8,9-HXCDD	0.954	0.944-0.970
13C-1,2,3,6,7,8-HXCDF	13C-1,2,3,7,8,9-HXCDD	0.958	0.949-0.975
13C-1,2,3,7,8,9-HXCDF	13C-1,2,3,7,8,9-HXCDD	1.004	0.977-1.047
13C-2,3,4,6,7,8-HXCDF	13C-1,2,3,7,8,9-HXCDD	0.980	0.959-1.021
13C-1,2,3,4,6,7,8-HPCDF	13C-1,2,3,7,8,9-HXCDD	1.062	1.043-1.085
13C-1,2,3,4,7,8,9-HPCDF	13C-1,2,3,7,8,9-HXCDD	1.104	1.057-1.151
CLEANUP STANDARD			
37CL-2,3,7,8-TCDD	13C-1,2,3,4-TCDD	1.001	0.989-1.052

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for Relative Retention Times (RRT) as specified in Table 2, Method 1613.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ **Thong Do** _____

For Axys Internal Use Only [XSL Template: Form6B.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
 Report Filename: 1613_DIOXINS_DX8M_127AS12_Form6B_SJ2456910.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 6B
PCDD/PCDF RELATIVE RETENTION TIMES

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_127A S: 12

Instrument ID: HR GC/MS

Analysis Date: 29-Oct-2018

GC Column ID: DB5

Analysis Time: 18:07:09

LAB FLAG ¹	RETENTION TIME REFERENCE	RRT	RRT QC LIMITS ²
LABELED COMPOUND			
13C-2,3,7,8-TCDD	13C-1,2,3,4-TCDD	1.012	0.976-1.043
13C-1,2,3,7,8-PECDD	13C-1,2,3,4-TCDD	1.381	1.000-1.567
13C-1,2,3,4,7,8-HXCDD	13C-1,2,3,7,8,9-HXCDD	0.987	0.977-1.000
13C-1,2,3,6,7,8-HXCDD	13C-1,2,3,7,8,9-HXCDD	0.990	0.981-1.003
13C-1,2,3,4,6,7,8-HPCDD	13C-1,2,3,7,8,9-HXCDD	1.094	1.086-1.110
13C-OCDD	13C-1,2,3,7,8,9-HXCDD	1.179	1.032-1.311
13C-2,3,7,8-TCDF	13C-1,2,3,4-TCDD	0.965	0.923-1.103
13C-1,2,3,7,8-PECDF	13C-1,2,3,4-TCDD	1.283	1.000-1.425
13C-2,3,4,7,8-PECDF	13C-1,2,3,4-TCDD	1.350	1.011-1.526
13C-1,2,3,4,7,8-HXCDF	13C-1,2,3,7,8,9-HXCDD	0.954	0.944-0.970
13C-1,2,3,6,7,8-HXCDF	13C-1,2,3,7,8,9-HXCDD	0.958	0.949-0.975
13C-1,2,3,7,8,9-HXCDF	13C-1,2,3,7,8,9-HXCDD	1.004	0.977-1.047
13C-2,3,4,6,7,8-HXCDF	13C-1,2,3,7,8,9-HXCDD	0.980	0.959-1.021
13C-1,2,3,4,6,7,8-HPCDF	13C-1,2,3,7,8,9-HXCDD	1.062	1.043-1.085
13C-1,2,3,4,7,8,9-HPCDF	13C-1,2,3,7,8,9-HXCDD	1.104	1.057-1.151
CLEANUP STANDARD			
37CL-2,3,7,8-TCDD	13C-1,2,3,4-TCDD	1.001	0.989-1.052

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for Relative Retention Times (RRT) as specified in Table 2, Method 1613.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ Thong Do _____

For Axys Internal Use Only [XSL Template: Form6B.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DX8M_127AS12_Form6B_SJ2457401.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 4A
PCDD/PCDF CALIBRATION VERIFICATION

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_127A S: 24

Instrument ID: HR GC/MS

Analysis Date: 30-Oct-2018

GC Column ID: DB5

Analysis Time: 05:17:24

COMPOUND	LAB FLAG ¹	MZ's FORMING RATIO ²	ION ABUND. RATIO	QC LIMITS ³	CONC. FOUND (ng/mL)	CONC. RANGE (ng/mL) ⁴
2,3,7,8-TCDD		M/M+2	0.80	0.65-0.89	9.87	7.8 - 12.9
1,2,3,7,8-PECDD ⁵		M/M+2	0.62	0.52-0.70	50.2	39 - 65
1,2,3,4,7,8-HXCDD		M+2/M+4	1.25	1.05-1.43	49.7	39 - 64
1,2,3,6,7,8-HXCDD		M+2/M+4	1.25	1.05-1.43	49.9	39 - 64
1,2,3,7,8,9-HXCDD		M+2/M+4	1.21	1.05-1.43	50.8	41 - 61
1,2,3,4,6,7,8-HPCDD		M+2/M+4	1.04	0.88-1.20	51.8	43 - 58
OCDD		M+2/M+4	0.88	0.76-1.02	100	79 - 126
2,3,7,8-TCDF		M/M+2	0.77	0.65-0.89	9.83	8.4 - 12
1,2,3,7,8-PECDF		M+2/M+4	1.56	1.32-1.78	50.9	41 - 60
2,3,4,7,8-PECDF		M+2/M+4	1.55	1.32-1.78	49.7	41 - 61
1,2,3,4,7,8-HXCDF		M+2/M+4	1.23	1.05-1.43	51.4	45 - 56
1,2,3,6,7,8-HXCDF		M+2/M+4	1.23	1.05-1.43	50.2	44 - 57
1,2,3,7,8,9-HXCDF		M+2/M+4	1.22	1.05-1.43	48.5	45 - 56
2,3,4,6,7,8-HXCDF		M+2/M+4	1.24	1.05-1.43	50.7	44 - 57
1,2,3,4,6,7,8-HPCDF		M+2/M+4	1.03	0.88-1.20	50.1	45 - 55
1,2,3,4,7,8,9-HPCDF		M+2/M+4	1.04	0.88-1.20	51.3	43 - 58
OCDF		M+2/M+4	0.90	0.76-1.02	96.8	63 - 159

(1) Where applicable, custom lab flags have been used on this report.

(2) See Table 8, Method 1613, for m/z specifications.

(3) Ion Abundance Ratio Control Limits as specified in Table 9, Method 1613.

(4) Contract-required concentration range as determined from the percent of the test concentration in Table 6, Method 1613, under VER.

(5) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Thong Do_____

For Axy Internal Use Only [XSL Template: Form4A.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DX8M_127AS24__Form4A_SJ2457403.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 4B
PCDD/PCDF CALIBRATION VERIFICATION

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_127A S: 24

Instrument ID: HR GC/MS

Analysis Date: 30-Oct-2018

GC Column ID: DB5

Analysis Time: 05:17:24

LABELLED COMPOUND	LAB FLAG ¹	MZ's FORMING RATIO ²	ION ABUND. RATIO	QC LIMITS ³	CONC. FOUND (ng/mL)	CONC. RANGE (ng/mL) ⁴
13C-2,3,7,8-TCDD		M/M+2	0.77	0.65-0.89	101	82 - 121
13C-1,2,3,7,8-PECDD ⁵		M/M+2	0.63	0.52-0.70	96.4	62 - 160
13C-1,2,3,4,7,8-HXCDD		M+2/M+4	1.27	1.05-1.43	100	85 - 117
13C-1,2,3,6,7,8-HXCDD		M+2/M+4	1.25	1.05-1.43	101	85 - 118
13C-1,2,3,4,6,7,8-HPCDD		M+2/M+4	1.05	0.88-1.20	99.1	72 - 138
13C-OCDD		M+2/M+4	0.90	0.76-1.02	203	96 - 415
13C-2,3,7,8-TCDF		M/M+2	0.78	0.65-0.89	98.8	71 - 140
13C-1,2,3,7,8-PECDF		M+2/M+4	1.58	1.32-1.78	92.4	76 - 130
13C-2,3,4,7,8-PECDF		M+2/M+4	1.59	1.32-1.78	92.9	77 - 130
13C-1,2,3,4,7,8-HXCDF		M/M+2	0.52	0.43-0.59	95.0	76 - 131
13C-1,2,3,6,7,8-HXCDF		M/M+2	0.52	0.43-0.59	96.0	70 - 143
13C-1,2,3,7,8,9-HXCDF		M/M+2	0.52	0.43-0.59	98.3	74 - 135
13C-2,3,4,6,7,8-HXCDF		M/M+2	0.50	0.43-0.59	94.8	73 - 137
13C-1,2,3,4,6,7,8-HPCDF		M/M+2	0.44	0.37-0.51	95.4	78 - 129
13C-1,2,3,4,7,8,9-HPCDF		M/M+2	0.45	0.37-0.51	97.8	77 - 129
CLEANUP STANDARD						
37CL-2,3,7,8-TCDD ⁶					9.53	7.9 - 12.7

(1) Where applicable, custom lab flags have been used on this report.

(2) See Table 8, Method 1613, for m/z specifications.

(3) Ion Abundance Ratio Control Limits as specified in Table 9, Method 1613.

(4) Contract-required concentration range as determined from the percent of the test concentration in Table 6, Method 1613, under VER.

(5) Alternate confirmation and quantitation ions used for native and labeled PECDD.

(6) No ion abundance ratio for 37Cl4-2,3,7,8-TCDD; concentration reported.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: Thong DoFor Axys Internal Use Only [XSL Template: Form4B.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DX8M_127AS24_Form4B_SJ2457403.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 6A
PCDD/PCDF RELATIVE RETENTION TIMES

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
 V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_127A S: 24

Instrument ID: HR GC/MS

Analysis Date: 30-Oct-2018

GC Column ID: DB5

Analysis Time: 05:17:24

COMPOUND	LAB FLAG ¹	RETENTION TIME REFERENCE	RRT	RRT QC LIMITS ²
2,3,7,8-TCDD		13C-2,3,7,8-TCDD	1.001	0.999-1.002
1,2,3,7,8-PECDD ³		13C-1,2,3,7,8-PECDD	1.001	0.999-1.002
1,2,3,4,7,8-HXCDD		13C-1,2,3,4,7,8-HXCDD	1.000	0.999-1.001
1,2,3,6,7,8-HXCDD		13C-1,2,3,6,7,8-HXCDD	1.000	0.998-1.004
1,2,3,7,8,9-HXCDD		13C-1,2,3,6,7,8-HXCDD	1.010	1.000-1.019
1,2,3,4,6,7,8-HPCDD		13C-1,2,3,4,6,7,8-HPCDD	1.000	0.999-1.001
OCDD		13C-OCDD	1.000	0.999-1.001
2,3,7,8-TCDF		13C-2,3,7,8-TCDF	1.001	0.999-1.003
1,2,3,7,8-PECDF		13C-1,2,3,7,8-PECDF	1.001	0.999-1.002
2,3,4,7,8-PECDF		13C-2,3,4,7,8-PECDF	1.000	0.999-1.002
1,2,3,4,7,8-HXCDF		13C-1,2,3,4,7,8-HXCDF	1.001	0.999-1.001
1,2,3,6,7,8-HXCDF		13C-1,2,3,6,7,8-HXCDF	1.000	0.997-1.005
1,2,3,7,8,9-HXCDF		13C-1,2,3,7,8,9-HXCDF	1.001	0.999-1.001
2,3,4,6,7,8-HXCDF		13C-2,3,4,6,7,8-HXCDF	1.001	0.999-1.001
1,2,3,4,6,7,8-HPCDF		13C-1,2,3,4,6,7,8-HPCDF	1.000	0.999-1.001
1,2,3,4,7,8,9-HPCDF		13C-1,2,3,4,7,8,9-HPCDF	1.000	0.999-1.001
OCDF		13C-OCDD	1.002	0.999-1.008

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for Relative Retention Times (RRT) as specified in Table 2, Method 1613.

(3) Alternate confirmation and quantitation ions used for native and labeled PECDD.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ **Thong Do** _____

For Axys Internal Use Only [XSL Template: Form6A.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
 Report Filename: 1613_DIOXINS_DX8M_127AS24__Form6A_SJ2457403.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 6B
PCDD/PCDF RELATIVE RETENTION TIMES

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 24-Oct-2018

VER Data Filename: DX8M_127A S: 24

Instrument ID: HR GC/MS

Analysis Date: 30-Oct-2018

GC Column ID: DB5

Analysis Time: 05:17:24

LAB FLAG ¹	RETENTION TIME REFERENCE	RRT	RRT QC LIMITS ²
LABELED COMPOUND			
13C-2,3,7,8-TCDD	13C-1,2,3,4-TCDD	1.013	0.976-1.043
13C-1,2,3,7,8-PECDD	13C-1,2,3,4-TCDD	1.382	1.000-1.567
13C-1,2,3,4,7,8-HXCDD	13C-1,2,3,7,8,9-HXCDD	0.987	0.977-1.000
13C-1,2,3,6,7,8-HXCDD	13C-1,2,3,7,8,9-HXCDD	0.990	0.981-1.003
13C-1,2,3,4,6,7,8-HPCDD	13C-1,2,3,7,8,9-HXCDD	1.095	1.086-1.110
13C-OCDD	13C-1,2,3,7,8,9-HXCDD	1.179	1.032-1.311
13C-2,3,7,8-TCDF	13C-1,2,3,4-TCDD	0.966	0.923-1.103
13C-1,2,3,7,8-PECDF	13C-1,2,3,4-TCDD	1.283	1.000-1.425
13C-2,3,4,7,8-PECDF	13C-1,2,3,4-TCDD	1.351	1.011-1.526
13C-1,2,3,4,7,8-HXCDF	13C-1,2,3,7,8,9-HXCDD	0.954	0.944-0.970
13C-1,2,3,6,7,8-HXCDF	13C-1,2,3,7,8,9-HXCDD	0.958	0.949-0.975
13C-1,2,3,7,8,9-HXCDF	13C-1,2,3,7,8,9-HXCDD	1.004	0.977-1.047
13C-2,3,4,6,7,8-HXCDF	13C-1,2,3,7,8,9-HXCDD	0.980	0.959-1.021
13C-1,2,3,4,6,7,8-HPCDF	13C-1,2,3,7,8,9-HXCDD	1.062	1.043-1.085
13C-1,2,3,4,7,8,9-HPCDF	13C-1,2,3,7,8,9-HXCDD	1.104	1.057-1.151
CLEANUP STANDARD			
37CL-2,3,7,8-TCDD	13C-1,2,3,4-TCDD	1.001	0.989-1.052

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for Relative Retention Times (RRT) as specified in Table 2, Method 1613.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ **Thong Do** _____

For Axys Internal Use Only [XSL Template: Form6B.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DX8M_127AS24_Form6B_SJ2457403.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 5

PCDD/PCDF RT WINDOW AND ISOMER SPECIFICITY STANDARDS

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Instrument ID:	HR GC/MS	Initial Calibration Date:	24-Oct-2018
RT Window Data Filename:	DX8M_127A S: 1	Analysis Date:	29-Oct-2018
DB-5 IS Data Filename:	DX8M_127A S: 1	Analysis Date:	29-Oct-2018
DB-225 IS Data Filename:		Analysis Date:	
		Time:	07:59:05
		Time:	07:59:05
		Time:	

DB5 RT WINDOW DEFINING STANDARDS RESULT

ISOMERS	ABSOLUTE RT	ISOMERS	ABSOLUTE RT
1,3,6,8-TCDD (F)	23:01	1,3,6,8-TCDF (F)	21:29
1,2,8,9-TCDD (L)	28:24	1,2,8,9-TCDF (L)	28:14
1,2,4,7,9-PECDD (F)	32:08	1,3,4,6,8-PECDF (F)	28:57
1,2,3,8,9-PECDD (L)	37:07	1,2,3,8,9-PECDF (L)	37:10
1,2,4,6,7,9-HXCDD (F)	40:05	1,2,3,4,6,8-HXCDF (F)	39:02
1,2,3,4,6,7-HXCDD (L)	42:43	1,2,3,4,8,9-HXCDF (L)	43:03
1,2,3,4,6,7,9-HPCDD (F)	45:50	1,2,3,4,6,7,8-HPCDF (F)	45:21
1,2,3,4,6,7,8-HPCDD (L)	46:44	1,2,3,4,7,8,9-HPCDF (L)	47:08

(F) = First eluting isomer (DB-5); (L) = Last eluting isomer (DB-5)

ISOMER SPECIFICITY (IS) TEST STANDARDS RESULT

Isomers	% Valley Height Between Compared Peaks	Isomers	% Valley Height Between Compared Peaks
1,2,3,4-TCDD 1,2,7,8-TCDD	0	1,2,3,8-TCDD 2,3,7,8-TCDD	9
1,2,7,8-TCDD 1,4,7,8-TCDD	0	2,3,4,7-TCDF 2,3,7,8-TCDF	N/A
1,4,7,8-TCDD 1,2,3,7-TCDD	0	2,3,7,8-TCDF 1,2,3,9-TCDF	N/A
1,2,3,7-TCDD 1,2,3,8-TCDD	DB-5 column; co-elute as per Figure 6 in Method		

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ Thong Do _____

For Axys Internal Use Only [XSL Template: DXForm5.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS__DX8M_127AS1_Form5_SJ2456927.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 5

PCDD/PCDF RT WINDOW AND ISOMER SPECIFICITY STANDARDS

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Instrument ID:	HR GC/MS	Initial Calibration Date:	24-Oct-2018
RT Window Data Filename:	DX8M_127A S: 12	Analysis Date:	29-Oct-2018
DB-5 IS Data Filename:	DX8M_127A S: 12	Analysis Date:	29-Oct-2018
DB-225 IS Data Filename:		Analysis Date:	
		Time:	18:07:09
		Time:	18:07:09
		Time:	

DB5 RT WINDOW DEFINING STANDARDS RESULT

ISOMERS	ABSOLUTE RT	ISOMERS	ABSOLUTE RT
1,3,6,8-TCDD (F)	22:59	1,3,6,8-TCDF (F)	21:27
1,2,8,9-TCDD (L)	28:21	1,2,8,9-TCDF (L)	28:11
1,2,4,7,9-PECDD (F)	32:06	1,3,4,6,8-PECDF (F)	28:55
1,2,3,8,9-PECDD (L)	37:05	1,2,3,8,9-PECDF (L)	37:08
1,2,4,6,7,9-HXCDD (F)	40:03	1,2,3,4,6,8-HXCDF (F)	39:00
1,2,3,4,6,7-HXCDD (L)	42:42	1,2,3,4,8,9-HXCDF (L)	43:02
1,2,3,4,6,7,9-HPCDD (F)	45:49	1,2,3,4,6,7,8-HPCDF (F)	45:20
1,2,3,4,6,7,8-HPCDD (L)	46:43	1,2,3,4,7,8,9-HPCDF (L)	47:07

(F) = First eluting isomer (DB-5); (L) = Last eluting isomer (DB-5)

ISOMER SPECIFICITY (IS) TEST STANDARDS RESULT

Isomers	% Valley Height Between Compared Peaks	Isomers	% Valley Height Between Compared Peaks
1,2,3,4-TCDD 1,2,7,8-TCDD	0	1,2,3,8-TCDD 2,3,7,8-TCDD	9
1,2,7,8-TCDD 1,4,7,8-TCDD	0	2,3,4,7-TCDF 2,3,7,8-TCDF	N/A
1,4,7,8-TCDD 1,2,3,7-TCDD	0	2,3,7,8-TCDF 1,2,3,9-TCDF	N/A
1,2,3,7-TCDD 1,2,3,8-TCDD	DB-5 column; co-elute as per Figure 6 in Method		

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ Thong Do _____

For Axy Internal Use Only [XSL Template: DXForm5.xsl; Created: 21-Nov-2018 17:03:21; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS__DX8M_127AS12_Form5_SJ2457442.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 4A
PCDD/PCDF CALIBRATION VERIFICATION

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
 V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date:	10-Jul-2018	VER Data Filename:	DB83_183 S: 2
Instrument ID:	HR GC/MS	Analysis Date:	23-Oct-2018
GC Column ID:	DB225	Analysis Time:	20:13:08

COMPOUND	LAB FLAG ¹	MZ's FORMING RATIO ²	ION ABUND. RATIO	QC LIMITS ³	CONC. FOUND (ng/mL)	CONC. RANGE (ng/mL) ⁴
1,2,3,7,8,9-HXCDD		M+2/M+4	1.21	1.05-1.43	48.8	41 - 61
2,3,7,8-TCDF		M/M+2	0.84	0.65-0.89	9.34	8.4 - 12

(1) Where applicable, custom lab flags have been used on this report.

(2) See Table 8, Method 1613, for m/z specifications.

(3) Ion Abundance Ratio Control Limits as specified in Table 9, Method 1613.

(4) Contract-required concentration range as determined from the percent of the test concentration in Table 6, Method 1613, under VER.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ **Ting Chen** _____

For Axys Internal Use Only [XSL Template: Form4A.xsl; Created: 21-Nov-2018 17:04:30; Application: XMLTransformer-1.16.51;
 Report Filename: 1613_DIOXINS_DB83_183S2__Form4A_SJ2455690.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 6A
PCDD/PCDF RELATIVE RETENTION TIMES

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 10-Jul-2018

VER Data Filename: DB83_183 S: 2

Instrument ID: HR GC/MS

Analysis Date: 23-Oct-2018

GC Column ID: DB225

Analysis Time: 20:13:08

COMPOUND	LAB FLAG ¹	RETENTION TIME REFERENCE	RRT	RRT QC LIMITS ²
1,2,3,7,8,9-HXCDD		13C-1,2,3,6,7,8-HXCDD	1.064	1.052-1.075
2,3,7,8-TCDF		13C-2,3,7,8-TCDF	1.001	0.999-1.003

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for Relative Retention Times (RRT) as specified in Table 2, Method 1613.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____Ting Chen_____

For Axs Internal Use Only [XSL Template: Form6A.xsl; Created: 21-Nov-2018 17:04:30; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DB83_183S2__Form6A_SJ2455690.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 4A
PCDD/PCDF CALIBRATION VERIFICATION

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 10-Jul-2018

VER Data Filename: DB83_183 S: 15

Instrument ID: HR GC/MS

Analysis Date: 24-Oct-2018

GC Column ID: DB225

Analysis Time: 05:09:16

COMPOUND	LAB FLAG ¹	MZ's FORMING RATIO ²	ION ABUND. RATIO	QC LIMITS ³	CONC. FOUND (ng/mL)	CONC. RANGE (ng/mL) ⁴
1,2,3,7,8,9-HXCDD		M+2/M+4	1.24	1.05-1.43	48.0	41 - 61
2,3,7,8-TCDF		M/M+2	0.78	0.65-0.89	9.44	8.4 - 12

(1) Where applicable, custom lab flags have been used on this report.

(2) See Table 8, Method 1613, for m/z specifications.

(3) Ion Abundance Ratio Control Limits as specified in Table 9, Method 1613.

(4) Contract-required concentration range as determined from the percent of the test concentration in Table 6, Method 1613, under VER.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ Thong Do _____

For Axys Internal Use Only [XSL Template: Form4A.xsl; Created: 21-Nov-2018 17:04:30; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DB83_183S15__Form4A_SJ2455703.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 6A
PCDD/PCDF RELATIVE RETENTION TIMES

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Initial Calibration Date: 10-Jul-2018

VER Data Filename: DB83_183 S: 15

Instrument ID: HR GC/MS

Analysis Date: 24-Oct-2018

GC Column ID: DB225

Analysis Time: 05:09:16

COMPOUND	LAB FLAG ¹	RETENTION TIME REFERENCE	RRT	RRT QC LIMITS ²
1,2,3,7,8,9-HXCDD		13C-1,2,3,6,7,8-HXCDD	1.064	1.052-1.075
2,3,7,8-TCDF		13C-2,3,7,8-TCDF	1.001	0.999-1.003

(1) Where applicable, custom lab flags have been used on this report.

(2) Contract-required limits for Relative Retention Times (RRT) as specified in Table 2, Method 1613.

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ Thong Do _____

For Axs Internal Use Only [XSL Template: Form6A.xsl; Created: 21-Nov-2018 17:04:30; Application: XMLTransformer-1.16.51;
Report Filename: 1613_DIOXINS_DB83_183S15__Form6A_SJ2455703.html; Workgroup: WG65583; Design ID: 3359]

SGS AXYS METHOD MLA-017 Rev 20

Form 5

PCDD/PCDF RT WINDOW AND ISOMER SPECIFICITY STANDARDS

SGS AXYS ANALYTICAL SERVICES

2045 MILLS RD., SIDNEY, B.C., CANADA
V8L 5X2 TEL (250) 655-5800 FAX (250) 655-5811

Instrument ID:	HR GC/MS	Initial Calibration Date:	10-Jul-2018
RT Window Data Filename:		Analysis Date:	Time:
DB-5 IS Data Filename:		Analysis Date:	Time:
DB-225 IS Data Filename:	DB83_183 S: 1	Analysis Date:	23-Oct-2018
		Time:	19:31:52

DB225 RT WINDOW DEFINING STANDARDS RESULT

ISOMERS	ABSOLUTE RT	ISOMERS	ABSOLUTE RT
1,3,6,8-TCDD (F)	N/A	1,3,6,8-TCDF (F)	N/A
1,2,8,9-TCDD (L)	N/A	1,2,8,9-TCDF (L)	N/A
1,2,4,7,9-PECDD (F)	N/A	1,3,4,6,8-PECDF (F)	N/A
1,2,3,8,9-PECDD (L)	N/A	1,2,3,8,9-PECDF (L)	N/A
1,2,4,6,7,9-HXCDD (F)	N/A	1,2,3,4,6,8-HXCDF (F)	N/A
1,2,3,4,6,7-HXCDD (L)	N/A	1,2,3,4,8,9-HXCDF (L)	N/A
1,2,3,4,6,7,9-HPCDD (F)	N/A	1,2,3,4,6,7,8-HPCDF (F)	N/A
1,2,3,4,6,7,8-HPCDD (L)	N/A	1,2,3,4,7,8,9-HPCDF (L)	N/A

(F) = First eluting isomer (DB-5); (L) = Last eluting isomer (DB-5)

ISOMER SPECIFICITY (IS) TEST STANDARDS RESULT

Isomers	% Valley Height Between Compared Peaks	Isomers	% Valley Height Between Compared Peaks
1,2,3,4-TCDD 1,2,7,8-TCDD	N/A	1,2,3,8-TCDD 2,3,7,8-TCDD	N/A
1,2,7,8-TCDD 1,4,7,8-TCDD	N/A	2,3,4,7-TCDF 2,3,7,8-TCDF	10
1,4,7,8-TCDD 1,2,3,7-TCDD	N/A	2,3,7,8-TCDF 1,2,3,9-TCDF	20
1,2,3,7-TCDD 1,2,3,8-TCDD	N/A		

These data are validated and reported as accurate and in accord with SGS AXYS Analytical Services Ltd. ISO17025 compliant quality assurance processes.

Signed: _____ **Thong Do** _____

For Axys Internal Use Only [XSL Template: DXForm5.xsl; Created: 21-Nov-2018 17:04:30; Application: XMLTransformer-1.16.51; Report Filename: 1613_DIOXINS__DB83_183S1_Form5_SJ2455723.html; Workgroup: WG65583; Design ID: 3359]

Accreditation Scope																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																														
---------------------	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--

Accreditation Scope

SGS AXYS Analytical Services Ltd.
file ref.: ACC-101 Rev. 40

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB ISO 17025	ANAB DoD **	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB ISO 17025	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE *	Maine DOH	Pennsylvania DEP	ANAB ISO 17025	ANAB DoD **
		EPA 1699	MLA-028				Y							Y											Y								Y	
		SGS AXYS MLA-028	MLA-028	Y	Y		Y					Y		Y		Y							Y	Y						Y		Y		
		SGS AXYS MLA-007	MLA-007	Y	Y		Y								Y		Y							Y	Y							Y		
	4,4'-DDT	EPA 625	MLA-007																							Y		Y	Y		Y		Y	
		EPA 8270	MLA-007				Y				Y	Y	Y	Y	Y																			
		EPA 1699	MLA-028				Y							Y												Y							Y	
		SGS AXYS MLA-028	MLA-028	Y	Y		Y						Y		Y		Y						Y	Y					Y			Y		
		SGS AXYS MLA-007	MLA-007	Y	Y		Y								Y		Y						Y	Y								Y		
		Aldrin	EPA 625	MLA-007																							Y		Y	Y		Y		Y
	EPA 8270		MLA-007				Y				Y	Y	Y	Y	Y																			
	EPA 1699		MLA-028				Y							Y												Y							Y	
	SGS AXYS MLA-028		MLA-028	Y	Y		Y						Y		Y		Y						Y	Y					Y			Y		
	Alpha-HCH	SGS AXYS MLA-007	MLA-007	Y	Y		Y								Y		Y						Y	Y									Y	
		EPA 625	MLA-007																							Y		Y		Y	Y		Y	
		EPA 8270	MLA-007				Y				Y	Y	Y	Y	Y																			
		EPA 1699	MLA-028				Y							Y												Y							Y	
	Beta-HCH	SGS AXYS MLA-028	MLA-028	Y	Y		Y						Y		Y		Y						Y	Y									Y	
		SGS AXYS MLA-007	MLA-007	Y	Y		Y								Y		Y						Y	Y									Y	
		EPA 625	MLA-007				Y				Y	Y	Y	Y	Y											Y		Y	Y		Y		Y	
		EPA 8270	MLA-007				Y				Y	Y	Y	Y	Y																			
	Chlordane, technical	EPA 1699	MLA-028				Y								Y											Y								Y
		SGS AXYS MLA-028	MLA-028	Y	Y		Y						Y		Y		Y						Y	Y					Y				Y	
		SGS AXYS MLA-007	MLA-007	Y	Y		Y								Y		Y						Y	Y									Y	
		EPA 625	MLA-007																												Y			
	cis-Chlordane (alpha-Chlordane)	EPA 8270	MLA-007				Y				Y	Y		Y	Y											Y		Y	Y		Y		Y	
		EPA 1699	MLA-028				Y								Y										Y								Y	
		SGS AXYS MLA-028	MLA-028	Y	Y		Y						Y		Y		Y						Y	Y				Y					Y	
		SGS AXYS MLA-007	MLA-007	Y	Y		Y								Y		Y						Y	Y									Y	
	cis-Nonachlor	EPA 8270	MLA-007				Y					Y		Y												Y								Y
		EPA 1699	MLA-028				Y								Y										Y								Y	
		SGS AXYS MLA-028	MLA-028	Y	Y		Y						Y		Y		Y						Y	Y				Y					Y	
		SGS AXYS MLA-007	MLA-007	Y	Y		Y								Y		Y						Y	Y									Y	
	Delta-HCH	EPA 608	MLA-007																							Y			Y	Y		Y		Y
		EPA 8081	MLA-007				Y				Y	Y	Y	Y	Y																			
		EPA 1699	MLA-028				Y								Y										Y								Y	
		SGS AXYS MLA-028	MLA-028	Y	Y		Y						Y		Y		Y						Y	Y				Y					Y	
		SGS AXYS MLA-007	MLA-007	Y	Y		Y								Y		Y						Y	Y									Y	
	Dieldrin	EPA 608	MLA-007																							Y			Y	Y		Y		Y
		EPA 8081	MLA-007				Y				Y	Y	Y	Y	Y																			
		EPA 1699	MLA-028				Y								Y										Y								Y	
		SGS AXYS MLA-028	MLA-028	Y	Y		Y						Y		Y		Y						Y	Y				Y					Y	
		SGS AXYS MLA-007	MLA-007	Y	Y		Y								Y		Y						Y	Y									Y	
	Endosulphan I	EPA 608	MLA-007																							Y			Y	Y		Y		Y
		EPA 8081	MLA-007				Y				Y	Y	Y	Y	Y																			
		EPA 1699	MLA-028				Y								Y										Y								Y	
		SGS AXYS MLA-028	MLA-028	Y	Y		Y						Y		Y		Y						Y	Y				Y					Y	
		SGS AXYS MLA-007	MLA-007	Y	Y		Y								Y		Y						Y	Y									Y	
	Endosulphan II	EPA 608	MLA-007																							Y			Y	Y		Y		Y

Accreditation Scope

SGS AXYS Analytical Services Ltd.
file ref.: ACC-101 Rev. 40

Accreditation Scope																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																							
---------------------	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--

Accreditation Scope																																																															
SGS AXYS Analytical Services Ltd. file ref.: ACC-101 Rev. 40																																																															
				Serum										Solids										Tissue										Urine										Water										Water, Non-Potable									
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB ISO 17025	ANAB DoD **	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB ISO 17025	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE *	Maine DOH	Pennsylvania DEP	ANAB ISO 17025	ANAB DoD **																													
		SGS AXYS MLA-028	MLA-028	Y	Y	Y					Y			Y		Y							Y	Y	Y				Y			Y																															
		SGS AXYS MLA-007	MLA-007	Y	Y	Y								Y		Y							Y	Y	Y							Y																															
	Oxychlordane	EPA 8270	MLA-007			Y						Y													Y								Y																														
		EPA 1699	MLA-028			Y								Y											Y								Y																														
		SGS AXYS MLA-028	MLA-028	Y	Y	Y						Y	Y		Y		Y						Y	Y	Y				Y			Y																															
		SGS AXYS MLA-007	MLA-007	Y	Y	Y										Y		Y					Y	Y	Y							Y																															
	Toxaphene	EPA 8270	MLA-007									Y														Y								Y																													
		SGS AXYS MLA-007	MLA-007		Y												Y						Y																																								
	trans-Chlordane (gamma-Chlordane)	EPA 8270	MLA-007				Y			Y		Y	Y	Y												Y		Y			Y		Y																														
		EPA 1699	MLA-028				Y							Y											Y								Y																														
		SGS AXYS MLA-028	MLA-028	Y	Y	Y						Y	Y		Y		Y						Y	Y	Y				Y			Y																															
		SGS AXYS MLA-007	MLA-007	Y	Y	Y									Y		Y						Y	Y	Y							Y																															
	trans-Nonachlor	EPA 8270	MLA-007				Y					Y	Y												Y								Y																														
		EPA 1699	MLA-028				Y							Y											Y								Y																														
		SGS AXYS MLA-028	MLA-028	Y	Y	Y						Y	Y		Y		Y						Y	Y	Y				Y			Y																															
		SGS AXYS MLA-007	MLA-007	Y	Y	Y									Y		Y						Y	Y	Y							Y																															
PAH	1,2,6-Trimethylphenanthrene	SGS AXYS MLA-021	MLA-021		Y																	Y																																									
	1,2-Dimethylnaphthalene	SGS AXYS MLA-021	MLA-021		Y																	Y																																									
	1,4,6,7-Tetramethylnaphthalene	SGS AXYS MLA-021	MLA-021		Y																	Y																																									
	1,7-Dimethylfluorene	SGS AXYS MLA-021	MLA-021		Y																	Y																																									
	1,7-Dimethylphenanthrene	SGS AXYS MLA-021	MLA-021		Y																	Y																																									
	1,8-Dimethylphenanthrene	SGS AXYS MLA-021	MLA-021		Y																	Y																																									
	1-Methylchrysene	SGS AXYS MLA-021	MLA-021		Y																	Y																																									
	1-Methylnaphthalene	SGS AXYS MLA-021	MLA-021		Y																	Y																																									
	1-Methylphenanthrene	SGS AXYS MLA-021	MLA-021		Y																	Y																																									
	2,3,5-Trimethylnaphthalene	SGS AXYS MLA-021	MLA-021		Y																	Y																																									
	2,3,6-Trimethylnaphthalene	SGS AXYS MLA-021	MLA-021		Y																	Y																																									
	2,4-Dimethyldibenzothiophene	SGS AXYS MLA-021	MLA-021		Y																	Y																																									
	2,6-Dimethylnaphthalene	SGS AXYS MLA-021	MLA-021		Y																	Y																																									
	2,6-Dimethylphenanthrene	SGS AXYS MLA-021	MLA-021		Y																	Y																																									
	2-Methylantracene	SGS AXYS MLA-021	MLA-021		Y																	Y																																									
	2-Methyldibenzothiophene	SGS AXYS MLA-021	MLA-021		Y																	Y																																									
	2-Methylfluorene	SGS AXYS MLA-021	MLA-021		Y																		Y																																								
	2-Methylnaphthalene	EPA 1625	MLA-021																															Y																													
		EPA 8270	MLA-021				Y			Y					Y											Y																																					
		SGS AXYS MLA-021	MLA-021		Y	Y									Y								Y										Y																														
		SGS AXYS MLA-021	MLA-021		Y																		Y																																								
	2-Methylphenanthrene	SGS AXYS MLA-021	MLA-021		Y																		Y																																								
	3,6-Dimethylphenanthrene	SGS AXYS MLA-021	MLA-021		Y																		Y																																								
	3-Methyldibenzothiophene	SGS AXYS MLA-021	MLA-021		Y																		Y																																								
	3-Methylfluoranthene/ Benzo(a)fluorene	SGS AXYS MLA-021	MLA-021		Y																		Y																																								
	3-Methylphenanthrene	SGS AXYS MLA-021	MLA-021		Y																		Y																																								
	5,9-Dimethylchrysene	SGS AXYS MLA-021	MLA-021		Y																		Y																																								
	5/6-Methylchrysenes	SGS AXYS MLA-021	MLA-021		Y																		Y																																								
	7-Methylbenzo(a)pyrene	SGS AXYS MLA-021	MLA-021		Y																		Y																																								
	9/4-Methylphenanthrenes	SGS AXYS MLA-021	MLA-021		Y																		Y																																								
	Acenaphthene	EPA 1625	MLA-021																						Y		Y	Y					Y																														
		EPA 8270	MLA-021				Y			Y	Y			Y	Y													Y																																			
		SGS AXYS MLA-021	MLA-021		Y	Y									Y		Y						Y		Y							Y																															
		EPA 8270	MLA-021				Y			Y	Y			Y	Y												Y	Y				Y																															

Accreditation Scope				Serum		Solids		Tissue		Urine		Water		Water, Non-Potable	
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB ISO 17025	ANAB DoD **
Anthracene		SGS AXYS MLA-021	MLA-021		Y	Y									
		EPA 1625	MLA-021												
		EPA 8270	MLA-021			Y			Y	Y					
		SGS AXYS MLA-021	MLA-021	Y	Y					Y				Y	
Benz[a]anthracene		EPA 1625	MLA-021												
		EPA 8270	MLA-021			Y			Y	Y					
		SGS AXYS MLA-021	MLA-021	Y	Y					Y				Y	
		EPA 1625	MLA-021												
Benzo[a]pyrene		EPA 1625	MLA-021												
		EPA 8270	MLA-021			Y			Y	Y					
		SGS AXYS MLA-021	MLA-021	Y	Y					Y				Y	
		EPA 1625	MLA-021												
Benzo[b]fluoranthene		EPA 1625	MLA-021												
		EPA 8270	MLA-021			Y			Y	Y					
		SGS AXYS MLA-021	MLA-021	Y	Y					Y				Y	
		EPA 1625	MLA-021												
Benzo[e]pyrene		SGS AXYS MLA-021	MLA-021	Y											
		EPA 1625	MLA-021												
		EPA 8270	MLA-021			Y			Y	Y					
		SGS AXYS MLA-021	MLA-021	Y	Y					Y				Y	
Benzo[ghi]perylene		SGS AXYS MLA-021	MLA-021	Y											
		EPA 1625	MLA-021												
		EPA 8270	MLA-021			Y			Y	Y					
		SGS AXYS MLA-021	MLA-021	Y	Y					Y				Y	
Benzofluoranthenes		SGS AXYS MLA-021	MLA-021	Y											
		EPA 1625	MLA-021												
		EPA 8270	MLA-021			Y			Y	Y					
		SGS AXYS MLA-021	MLA-021			Y				Y				Y	
Biphenyl		SGS AXYS MLA-021	MLA-021	Y											
		EPA 1625	MLA-021												
		EPA 8270	MLA-021			Y			Y	Y					
		SGS AXYS MLA-021	MLA-021			Y				Y				Y	
C1-Acenaphthenes		SGS AXYS MLA-021	MLA-021	Y											
C1-Benz(a)anthracenes/chrysenes		SGS AXYS MLA-021	MLA-021	Y											
C1-Benzofluoranthenes/ Benzopyrenes		SGS AXYS MLA-021	MLA-021	Y											
C1-Biphenyls		SGS AXYS MLA-021	MLA-021	Y											
C1-Dibenzothiophene		SGS AXYS MLA-021	MLA-021	Y											
C1-Fluoranthenes/Pyrenes		SGS AXYS MLA-021	MLA-021	Y											
C1-Fluorenes		SGS AXYS MLA-021	MLA-021	Y											
C1-Naphthalenes		SGS AXYS MLA-021	MLA-021	Y											
C1-Phenanthrenes/Anthracenes		SGS AXYS MLA-021	MLA-021	Y											
C2-Benz(a)anthracenes/Chrysenes		SGS AXYS MLA-021	MLA-021	Y											
C2-Benzofluoranthenes/ Benzopyrenes		SGS AXYS MLA-021	MLA-021	Y											
C2-Biphenyls		SGS AXYS MLA-021	MLA-021	Y											
C2-Dibenzothiophene		SGS AXYS MLA-021	MLA-021	Y											
C2-Fluoranthenes/Pyrenes		SGS AXYS MLA-021	MLA-021	Y											
C2-Fluorenes		SGS AXYS MLA-021	MLA-021	Y											
C2-Naphthalenes		SGS AXYS MLA-021	MLA-021	Y											
C2-Phenanthrenes/Anthracenes		SGS AXYS MLA-021	MLA-021	Y											
C3-Benz(a)anthracenes/Chrysenes		SGS AXYS MLA-021	MLA-021	Y											
C3-Dibenzothiophene		SGS AXYS MLA-021	MLA-021	Y											
C3-Fluoranthenes/Pyrenes		SGS AXYS MLA-021	MLA-021	Y											
C3-Fluorenes		SGS AXYS MLA-021	MLA-021	Y											
C3-Naphthalenes		SGS AXYS MLA-021	MLA-021	Y											
C3-Phenanthrenes/Anthracenes		SGS AXYS MLA-021	MLA-021	Y											
C4-Benz(a)anthracenes/Chrysenes		SGS AXYS MLA-021	MLA-021	Y											
C4-Dibenzothiophene		SGS AXYS MLA-021	MLA-021	Y											
C4-Fluoranthenes/Pyrenes		SGS AXYS MLA-021	MLA-021	Y											
C4-Naphthalenes		SGS AXYS MLA-021	MLA-021	Y											
C4-Phenanthrenes/Anthracenes		SGS AXYS MLA-021	MLA-021	Y											

Accreditation ScopeSGS AXYS Analytical Services Ltd.
file ref.: ACC-101 Rev. 40

Accreditation Scope				Serum	Solids	Tissue	Urine	Water	Water, Non-Potable																									
SGS AXYS Analytical Services Ltd. file ref.: ACC-101 Rev. 40																																		
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB ISO 17025	ANAB DoD **	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB ISO 17025	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE *	Maine DOH	Pennsylvania DEP	ANAB ISO 17025	ANAB DoD **
	Chrysene	EPA 1625	MLA-021																						Y		Y	Y				Y		
		EPA 8270	MLA-021				Y				Y	Y		Y	Y																			
		SGS AXYS MLA-021	MLA-021		Y	Y									Y		Y						Y		Y							Y		
	Dibenz[a,h]anthracene	EPA 1625	MLA-021																						Y			Y	Y				Y	
		EPA 8270	MLA-021				Y				Y	Y		Y	Y									Y									Y	
		SGS AXYS MLA-021	MLA-021		Y	Y									Y		Y						Y		Y							Y		
	Dibenzothiophene	SGS AXYS MLA-021	MLA-021		Y																		Y											
	Fluoranthene	EPA 1625	MLA-021																						Y		Y	Y				Y		
		EPA 8270	MLA-021				Y				Y	Y		Y	Y									Y									Y	
		SGS AXYS MLA-021	MLA-021		Y	Y									Y		Y						Y		Y							Y		
	Fluorene	EPA 1625	MLA-021																						Y			Y	Y				Y	
		EPA 8270	MLA-021				Y				Y	Y		Y	Y										Y								Y	
		SGS AXYS MLA-021	MLA-021		Y	Y									Y		Y						Y		Y								Y	
	Indeno[1,2,3-cd]pyrene	EPA 1625	MLA-021																						Y			Y	Y				Y	
		EPA 8270	MLA-021				Y				Y	Y		Y	Y										Y								Y	
		SGS AXYS MLA-021	MLA-021		Y	Y									Y		Y						Y		Y								Y	
	Naphthalene	EPA 1625	MLA-021																						Y			Y	Y				Y	
		EPA 8270	MLA-021				Y				Y	Y		Y	Y																		Y	
		SGS AXYS MLA-021	MLA-021		Y	Y									Y		Y						Y		Y								Y	
	Perylene	SGS AXYS MLA-021	MLA-021		Y																		Y		Y								Y	
	Phenanthrene	EPA 1625	MLA-021																						Y			Y	Y				Y	
		EPA 8270	MLA-021				Y				Y	Y		Y	Y																		Y	
		SGS AXYS MLA-021	MLA-021		Y	Y									Y		Y						Y		Y								Y	
	Pyrene	EPA 1625	MLA-021																						Y			Y	Y				Y	
		EPA 8270	MLA-021				Y				Y	Y		Y	Y																		Y	
		SGS AXYS MLA-021	MLA-021		Y	Y									Y		Y						Y		Y								Y	
	Retene	SGS AXYS MLA-021	MLA-021		Y																			Y										
PBDPE	BDE 10 2,6-dibromodiphenylether	EPA 1614	MLA-033									Y																				Y		
		SGS AXYS MLA-033	MLA-033	Y	Y												Y						Y		Y									
	BDE 100 2,2',4,4',6-pentabromodiphenylether	EPA 1614	MLA-033										Y																			Y		
		SGS AXYS MLA-033	MLA-033	Y	Y													Y						Y		Y								
	BDE 105 2,3,3',4,4'-pentabromodiphenylether	EPA 1614	MLA-033										Y																			Y		
		SGS AXYS MLA-033	MLA-033	Y	Y													Y						Y		Y								
	BDE 11 3,3'-dibromodiphenylether	EPA 1614	MLA-033										Y																			Y		
		SGS AXYS MLA-033	MLA-033	Y	Y													Y						Y		Y						Y		
	BDE 116 2,3,4,5,6-pentabromodiphenylether	EPA 1614	MLA-033										Y																			Y		
		SGS AXYS MLA-033	MLA-033	Y	Y													Y						Y		Y								
	BDE 119 2,3',4,4',6-pentabromodiphenylether	EPA 1614	MLA-033										Y																			Y		
		SGS AXYS MLA-033	MLA-033	Y	Y													Y						Y		Y							Y	
	BDE 12 3,4-dibromodiphenylether	EPA 1614	MLA-033										Y																			Y		
		SGS AXYS MLA-033	MLA-033	Y	Y													Y						Y		Y								
	BDE 126 3,3',4,4',5-pentabromodiphenylether	EPA 1614	MLA-033										Y																			Y		
		SGS AXYS MLA-033	MLA-033	Y	Y													Y						Y		Y								
	BDE 13 3,4'-dibromodiphenylether	EPA 1614	MLA-033										Y																			Y		
		SGS AXYS MLA-033	MLA-033	Y	Y													Y						Y		Y								
BDE 140 2,2',3,4,4',6'-hexabromodiphenylether	EPA 1614	MLA-033										Y																			Y			
	SGS AXYS MLA-033	MLA-033	Y	Y													Y						Y		Y									
BDE 15 4,4'-dibromodiphenylether	EPA 1614	MLA-033										Y																			Y			
	SGS AXYS MLA-033	MLA-033	Y	Y													Y						Y		Y									
BDE 153 2,2',4,4',5,5'-hexabromodiphenylether	EPA 1614	MLA-033										Y																		Y				

Accreditation Scope																			Serum	Solids																Tissue		Urine	Water	Water, Non-Potable																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																											
SGS AXYS Analytical Services Ltd. file ref.: ACC-101 Rev. 40																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																			

Accreditation Scope																Serum									Tissue			Urine	Water	Water, Non-Potable							
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB ISO 17025	ANAB DoD **	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB ISO 17025	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE *	Maine DOH	Pennsylvania DEP	ANAB ISO 17025	ANAB DoD **			
	BDE-33 2',3,4-tribromodiphenylether	SGS AXYS MLA-033	MLA-033	Y	Y											Y							Y														
		EPA 1614	MLA-033										Y																	Y							
	"PCBs" category (CA only)	SGS AXYS MLA-033	MLA-033	Y	Y											Y							Y				Y										
		EPA 625	MLA-007																								Y										
	PCB Aroclor 1016	EPA 8270	MLA-007			Y																															
		EPA 1668	MLA-010				Y						Y	Y												Y					Y		Y				
		EPA 625	MLA-007																									Y			Y		Y				
		EPA 8270	MLA-007				Y			Y	Y			Y													Y		Y		Y		Y				
		SGS AXYS MLA-010	MLA-010				Y							Y												Y							Y				
		SGS AXYS MLA-007	MLA-007		Y		Y							Y		Y						Y	Y			Y						Y					
	PCB Aroclor 1016/1242	EPA 8270	MLA-007								Y												Y			Y							Y				
		EPA 1668	MLA-010				Y						Y	Y												Y				Y		Y					
	PCB Aroclor 1221	EPA 625	MLA-007																							Y					Y		Y				
		EPA 8270	MLA-007				Y			Y	Y	Y		Y											Y		Y	Y		Y		Y					
		SGS AXYS MLA-010	MLA-010				Y							Y											Y								Y				
		SGS AXYS MLA-007	MLA-007		Y		Y							Y		Y						Y	Y		Y							Y					
	PCB Aroclor 1232	EPA 1668	MLA-010				Y						Y	Y											Y					Y		Y					
		EPA 625	MLA-007																							Y					Y		Y				
		EPA 8270	MLA-007				Y			Y	Y	Y		Y													Y	Y			Y		Y				
		SGS AXYS MLA-010	MLA-010				Y							Y											Y								Y				
		SGS AXYS MLA-007	MLA-007		Y		Y							Y		Y							Y	Y		Y							Y				
		EPA 1668	MLA-010				Y							Y	Y										Y					Y		Y					
	PCB Aroclor 1242	EPA 625	MLA-007																						Y					Y		Y					
		EPA 8270	MLA-007				Y			Y	Y			Y											Y		Y	Y		Y		Y					
		SGS AXYS MLA-010	MLA-010				Y							Y											Y								Y				
		SGS AXYS MLA-007	MLA-007		Y		Y								Y		Y						Y	Y		Y							Y				
	PCB Aroclor 1248	EPA 1668	MLA-010				Y						Y	Y											Y					Y		Y					
		EPA 625	MLA-007																							Y					Y		Y				
		EPA 8270	MLA-007				Y			Y	Y	Y		Y													Y	Y			Y		Y				
		SGS AXYS MLA-010	MLA-010				Y								Y										Y								Y				
		SGS AXYS MLA-007	MLA-007		Y		Y							Y		Y							Y	Y		Y							Y				
		EPA 1668	MLA-010				Y							Y	Y										Y					Y		Y					
	PCB Aroclor 1254	EPA 625	MLA-007																							Y					Y		Y				
		EPA 8270	MLA-007				Y			Y	Y	Y		Y												Y	Y			Y		Y					
		SGS AXYS MLA-010	MLA-010				Y							Y											Y								Y				
		SGS AXYS MLA-007	MLA-007		Y		Y								Y		Y						Y	Y		Y							Y				
	PCB Aroclor 1260	EPA 1668	MLA-010				Y						Y	Y											Y					Y		Y					
		EPA 625	MLA-007																						Y								Y				
		EPA 8270	MLA-007				Y			Y	Y	Y		Y												Y	Y		Y		Y		Y				
		SGS AXYS MLA-010	MLA-010				Y								Y										Y								Y				
		SGS AXYS MLA-007	MLA-007		Y		Y							Y		Y						Y	Y		Y								Y				
		SGS AXYS MLA-007	MLA-007		Y												Y							Y	Y								Y				
	PCB Aroclor 1268	SGS AXYS MLA-007	MLA-007		Y											Y																					
	PCB 1 2-Chlorobiphenyl	EPA 1668	MLA-010				Y			Y	Y	Y	Y	Y	Y							Y				Y			Y		Y		Y				
		EPA 8270	MLA-007										Y																								
		SGS AXYS MLA-010	MLA-010	Y	Y	Y									Y		Y					Y	Y		Y							Y					
		EPA 1668	MLA-010				Y			Y	Y	Y	Y		Y								Y			Y		Y	Y		Y		Y				
	PCB 10 2,6-Dichlorobiphenyl	SGS AXYS MLA-010	MLA-010	Y	Y	Y									Y		Y					Y	Y		Y								Y				
		EPA 1668	MLA-010				Y			Y	Y	Y	Y		Y								Y			Y		Y	Y		Y		Y				
	PCB 100 2,2',4,4',6-Pentachlorobiphenyl	SGS AXYS MLA-010	MLA-010	Y	Y	Y									Y		Y					Y	Y		Y								Y				
		EPA 1668	MLA-010				Y			Y	Y	Y	Y	Y	Y								Y			Y		Y	Y		Y		Y				
		EPA 8270	MLA-007									Y																									
		SGS AXYS MLA-010	MLA-010	Y	Y	Y									Y		Y					Y	Y		Y								Y				

Accreditation Scope																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																
---------------------	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--

Accreditation Scope																Serum	Solids													Tissue		Urine	Water	Water, Non-Potable																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																											
SGS AXYS Analytical Services Ltd. file ref.: ACC-101 Rev. 40																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													

Accreditation Scope				Serum										Tissue					Urine		Water		Water, Non-Potable										
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Solids										Tissue					Urine		Water		Water, Non-Potable										
				CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB ISO 17025	ANAB DoD **	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB ISO 17025	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE *	Maine DOH	Pennsylvania DEP	ANAB ISO 17025
	PCB 158/160	EPA 8270	MLA-007									Y																					
		SGS AXYS MLA-007	MLA-007		Y											Y						Y		Y									
	PCB 159 2,3,3',4,5,5'-Hexachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y					Y						Y		Y	Y	Y	Y		Y	Y
		EPA 8270	MLA-007								Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y		Y		Y		Y	Y	Y								Y	
		SGS AXYS MLA-007	MLA-007		Y											Y						Y	Y										
	PCB 16 2,2',3-Trichlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y					Y						Y		Y	Y	Y	Y		Y	Y
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y		Y		Y		Y	Y		Y		Y					Y	
	PCB 16/32	EPA 8270	MLA-007									Y																					
		SGS AXYS MLA-007	MLA-007		Y											Y						Y	Y										
	PCB 160 2,3,3',4,5,6-Hexachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y					Y						Y		Y	Y	Y	Y		Y	Y
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y		Y		Y		Y	Y		Y							Y	
		SGS AXYS MLA-007	MLA-007		Y											Y						Y	Y										
	PCB 161 2,3,3',4,5',6-Hexachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y					Y						Y		Y	Y	Y	Y		Y	Y
		EPA 8270	MLA-007								Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y		Y		Y		Y	Y		Y							Y	
		SGS AXYS MLA-007	MLA-007		Y											Y						Y	Y										
	PCB 162 2,3,3',4',5,5'-Hexachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y					Y						Y		Y	Y	Y	Y		Y	Y
		EPA 8270	MLA-007								Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y		Y		Y		Y	Y		Y		Y					Y	
	PCB 163 2,3,3',4',5,6-Hexachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y					Y						Y		Y	Y	Y	Y		Y	Y
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y		Y		Y		Y	Y		Y		Y					Y	
	PCB 164 2,3,3',4',5',6-Hexachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y					Y						Y		Y	Y	Y	Y		Y	Y
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y		Y		Y		Y	Y		Y		Y					Y	
	PCB 165 2,3,3',5,5',6-Hexachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y					Y						Y		Y	Y	Y	Y		Y	Y
		EPA 8270	MLA-007								Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y		Y		Y		Y	Y		Y		Y					Y	
	PCB 166 2,3,4,4',5,6-Hexachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y					Y						Y		Y	Y	Y	Y		Y	Y
		EPA 8270	MLA-007								Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y		Y		Y		Y	Y		Y		Y					Y	
	PCB 167 2,3',4,4',5,5'-Hexachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y											Y		Y	Y	Y	Y		Y	Y
		EPA 8270	MLA-007								Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y		Y		Y		Y	Y		Y		Y					Y	
	PCB 168 2,3',4,4',5',6-Hexachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y					Y						Y		Y	Y	Y	Y		Y	Y
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y		Y		Y		Y	Y		Y		Y					Y	
		EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y					Y						Y		Y	Y	Y	Y		Y	Y
		SGS AXYS MLA-007	MLA-007		Y											Y						Y	Y										
	PCB 169 3,3',4,4',5,5'-Hexachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y					Y						Y		Y	Y	Y	Y		Y	Y
		EPA 8270	MLA-007								Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y		Y		Y		Y	Y		Y		Y					Y	
		SGS AXYS MLA-007	MLA-007		Y											Y						Y	Y										
	PCB 17 2,2',4-Trichlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y					Y						Y		Y	Y	Y	Y		Y	Y
		EPA 8270	MLA-007								Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y		Y		Y		Y	Y		Y		Y					Y	
		SGS AXYS MLA-007	MLA-007		Y											Y						Y	Y										
	PCB 170 2,2',3,3',4,4',5-Heptachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y					Y						Y		Y	Y	Y	Y		Y	Y
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y		Y		Y		Y	Y		Y		Y					Y	
		SGS AXYS MLA-901	MLA-901	Y																													
	PCB 170/190	EPA 8270	MLA-007								Y																						
		SGS AXYS MLA-007	MLA-007		Y											Y						Y	Y										
	PCB 171 2,2',3,3',4,4',6-Heptachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y					Y						Y		Y	Y	Y	Y		Y	Y
		EPA 8270	MLA-007									Y																					

Accreditation Scope				Serum										Tissue					Urine		Water		Water, Non-Potable											
SGS AXYS Analytical Services Ltd. file ref.: ACC-101 Rev. 40				Solids																														
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB ISO 17025	ANAB DoD **	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB ISO 17025	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE *	Maine DOH	Pennsylvania DEP	ANAB ISO 17025	ANAB DoD **
		SGS AXYS MLA-007	MLA-007		Y									Y		Y					Y	Y			Y	Y	Y	Y	Y		Y	Y		
	PCB 197 2,2',3,3',4,4',6,6'-Octachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y						Y				Y		Y	Y	Y	Y		Y	Y		
		EPA 8270	MLA-007									Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y			Y		Y		Y		Y							Y		
		SGS AXYS MLA-007	MLA-007		Y											Y							Y											
	PCB 198 2,2',3,3',4,5,5',6-Octachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y						Y				Y		Y	Y	Y	Y		Y	Y		
		EPA 8270	MLA-007									Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y			Y		Y		Y		Y							Y		
		SGS AXYS MLA-007	MLA-007		Y											Y							Y											
	PCB 199 2,2',3,3',4,5,5',6'-Octachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y						Y				Y		Y	Y	Y	Y		Y	Y		
		EPA 8270	MLA-007									Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y			Y		Y		Y		Y							Y		
		SGS AXYS MLA-007	MLA-007		Y											Y							Y											
	PCB 2 3-Chlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y						Y				Y		Y	Y	Y	Y		Y	Y		
		EPA 8270	MLA-007									Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y			Y		Y		Y		Y							Y		
		SGS AXYS MLA-007	MLA-007																															
	PCB 20 2,3,3'-Trichlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y						Y				Y		Y	Y	Y	Y		Y	Y		
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y			Y		Y		Y		Y									
	PCB 200 2,2',3,3',4,5,6,6'-Octachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y						Y				Y		Y	Y	Y	Y		Y	Y		
		EPA 8270	MLA-007									Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y			Y		Y		Y		Y							Y		
	PCB 201 2,2',3,3',4,5',6,6'-Octachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y						Y				Y		Y	Y	Y	Y		Y	Y		
		EPA 8270	MLA-007									Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y			Y		Y		Y		Y							Y		
		SGS AXYS MLA-007	MLA-007		Y											Y							Y											
	PCB 202 2,2',3,3',5,5',6,6'-Octachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y						Y				Y		Y	Y	Y	Y		Y	Y		
		EPA 8270	MLA-007									Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y			Y		Y		Y		Y							Y		
		SGS AXYS MLA-007	MLA-007		Y											Y							Y											
	PCB 203 2,2',3,4,4',5,5',6-Octachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y						Y				Y		Y	Y	Y	Y		Y	Y		
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y			Y		Y		Y		Y							Y		
		SGS AXYS MLA-007	MLA-007		Y											Y							Y											
	PCB 204 2,2',3,4,4',5,6,6'-Octachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y						Y				Y		Y	Y	Y	Y		Y	Y		
		EPA 8270	MLA-007									Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y			Y		Y		Y		Y							Y		
	PCB 205 2,3,3',4,4',5,5',6-Octachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y						Y				Y		Y	Y	Y	Y		Y	Y		
		EPA 8270	MLA-007									Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y			Y		Y		Y		Y							Y		
		SGS AXYS MLA-007	MLA-007		Y											Y							Y											
	PCB 206 2,2',3,3',4,4',5,5',6-Nonachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y						Y				Y		Y	Y	Y	Y		Y	Y		
		EPA 8270	MLA-007									Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y			Y		Y		Y		Y							Y		
		SGS AXYS MLA-007	MLA-007		Y											Y							Y											
	PCB 207 2,2',3,3',4,4',5,6,6'-Nonachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y						Y				Y		Y	Y	Y	Y		Y	Y		
		EPA 8270	MLA-007									Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y			Y		Y		Y		Y							Y		
		SGS AXYS MLA-007	MLA-007		Y											Y							Y											
	PCB 208 2,2',3,3',4,5,5',6,6'-Nonachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y						Y				Y		Y	Y	Y	Y		Y	Y		
		EPA 8270	MLA-007									Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y			Y		Y		Y		Y							Y		
		SGS AXYS MLA-007	MLA-007		Y											Y							Y											
	PCB 209 Decachlorobiphenyl	EPA 1668	MLA-010			Y		Y	Y	Y	Y	Y	Y	Y						Y				Y		Y	Y	Y	Y		Y	Y		

Accreditation Scope

SGS AXYS Analytical Services Ltd.
file ref.: ACC-101 Rev. 40

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Serum										Tissue					Urine		Water		Water, Non-Potable										
				CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB ISO 17025	ANAB DoD **	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB ISO 17025	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE *	Maine DOH	Pennsylvania DEP	ANAB ISO 17025
	PCB 49 2,2',4,5'-Tetrachlorobiphenyl	EPA 1668	MLA-010				Y														Y				Y								
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y					Y		Y		Y								Y
	PCB 49/43	EPA 8270	MLA-007									Y																					
		SGS AXYS MLA-007	MLA-007		Y											Y							Y										
	PCB 5 2,3-Dichlorobiphenyl	EPA 1668	MLA-010				Y		Y	Y	Y	Y	Y	Y							Y					Y	Y	Y	Y	Y		Y	Y
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y					Y		Y		Y								Y
	PCB 50 2,2',4,6-Tetrachlorobiphenyl	EPA 1668	MLA-010				Y		Y	Y	Y	Y	Y	Y							Y					Y	Y	Y	Y	Y		Y	Y
		EPA 8270	MLA-007									Y																					
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y					Y		Y		Y								Y
		SGS AXYS MLA-007	MLA-007		Y											Y						Y		Y									
	PCB 51 2,2',4,6'-Tetrachlorobiphenyl	EPA 1668	MLA-010				Y		Y	Y	Y	Y	Y	Y							Y					Y	Y	Y	Y	Y		Y	Y
		EPA 8270	MLA-007									Y																					
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y					Y		Y		Y								Y
	PCB 52 2,2',5,5'-Tetrachlorobiphenyl	EPA 1668	MLA-010				Y		Y	Y	Y	Y	Y	Y							Y				Y		Y	Y	Y	Y		Y	Y
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y					Y		Y		Y								Y
	PCB 52/73	EPA 8270	MLA-007									Y																					
		SGS AXYS MLA-007	MLA-007		Y											Y							Y										
	PCB 53 2,2',5,6'-Tetrachlorobiphenyl	EPA 1668	MLA-010				Y		Y	Y	Y	Y	Y	Y							Y					Y	Y	Y	Y	Y		Y	Y
		EPA 8270	MLA-007									Y																					
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y					Y		Y		Y								Y
	PCB 54 2,2',6,6'-Tetrachlorobiphenyl	EPA 1668	MLA-010				Y		Y	Y	Y	Y	Y	Y							Y					Y	Y	Y	Y	Y		Y	Y
		EPA 8270	MLA-007									Y																					
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y					Y		Y		Y								Y
	PCB 55 2,3,3',4-Tetrachlorobiphenyl	EPA 1668	MLA-010				Y		Y	Y	Y	Y	Y	Y							Y					Y	Y	Y	Y	Y		Y	Y
		EPA 8270	MLA-007									Y																					
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y					Y		Y		Y								Y
	PCB 56 2,3,3',4'-Tetrachlorobiphenyl	EPA 1668	MLA-010				Y		Y	Y	Y	Y	Y	Y							Y					Y	Y	Y	Y	Y		Y	Y
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y					Y		Y		Y								Y
	PCB 56/60	EPA 8270	MLA-007									Y																					
		SGS AXYS MLA-007	MLA-007		Y											Y							Y										
	PCB 57 2,3,3',5-Tetrachlorobiphenyl	EPA 1668	MLA-010				Y		Y	Y	Y	Y	Y	Y							Y					Y	Y	Y	Y	Y		Y	Y
		EPA 8270	MLA-007									Y																					
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y					Y		Y		Y								Y
	PCB 58 2,3,3',5'-Tetrachlorobiphenyl	EPA 1668	MLA-010				Y		Y	Y	Y	Y	Y	Y							Y					Y	Y	Y	Y	Y		Y	Y
		EPA 8270	MLA-007									Y																					
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y					Y		Y		Y								Y
	PCB 59 2,3,3',6-Tetrachlorobiphenyl	EPA 1668	MLA-010				Y		Y	Y	Y	Y	Y	Y							Y					Y	Y	Y	Y	Y		Y	Y
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y					Y		Y		Y								Y
	PCB 6 2,3'-Dichlorobiphenyl	EPA 1668	MLA-010				Y		Y	Y	Y	Y	Y	Y							Y					Y	Y	Y	Y	Y		Y	Y
		EPA 8270	MLA-007									Y																					
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y					Y		Y		Y								Y
	PCB 60 2,3,4,4'-Tetrachlorobiphenyl	EPA 1668	MLA-010				Y		Y	Y	Y	Y	Y	Y							Y					Y	Y	Y	Y	Y		Y	Y
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y					Y		Y		Y								Y
	PCB 61 2,3,4,5-Tetrachlorobiphenyl	EPA 1668	MLA-010				Y		Y	Y	Y	Y	Y	Y							Y					Y	Y	Y	Y	Y		Y	Y
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y					Y		Y		Y								Y
	PCB 62 2,3,4,6-Tetrachlorobiphenyl	EPA 1668	MLA-010				Y		Y	Y	Y	Y	Y	Y							Y					Y	Y	Y	Y	Y		Y	Y
		SGS AXYS MLA-010	MLA-010	Y	Y	Y								Y		Y					Y		Y		Y								Y
	PCB 62/65	EPA 8270	MLA-007									Y																					
		EPA 1668	MLA-010				Y		Y	Y	Y	Y	Y	Y							Y					Y	Y	Y	Y	Y		Y	Y
	PCB 63 2,3,4',5-Tetrachlorobiphenyl	EPA 8270	MLA-007									Y																					

Accreditation Scope																Serum	Solids													Tissue													Urine	Water	Water, Non-Potable																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																									
																CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB ISO 17025	ANAB DoD **	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB ISO 17025	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE *	Maine DOH	Pennsylvania DEP	ANAB ISO 17025	ANAB DoD **																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																								
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																			

Accreditation Scope				Serum										Tissue					Urine		Water		Water, Non-Potable													
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	CALA		SOLIDS										CALA					CALA		CALA		CALA											
						CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB ISO 17025	ANAB DoD **	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB ISO 17025	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE *	Maine DOH	Pennsylvania DEP	ANAB ISO 17025	ANAB DoD **
		SGS AXYS MLA-010	MLA-010	Y	Y			Y					Y	Y	Y	Y		Y				Y		Y		Y								Y		
	PCB 95 2,2',3,5',6-Pentachlorobiphenyl	EPA 1668	MLA-010					Y			Y	Y	Y	Y	Y	Y						Y				Y			Y	Y	Y	Y		Y	Y	
		SGS AXYS MLA-010	MLA-010	Y	Y		Y									Y		Y				Y		Y		Y								Y		
	PCB 95/93	EPA 8270	MLA-007											Y																						
		SGS AXYS MLA-007	MLA-007		Y													Y						Y												
	PCB 96 2,2',3,6,6'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y			Y	Y	Y	Y	Y	Y						Y				Y			Y	Y	Y	Y		Y	Y	
		EPA 8270	MLA-007											Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y		Y									Y		Y				Y		Y		Y								Y		
	PCB 97 2,2',3,4',5'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y			Y	Y	Y	Y	Y	Y						Y				Y			Y	Y	Y	Y		Y	Y	
		SGS AXYS MLA-010	MLA-010	Y	Y		Y									Y		Y				Y		Y		Y								Y		
	PCB 97/86	EPA 8270	MLA-007											Y																						
		SGS AXYS MLA-007	MLA-007		Y													Y						Y												
	PCB 98 2,2',3,4',6'-Pentachlorobiphenyl	EPA 1668	MLA-010					Y			Y	Y	Y	Y	Y	Y						Y				Y			Y	Y	Y	Y		Y	Y	
		SGS AXYS MLA-010	MLA-010	Y	Y		Y									Y		Y				Y		Y		Y								Y		
	PCB 98/102	EPA 8270	MLA-007											Y																						
	PCB 99 2,2',4,4',5-Pentachlorobiphenyl	EPA 1668	MLA-010					Y			Y	Y	Y	Y	Y	Y						Y				Y			Y	Y	Y	Y		Y	Y	
		EPA 8270	MLA-007											Y																						
		SGS AXYS MLA-010	MLA-010	Y	Y		Y									Y		Y				Y		Y		Y								Y		
		SGS AXYS MLA-007	MLA-007		Y													Y						Y												
		SGS AXYS MLA-901	MLA-901	Y																																
	PCB congeners, total	EPA 1668	MLA-010									Y																				Y				
	Sum - Dichlorobiphenyls (BZ-12+ BZ-13)	EPA 1668	MLA-010													Y						Y												Y		
		SGS AXYS MLA-010	MLA-010													Y						Y												Y		
	Sum - Heptachlorobiphenyls (BZ-171 + BZ-173)	EPA 1668	MLA-010													Y						Y												Y		
		SGS AXYS MLA-010	MLA-010													Y						Y												Y		
	Sum - Heptachlorobiphenyls (BZ-180 + BZ-193)	EPA 1668	MLA-010													Y						Y												Y		
		SGS AXYS MLA-010	MLA-010													Y						Y												Y		
	Sum - Heptachlorobiphenyls (BZ-183 + BZ-185)	EPA 1668	MLA-010													Y						Y												Y		
		SGS AXYS MLA-010	MLA-010													Y						Y												Y		
	Sum - Hexachlorobiphenyls (BZ-128 + BZ-166)	EPA 1668	MLA-010													Y						Y												Y		
		SGS AXYS MLA-010	MLA-010													Y						Y												Y		
	Sum - Hexachlorobiphenyls (BZ-129 + BZ-138 + BZ-160 + BZ-163)	EPA 1668	MLA-010													Y						Y												Y		
		SGS AXYS MLA-010	MLA-010													Y						Y												Y		
	Sum - Hexachlorobiphenyls (BZ-134 + BZ-143)	EPA 1668	MLA-010													Y						Y												Y		
		SGS AXYS MLA-010	MLA-010													Y						Y												Y		
	Sum - Hexachlorobiphenyls (BZ-135 + BZ-151 + BZ-154)	EPA 1668	MLA-010													Y						Y												Y		
		SGS AXYS MLA-010	MLA-010													Y						Y												Y		
	Sum - Hexachlorobiphenyls (BZ-139 + BZ-140)	EPA 1668	MLA-010													Y						Y												Y		
		SGS AXYS MLA-010	MLA-010													Y						Y												Y		
	Sum - Hexachlorobiphenyls (BZ-147 + BZ-149)	EPA 1668	MLA-010													Y						Y												Y		
		SGS AXYS MLA-010	MLA-010													Y						Y												Y		
	Sum - Hexachlorobiphenyls (BZ-153 + BZ-168)	EPA 1668	MLA-010													Y						Y												Y		
		SGS AXYS MLA-010	MLA-010													Y						Y												Y		
	Sum - Hexachlorobiphenyls (BZ-156 + BZ-157)	EPA 1668	MLA-010													Y						Y												Y		
		SGS AXYS MLA-010	MLA-010													Y						Y												Y		
	Sum - Pentachlorobiphenyls (BZ-107 + BZ-124)	EPA 1668	MLA-010													Y						Y												Y		
		SGS AXYS MLA-010	MLA-010													Y						Y												Y		
	Sum - Pentachlorobiphenyls (BZ-108 + BZ-124)	EPA 1668	MLA-010													Y						Y												Y		
		SGS AXYS MLA-010	MLA-010													Y						Y												Y		
	Sum - Pentachlorobiphenyls (BZ-110 + BZ-115)	EPA 1668	MLA-010													Y						Y												Y		

Accreditation Scope				SGS AXYS Analytical Services Ltd. file ref.: ACC-101 Rev. 40																			
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Serum										Tissue									
				CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB ISO 17025	ANAB DoD **	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB ISO 17025	CALA	CALA
		SGS AXYS MLA-010	MLA-010											Y							Y		
	Sum - Pentachlorobiphenyls (BZ-83 + BZ-99)	EPA 1668	MLA-010											Y							Y		
		SGS AXYS MLA-010	MLA-010											Y							Y		
	Sum - Pentachlorobiphenyls (BZ-85 + BZ-116 + BZ-117)	EPA 1668	MLA-010											Y							Y		
		SGS AXYS MLA-010	MLA-010											Y							Y		
	Sum - Pentachlorobiphenyls (BZ-86 + BZ-87 + BZ 97 + BZ-109 + BZ-119 + BZ-125)	EPA 1668	MLA-010											Y							Y		
		SGS AXYS MLA-010	MLA-010											Y							Y		
	Sum - Pentachlorobiphenyls (BZ-86 + BZ-87 + BZ-97 + BZ-108 + BZ-119 +BZ-125)	EPA 1668	MLA-010											Y							Y		
		SGS AXYS MLA-010	MLA-010											Y							Y		
	Sum - Pentachlorobiphenyls (BZ-88 + BZ-91)	EPA 1668	MLA-010											Y							Y		
		SGS AXYS MLA-010	MLA-010											Y							Y		
	Sum - Pentachlorobiphenyls (BZ-90 + BZ-101 + BZ-113)	EPA 1668	MLA-010											Y							Y		
		SGS AXYS MLA-010	MLA-010											Y							Y		
	Sum - Pentachlorobiphenyls (BZ-93 + BZ-95 + BZ-98 + BZ-100 + BZ-102)	EPA 1668	MLA-010											Y							Y		
		SGS AXYS MLA-010	MLA-010											Y							Y		
	Sum - Tetrachlorobiphenyls (BZ-40 + BZ-41 + BZ-71)	EPA 1668	MLA-010											Y							Y		
		SGS AXYS MLA-010	MLA-010											Y							Y		
	Sum - Tetrachlorobiphenyls (BZ-44 + BZ-47 + BZ-65)	EPA 1668	MLA-010											Y							Y		
		SGS AXYS MLA-010	MLA-010											Y							Y		
	Sum - Tetrachlorobiphenyls (BZ-45 + BZ-51)	EPA 1668	MLA-010											Y							Y		
		SGS AXYS MLA-010	MLA-010											Y							Y		
	Sum - Tetrachlorobiphenyls (BZ-49 + BZ-69)	EPA 1668	MLA-010											Y							Y		
		SGS AXYS MLA-010	MLA-010											Y							Y		
	Sum - Tetrachlorobiphenyls (BZ-50 + BZ-53)	EPA 1668	MLA-010											Y							Y		
		SGS AXYS MLA-010	MLA-010											Y							Y		
	Sum - Tetrachlorobiphenyls (BZ-59 + BZ-62 + BZ-75)	EPA 1668	MLA-010											Y							Y		
		SGS AXYS MLA-010	MLA-010											Y							Y		
	Sum - Tetrachlorobiphenyls (BZ-61 + BZ-70 + BZ-74 + BZ-76)	EPA 1668	MLA-010											Y							Y		
		SGS AXYS MLA-010	MLA-010											Y							Y		
	Sum - Trichlorobiphenyls (BZ-18 + BZ-30)	EPA 1668	MLA-010											Y							Y		
		SGS AXYS MLA-010	MLA-010											Y							Y		
	Sum - Trichlorobiphenyls (BZ-20 + BZ-28)	EPA 1668	MLA-010											Y							Y		
		SGS AXYS MLA-010	MLA-010											Y							Y		
	Sum - Trichlorobiphenyls (BZ-21 + BZ-33)	EPA 1668	MLA-010											Y							Y		
		SGS AXYS MLA-010	MLA-010											Y							Y		
	Sum - Trichlorobiphenyls (BZ-26 + BZ-29)	EPA 1668	MLA-010											Y							Y		
		SGS AXYS MLA-010	MLA-010											Y							Y		
	Total Dichlorobiphenyls	EPA 1668	MLA-010											Y							Y		
		EPA 8270	MLA-007									Y											
		SGS AXYS MLA-010	MLA-010	Y	Y									Y		Y		Y		Y			Y
		SGS AXYS MLA-007	MLA-007		Y											Y				Y			
	Total Heptachlorobiphenyls	EPA 1668	MLA-010											Y							Y		
		EPA 8270	MLA-007									Y											
		SGS AXYS MLA-010	MLA-010	Y	Y									Y		Y		Y		Y			Y
		SGS AXYS MLA-007	MLA-007		Y											Y				Y			
	Total Hexachlorobiphenyls	EPA 1668	MLA-010											Y							Y		
		EPA 8270	MLA-007									Y											
		SGS AXYS MLA-010	MLA-010	Y	Y									Y		Y		Y		Y			Y
		SGS AXYS MLA-007	MLA-007		Y											Y				Y			
	Total Monochlorobiphenyls	EPA 1668	MLA-010											Y							Y		

Accreditation Scope

SGS AXYS Analytical Services Ltd.
file ref.: ACC-101 Rev. 40

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Serum										Tissue					Urine		Water		Water, Non-Potable												
				CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB ISO 17025	ANAB DoD **	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB ISO 17025	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE *	Maine DOH	Pennsylvania DEP	ANAB ISO 17025	ANAB DoD **	
PCDDF	Total Nonachlorobiphenyls	SGS AXYS MLA-010	MLA-010	Y	Y									Y		Y			Y		Y		Y									Y			
		EPA 1668	MLA-010											Y																			Y		
	Total Octachlorobiphenyls	EPA 8270	MLA-007									Y																							
		SGS AXYS MLA-010	MLA-010	Y	Y									Y		Y			Y		Y		Y										Y		
		SGS AXYS MLA-007	MLA-007		Y											Y							Y												
		EPA 1668	MLA-010											Y							Y												Y		
	Total PCBs	EPA 8270	MLA-007									Y																							
		SGS AXYS MLA-010	MLA-010	Y	Y									Y		Y			Y		Y		Y										Y		
		SGS AXYS MLA-007	MLA-007		Y											Y							Y												
		EPA 1668	MLA-010											Y							Y													Y	
	Total Pentachlorobiphenyls	SGS AXYS MLA-010	MLA-010											Y																					
		EPA 8270	MLA-007									Y																							
		SGS AXYS MLA-010	MLA-010	Y	Y									Y		Y			Y		Y		Y										Y		
		SGS AXYS MLA-007	MLA-007		Y											Y							Y												
	Total Polychlorinated biphenyls	SGS AXYS MLA-007	MLA-007		Y											Y							Y												
	Total Tetrachlorobiphenyls	EPA 1668	MLA-010											Y							Y													Y	
		EPA 8270	MLA-007									Y																							
		SGS AXYS MLA-010	MLA-010	Y	Y									Y		Y			Y		Y		Y										Y		
		SGS AXYS MLA-007	MLA-007		Y											Y							Y												
	Total Trichlorobiphenyls	EPA 1668	MLA-010											Y							Y													Y	
		EPA 8270	MLA-007									Y																							
		SGS AXYS MLA-010	MLA-010	Y	Y									Y		Y			Y		Y		Y										Y		
		SGS AXYS MLA-007	MLA-007		Y											Y							Y												
PCDDF	"Dioxins and Dibenzofurans" category (CA only)	EPA 1613	MLA-017																					Y											
		EPA 8290	MLA-017			Y																													
	1,2,3,4,6,7,8-HpCDD	EPA 1613	MLA-017									Y		Y							Y				Y		Y	Y	Y	Y	Y	Y	Y		
		EPA 8290	MLA-017				Y		Y	Y	Y		Y	Y			Y		Y	Y	Y				Y			Y	Y					Y	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y								Y		Y	Y		Y		Y		Y		Y									Y	
	1,2,3,4,6,7,8-HpCDF	EPA 1613	MLA-017									Y		Y							Y				Y		Y	Y	Y	Y	Y	Y	Y		
		EPA 8290	MLA-017				Y		Y	Y	Y		Y	Y			Y		Y	Y	Y				Y			Y	Y					Y	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y								Y		Y	Y		Y		Y		Y		Y									Y	
	1,2,3,4,7,8,9-HpCDF	EPA 1613	MLA-017									Y		Y							Y				Y		Y	Y	Y	Y	Y	Y	Y		
		EPA 8290	MLA-017				Y		Y	Y	Y		Y	Y			Y		Y	Y	Y				Y			Y	Y					Y	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y								Y		Y	Y		Y		Y		Y		Y									Y	
	1,2,3,4,7,8-HxCDD	EPA 1613	MLA-017									Y		Y							Y				Y		Y	Y	Y	Y	Y	Y	Y		
		EPA 8290	MLA-017				Y		Y	Y	Y		Y	Y			Y		Y	Y	Y				Y			Y	Y					Y	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y								Y		Y	Y		Y		Y		Y		Y									Y	
	1,2,3,6,7,8-HxCDD	EPA 1613	MLA-017									Y		Y							Y				Y		Y	Y	Y	Y	Y	Y	Y		
		EPA 8290	MLA-017				Y		Y	Y	Y		Y	Y			Y		Y	Y	Y				Y			Y	Y					Y	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y								Y		Y	Y		Y		Y		Y		Y									Y	
	1,2,3,6,7,8-HxCDF	EPA 1613	MLA-017									Y		Y							Y				Y		Y	Y	Y	Y	Y	Y	Y		
		EPA 8290	MLA-017				Y		Y	Y	Y		Y	Y			Y		Y	Y	Y				Y			Y	Y					Y	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y								Y		Y	Y		Y		Y		Y		Y									Y	
PCDDF	1,2,3,7,8,9-HxCDD	EPA 1613	MLA-017									Y		Y							Y				Y		Y	Y	Y	Y	Y	Y	Y		
		EPA 8290	MLA-017				Y		Y	Y	Y		Y	Y			Y		Y	Y	Y				Y			Y	Y					Y	
		SGS AXYS MLA-017	MLA-017	Y	Y	Y								Y		Y	Y		Y		Y		Y		Y									Y	

Accreditation Scope

SGS AXYS Analytical Services Ltd.
file ref.: ACC-101 Rev. 40

Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	Serum										Tissue					Urine		Water		Water, Non-Potable											
				CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB ISO 17025	ANAB DoD **	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB ISO 17025	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE *	Maine DOH	Pennsylvania DEP	ANAB ISO 17025	ANAB DoD **
1,2,3,7,8,9-HxCDF	EPA 1613	MLA-017										Y		Y							Y												Y	
	EPA 8290	MLA-017					Y		Y	Y	Y		Y	Y			Y		Y	Y	Y				Y			Y	Y				Y	
	SGS AXYS MLA-017	MLA-017		Y	Y	Y								Y		Y	Y			Y		Y		Y									Y	
1,2,3,7,8-PeCDD	EPA 1613	MLA-017										Y		Y							Y												Y	
	EPA 8290	MLA-017					Y		Y	Y	Y		Y	Y			Y		Y	Y	Y				Y			Y	Y				Y	
	SGS AXYS MLA-017	MLA-017		Y	Y	Y								Y		Y	Y			Y		Y		Y									Y	
1,2,3,7,8-PeCDF	EPA 1613	MLA-017										Y		Y							Y						Y	Y	Y	Y	Y	Y		Y
	EPA 8290	MLA-017					Y		Y	Y	Y		Y	Y			Y		Y	Y	Y				Y			Y	Y				Y	
	SGS AXYS MLA-017	MLA-017		Y	Y	Y								Y		Y	Y			Y		Y		Y									Y	
2,3,4,6,7,8-HxCDF	EPA 1613	MLA-017										Y		Y							Y					Y		Y	Y	Y	Y	Y		Y
	EPA 8290	MLA-017					Y		Y	Y	Y		Y	Y			Y		Y	Y	Y				Y			Y	Y				Y	
	SGS AXYS MLA-017	MLA-017		Y	Y	Y								Y		Y	Y			Y		Y		Y									Y	
2,3,4,7,8-PeCDF	EPA 1613	MLA-017										Y		Y							Y					Y		Y	Y	Y	Y	Y		Y
	EPA 8290	MLA-017										Y		Y			Y		Y	Y	Y				Y			Y	Y				Y	
	SGS AXYS MLA-017	MLA-017		Y	Y	Y								Y		Y	Y			Y		Y		Y									Y	
2,3,7,8-TCDD	EPA 1613	MLA-017										Y		Y							Y					Y		Y	Y	Y	Y	Y		Y
	EPA 8290	MLA-017					Y		Y	Y	Y		Y	Y			Y		Y	Y	Y				Y			Y	Y				Y	
	SGS AXYS MLA-017	MLA-017		Y	Y	Y								Y		Y	Y			Y		Y		Y									Y	
2,3,7,8-TCDF	EPA 1613	MLA-017										Y		Y							Y					Y		Y	Y	Y	Y	Y		Y
	EPA 8290	MLA-017					Y		Y	Y	Y		Y	Y			Y		Y	Y	Y				Y			Y	Y				Y	
	SGS AXYS MLA-017	MLA-017		Y	Y	Y								Y		Y	Y			Y		Y		Y									Y	
OCDD	EPA 1613	MLA-017										Y		Y							Y					Y		Y	Y	Y	Y	Y		Y
	EPA 8290	MLA-017					Y		Y	Y	Y		Y	Y			Y		Y	Y	Y				Y			Y	Y				Y	
	SGS AXYS MLA-017	MLA-017		Y	Y	Y								Y		Y	Y			Y		Y		Y									Y	
OCDF	EPA 1613	MLA-017										Y		Y							Y					Y		Y	Y	Y	Y	Y		Y
	EPA 8290	MLA-017					Y		Y	Y	Y		Y	Y			Y		Y	Y	Y				Y			Y	Y				Y	
	SGS AXYS MLA-017	MLA-017		Y	Y	Y								Y		Y	Y			Y		Y		Y									Y	
Total HpCDD	EPA 1613	MLA-017												Y							Y					Y						Y	Y	
	EPA 8290	MLA-017					Y		Y		Y		Y	Y			Y		Y	Y	Y				Y				Y			Y	Y	
	SGS AXYS MLA-017	MLA-017					Y							Y			Y			Y				Y								Y	Y	
Total HpCDF	EPA 1613	MLA-017												Y							Y				Y							Y	Y	
	EPA 8290	MLA-017					Y		Y		Y		Y	Y			Y		Y	Y	Y				Y				Y			Y	Y	
	SGS AXYS MLA-017	MLA-017					Y							Y			Y			Y				Y								Y	Y	
Total HxCDD	EPA 1613	MLA-017												Y							Y				Y							Y	Y	
	EPA 8290	MLA-017					Y		Y		Y		Y	Y			Y		Y	Y	Y				Y				Y			Y	Y	
	SGS AXYS MLA-017	MLA-017					Y							Y			Y			Y				Y								Y	Y	
Total HxCDF	EPA 1613	MLA-017												Y							Y				Y							Y	Y	
	EPA 8290	MLA-017					Y		Y		Y		Y	Y			Y		Y	Y	Y				Y				Y			Y	Y	
	SGS AXYS MLA-017	MLA-017					Y							Y			Y			Y				Y								Y	Y	
Total PCDD	EPA 1613	MLA-017												Y							Y				Y							Y	Y	
	EPA 8290	MLA-017												Y							Y				Y							Y	Y	
	SGS AXYS MLA-017	MLA-017												Y							Y				Y							Y	Y	
Total PCDD+PCDF	EPA 1613	MLA-017												Y							Y				Y							Y	Y	
	EPA 8290	MLA-017												Y							Y				Y							Y	Y	
	SGS AXYS MLA-017	MLA-017												Y							Y				Y							Y	Y	
Total PCDF	EPA 1613	MLA-017												Y							Y				Y							Y	Y	
	EPA 8290	MLA-017												Y							Y				Y							Y	Y	
	SGS AXYS MLA-017	MLA-017												Y							Y				Y							Y	Y	
Total PeCDD	EPA 1613	MLA-017												Y							Y				Y							Y	Y	
	EPA 8290	MLA-017					Y		Y		Y		Y	Y			Y		Y	Y	Y				Y				Y			Y	Y	

Accreditation Scope				Serum																Tissue						Urine	Water	Water, Non-Potable						
				Solids																														
				CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	ANAB ISO 17025	ANAB DoD **	CALA	Florida DOH	Minnesota DOH	New Jersey DEP	Virginia DGS	ANAB ISO 17025	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE *	Maine DOH	Pennsylvania DEP	ANAB ISO 17025	ANAB DoD **
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID																															
	2-hydroxy-ibuprofen	SGS AXYS MLA-075	MLA-075		Y																		Y											
	4-Epianhydrochlortetracycline (EACTC)	EPA 1694	MLA-075											Y																	Y			
		SGS AXYS MLA-075	MLA-075		Y																		Y											
	4-Epianhydrotetracycline (EATC)	EPA 1694	MLA-075											Y																	Y			
		SGS AXYS MLA-075	MLA-075		Y																		Y											
	4-Epichlortetracycline (ECTC)	EPA 1694	MLA-075											Y																	Y			
		SGS AXYS MLA-075	MLA-075		Y																		Y											
	4-Epioxytetracycline (EOTC)	EPA 1694	MLA-075											Y																	Y			
		SGS AXYS MLA-075	MLA-075		Y																		Y											
	4-Epitetracycline (ETC)	EPA 1694	MLA-075											Y																	Y			
		SGS AXYS MLA-075	MLA-075		Y																		Y											
	Acetaminophen	EPA 1694	MLA-075											Y																	Y			
		SGS AXYS MLA-075	MLA-075		Y																		Y											
	Albuterol	EPA 1694	MLA-075											Y																	Y			
		SGS AXYS MLA-075	MLA-075		Y																		Y											
	Alprazolam	SGS AXYS MLA-075	MLA-075		Y																		Y											
	Amitriptyline	SGS AXYS MLA-075	MLA-075		Y																		Y											
	Amlodipine	SGS AXYS MLA-075	MLA-075		Y																		Y											
	Amphetamine	SGS AXYS MLA-075	MLA-075		Y																		Y											
	Anhydrochlortetracycline (ACTC)	EPA 1694	MLA-075											Y																	Y			
		SGS AXYS MLA-075	MLA-075		Y																		Y											
	Anhydrotetracycline (ATC)	EPA 1694	MLA-075											Y																	Y			
		SGS AXYS MLA-075	MLA-075		Y																		Y											
	Atenolol	SGS AXYS MLA-075	MLA-075		Y																		Y											
	Atorvastatin	SGS AXYS MLA-075	MLA-075		Y																		Y											
	Azithromycin	EPA 1694	MLA-075											Y																	Y			
		SGS AXYS MLA-075	MLA-075		Y																		Y											
	Benzoyllecgonine	SGS AXYS MLA-075	MLA-075		Y																		Y											
	Benztropine	SGS AXYS MLA-075	MLA-075		Y																		Y											
	Betamethasone	SGS AXYS MLA-075	MLA-075		Y																		Y											
	Bisphenol A	EPA 1694	MLA-075											Y																	Y			
		SGS AXYS MLA-075	MLA-075		Y																		Y											
	Caffeine	EPA 1694	MLA-075											Y																	Y			
		SGS AXYS MLA-075	MLA-075		Y																		Y											
	Carbadox	EPA 1694	MLA-075											Y																	Y			
		SGS AXYS MLA-075	MLA-075		Y																		Y											
	Carbamazepine	EPA 1694	MLA-075											Y																	Y			
		SGS AXYS MLA-075	MLA-075		Y																		Y											
	Cefotaxime	EPA 1694	MLA-075											Y																	Y			
		SGS AXYS MLA-075	MLA-075		Y																		Y											
	Chlortetracycline (CTC)	EPA 1694	MLA-075											Y																	Y			
		SGS AXYS MLA-075	MLA-075		Y																		Y											
	Cimetidine	EPA 1694	MLA-075											Y																	Y			
		SGS AXYS MLA-075	MLA-075		Y																		Y											
	Ciprofloxacin	EPA 1694	MLA-075											Y																	Y			
		SGS AXYS MLA-075	MLA-075		Y																		Y											
	Clarithromycin	EPA 1694	MLA-075											Y																	Y			
		SGS AXYS MLA-075	MLA-075		Y																		Y											
	Cinafloxacin	EPA 1694	MLA-075											Y																	Y			
		SGS AXYS MLA-075	MLA-075		Y																		Y											

Accreditation ScopeSGS AXYS Analytical Services Ltd.
file ref.: ACC-101 Rev. 40

Accreditation Scope																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																										
---------------------	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--

Accreditation Scope																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																							</
---------------------	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	----

Accreditation Scope				Serum		Solids		Tissue		Urine	Water	Water, Non-Potable				
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE *	Maine DOH	Pennsylvania DEP	ANAB ISO 17025	ANAB DoD **
	Decanoylcarnitine	SGS AXYS MLM-001	MLM-001	Y												
	Decenoylcarnitine	SGS AXYS MLM-001	MLM-001	Y												
	deoxycholic acid	SGS AXYS MLM-001	MLM-001	Y												
	docosahexaenoic acid (DHA)	SGS AXYS MLM-001	MLM-001													
	docosatetraenoic acid (adrenic acid)	SGS AXYS MLM-001	MLM-001													
	Dodecanedioylcarnitine	SGS AXYS MLM-001	MLM-001	Y												
	Dodecanoylcarnitine	SGS AXYS MLM-001	MLM-001	Y												
	Dodecenoylcarnitine	SGS AXYS MLM-001	MLM-001	Y												
	Dopamine	SGS AXYS MLM-001	MLM-001	Y												
	eicosapentaenoic acid (EPA)	SGS AXYS MLM-001	MLM-001													
	Eicosatetraenoic acid (arachidonic acid)	SGS AXYS MLM-001	MLM-001													
	eicosatrienoic acid (dihomo-γ-linolenic acid)	SGS AXYS MLM-001	MLM-001													
	Glutaconylcarnitine	SGS AXYS MLM-001	MLM-001	Y												
	Glutamate	SGS AXYS MLM-001	MLM-001	Y												
	Glutamine	SGS AXYS MLM-001	MLM-001	Y												
	Glutaryl carnitine (Hydroxyhexanoylcarnitine)	SGS AXYS MLM-001	MLM-001	Y												
	Glycine	SGS AXYS MLM-001	MLM-001	Y												
	glycochenodeoxycholic acid	SGS AXYS MLM-001	MLM-001	Y												
	glycocholic acid	SGS AXYS MLM-001	MLM-001	Y												
	glycodeoxycholic acid	SGS AXYS MLM-001	MLM-001	Y												
	Hexadecadienylcarnitine	SGS AXYS MLM-001	MLM-001	Y												
	hexadecanoic acid (palmitic acid)	SGS AXYS MLM-001	MLM-001													
	Hexadecanoylcarnitine	SGS AXYS MLM-001	MLM-001	Y												
	hexadecenoic acid (palmitoleic acid)	SGS AXYS MLM-001	MLM-001													
	Hexadecenoylcarnitine	SGS AXYS MLM-001	MLM-001	Y												
	Hexanoylcarnitine (Fumaryl carnitine)	SGS AXYS MLM-001	MLM-001	Y												
	Hexenoylcarnitine	SGS AXYS MLM-001	MLM-001	Y												
	Hexose (sum isomers)	SGS AXYS MLM-001	MLM-001	Y												
	Histamine	SGS AXYS MLM-001	MLM-001	Y												
	Histidine	SGS AXYS MLM-001	MLM-001	Y												
	Hydroxyhexadecadienylcarnitine	SGS AXYS MLM-001	MLM-001	Y												
	Hydroxyhexadecanoylcarnitine	SGS AXYS MLM-001	MLM-001	Y												
	Hydroxyhexadecenoylcarnitine	SGS AXYS MLM-001	MLM-001	Y												
	Hydroxybutyrylcarnitine	SGS AXYS MLM-001	MLM-001	Y												
	Hydroxyoctadecenoylcarnitine	SGS AXYS MLM-001	MLM-001	Y												
	Hydroxyproline	SGS AXYS MLM-001	MLM-001	Y												
	Hydroxypropionylcarnitine	SGS AXYS MLM-001	MLM-001	Y												
	Hydroxysphingomyeline C14:1	SGS AXYS MLM-001	MLM-001	Y												
	Hydroxysphingomyeline C16:1	SGS AXYS MLM-001	MLM-001	Y												
	Hydroxysphingomyeline C22:1	SGS AXYS MLM-001	MLM-001	Y												
	Hydroxysphingomyeline C22:2	SGS AXYS MLM-001	MLM-001	Y												
	Hydroxysphingomyeline C24:1	SGS AXYS MLM-001	MLM-001	Y												
	Hydroxytetradecadienylcarnitine	SGS AXYS MLM-001	MLM-001	Y												
	Hydroxytetradecenoylcarnitine	SGS AXYS MLM-001	MLM-001	Y												
	Hydroxyvaleryl carnitine (Methylmalonylcarnitine)	SGS AXYS MLM-001	MLM-001	Y												
	Isoleucine	SGS AXYS MLM-001	MLM-001	Y												
	Kynurenine	SGS AXYS MLM-001	MLM-001	Y												
	Leucine	SGS AXYS MLM-001	MLM-001	Y												
	lithocholic acid	SGS AXYS MLM-001	MLM-001	Y												
	Lysine	SGS AXYS MLM-001	MLM-001	Y												
	lysoPhosphatidylcholine acyl C14:0	SGS AXYS MLM-001	MLM-001	Y												
	lysoPhosphatidylcholine acyl C16:0	SGS AXYS MLM-001	MLM-001	Y												
	lysoPhosphatidylcholine acyl C16:1	SGS AXYS MLM-001	MLM-001	Y												
	lysoPhosphatidylcholine acyl C17:0	SGS AXYS MLM-001	MLM-001	Y												
	lysoPhosphatidylcholine acyl C18:0	SGS AXYS MLM-001	MLM-001	Y												
	lysoPhosphatidylcholine acyl C18:1	SGS AXYS MLM-001	MLM-001	Y												
	lysoPhosphatidylcholine acyl C18:2	SGS AXYS MLM-001	MLM-001	Y												
	lysoPhosphatidylcholine acyl C20:3	SGS AXYS MLM-001	MLM-001	Y												

Accreditation Scope				Serum		Solids		Tissue		Urine		Water		Water, Non-Potable	
SGS AXYS Analytical Services Ltd. file ref.: ACC-101 Rev. 40				CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	Pennsylvania DEP	ANAB ISO 17025
Compound Class	Compound	Accredited Method ID	SGS AXYS Method ID	CALA	CALA	California DPH	Florida DOH	Minnesota DOH	New Jersey DEP	New York DOH	Virginia DGS	Washington DE	Maine DOH	Pennsylvania DEP	ANAB ISO 17025
	Phosphatidylcholine diacyl C26:0	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C28:1	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C30:0	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C30:2	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C32:0	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C32:1	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C32:2	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C32:3	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C34:1	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C34:2	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C34:3	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C34:4	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C36:0	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C36:1	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C36:2	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C36:3	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C36:4	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C36:5	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C36:6	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C38:0	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C38:1	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C38:3	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C38:4	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C38:5	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C38:6	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C40:1	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C40:2	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C40:3	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C40:4	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C40:5	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C40:6	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C42:0	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C42:1	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C42:2	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C42:4	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C42:5	SGS AXYS MLM-001	MLM-001	Y											
	Phosphatidylcholine diacyl C42:6	SGS AXYS MLM-001	MLM-001	Y											
	Pimelylcarnitine	SGS AXYS MLM-001	MLM-001	Y											
	Proline	SGS AXYS MLM-001	MLM-001	Y											
	Propenoylcarnitine	SGS AXYS MLM-001	MLM-001	Y											
	Propionylcarnitine	SGS AXYS MLM-001	MLM-001	Y											
	Putrescine	SGS AXYS MLM-001	MLM-001	Y											
	Sarcosine	SGS AXYS MLM-001	MLM-001	Y											
	Serine	SGS AXYS MLM-001	MLM-001	Y											
	Serotonin	SGS AXYS MLM-001	MLM-001	Y											
	Spermidine	SGS AXYS MLM-001	MLM-001	Y											
	Spermine	SGS AXYS MLM-001	MLM-001	Y											
	Sphingomyeline C16:0	SGS AXYS MLM-001	MLM-001	Y											
	Sphingomyeline C16:1	SGS AXYS MLM-001	MLM-001	Y											
	Sphingomyeline C18:0	SGS AXYS MLM-001	MLM-001	Y											
	Sphingomyeline C18:1	SGS AXYS MLM-001	MLM-001	Y											
	Sphingomyeline C20:2	SGS AXYS MLM-001	MLM-001	Y											
	Sphingomyeline C22:3	SGS AXYS MLM-001	MLM-001	Y											
	Sphingomyeline C24:0	SGS AXYS MLM-001	MLM-001	Y											
	Sphingomyeline C24:1	SGS AXYS MLM-001	MLM-001	Y											
	Sphingomyeline C26:0	SGS AXYS MLM-001	MLM-001	Y											
	Sphingomyeline C26:1	SGS AXYS MLM-001	MLM-001	Y											
	Symmetric dimethylarginine	SGS AXYS MLM-001	MLM-001	Y											

Accreditation Scope																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																				
---------------------	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--

Legend

Y	Accreditation scope
BFR	Brominated flame retardants (non-PBDPE)
BPA and mPE	Bisphenol A and mono-Phthalate Esters
HBCDD	Hexabromocyclododecane
OC Pesticides	Organochlorine Pesticides
PAH	Polycyclic Aromatic Hydrocarbons
PBDPE	Polybrominated diphenylethers
PCB	Polychlorinated Biphenyls
PCDDF	Polychlorinated dibenzodioxins/furans
PFAS	Per- and Polyfluoroalkyl Substances
PPCP	Pharmaceutical and Personal Care Products
TBBPA	Tetrabromobisphenol A
TOP	Total Oxidizable Precursors
California DPH	California Department of Public Health, Lab ID 2911
Florida DOH	Florida Department of Health, Lab ID E871007, (NELAC Standard)
Pennsylvania DEP	Pennsylvania Department of Environmental Protection
Minnesota DOH	Minnesota Department of Health, Lab ID 232-999-430, (NELAC Standard)
New Jersey DEP	New Jersey Department of Environmental Protection, Lab ID CANA005, (NELAC Standard)
New York DOH	New York Department of Health, Lab ID 11674, (NELAC Standard)
Washington DE	Washington Department of Ecology, Lab ID C404
Virginia DGS	Virginia Department of General Services, Division of Consolidated Laboratory Services, Lab ID 460224, (NELAC Standard)
Maine DOH	Maine Center for Disease Control and Prevention, Department of Health and Human Services, Lab ID CN00003

ANAB DoD ANSI-ASQ National Accreditation Board, certificate ADE-1861, (US DoD QSM 5.1 Standard)



CALA Canadian Association for Laboratory Accreditation Inc., Lab ID A2637, (ISO/IEC 17025:2005 Standard)



CALA
Testing
Accreditation No. A 2637



ANAB ISO 17025 ANSI-ASQ National Accreditation Board, certificate ADE-1861.01, (ISO/IEC 17025:2005 Standard)

