

Data Validation Report

Project:	Portland Harbor Pre-Remedial Design Investigation and Baseline Sampling	
Laboratory:	SGS-AXYS, Sydney, British Columbia, Canada	
Laboratory Group:	WG66481-PAH	
Analyses/Method:	Polycyclic Aromatic Hydrocarbons (PAHs) / AXYS Method MLA-021 Rev12 Ver. 05	
Validation Level:	Stage 2A	
AECOM Project Number:	60566335.2.12	
Prepared by:	Paula DiMattei/AECOM	Completed on: 02/26/2019
Reviewed by:	Elissa McDonagh/AECOM	File Name: WG66481-PAH DVR

SUMMARY

The samples listed below were collected by AECOM in Portland Harbor in Portland, OR on November 27-30, 2018 and December 1, 2018.

Sample ID	Matrix/Sample Type
PDI-RB-XF-181129	Equipment Blank
PDI-WS-T01-1811	Surface Water
PDI-WS-T02-1811	Surface Water
PDI-WS-T03-1811	Surface Water
PDI-WS-T04-1812	Surface Water
PDI-WS-T05-1811	Surface Water
PDI-WS-T06-1811	Surface Water
PDI-WS-T07-1811	Surface Water

Data validation activities were conducted with reference to:

- *AXYS Laboratory SOP MLA-021 Rev.12 Ver. 05: Analytical Method for the Determination of Polycyclic Aromatic Hydrocarbons (PAHs), Alkylated PAHs and Alkanes,*
- *USEPA Contract Laboratory Program National Functional Guidelines for High Resolution Superfund Methods Data Review (April 2016),*
- *Quality Assurance Project Plan, Portland Harbor Pre-Remedial Design Investigation and Baseline Sampling, Portland Harbor Superfund Site (March 2018), and the*
- laboratory quality control (QC) limits.

The National Functional Guidelines were modified to accommodate the non-CLP methodologies. In the absence of method-specific information, laboratory QC limits, project-specific requirements and/or AECOM professional judgment were used as appropriate.

REVIEW ELEMENTS

The data were evaluated based on the following parameters (where applicable to the method):

- ✓ Data completeness (chain-of-custody (COC)/sample integrity)
- ✓ Holding times and sample preservation
- ✗ Laboratory blanks/equipment blanks
- NA Matrix spike (MS) and/or matrix spike duplicate (MSD) results
- ✓ Ongoing precision and recovery (OPR) results
- NA Field duplicate results
- ✗ Labeled compound recoveries
- ✗ Sample results/reporting issues

The symbol (✓) indicates that no validation qualifiers were applied based on this parameter. An NA indicates that the parameter was not included as part of this data set or was not applicable to this validation and therefore not reviewed. The symbol (✗) indicates that a QC nonconformance resulted in the qualification of data. Any QC nonconformance that resulted in the qualification of data is discussed below. In addition, nonconformances or other issues that were noted during validation, but did not result in qualification of data, may be discussed for informational purposes only.

The data appear valid as qualified and may be used for decision making purposes. Select data points were qualified as estimated due to nonconformances of certain QC criteria (see discussion below). Qualified sample results are presented in Table 1.

RESULTS

Data Completeness (COC)/Sample Integrity

The data package was reviewed and found to meet acceptance criteria for completeness:

- The COCs were reviewed for completeness of information relevant to the samples and requested analyses, and for signatures indicating transfer of sample custody.
- The laboratory sample login sheet(s) were reviewed for issues potentially affecting sample integrity, including the condition of sample containers upon receipt at the laboratory.
- Completeness of analyses was verified by comparing the reported results to the COC requests.

Holding Times and Sample Preservation

Sample preservation and preparation/analysis holding times were reviewed for conformance with method criteria. All method QC acceptance criteria were met.

Laboratory Blanks/Equipment Blanks

Laboratory method blank and equipment blank results are evaluated as to whether there are contaminants detected above the estimated detection limit (EDL).

Target compounds were detected in the laboratory method blank and equipment blank associated with the samples in this data set.

Detected compounds are summarized in Attachment A in Table A-1 and Table A-2. The results for the equipment blank PDI-RB-XF-181129 are provided for informational purposes only.

The NFG guidance stipulates that a conservative approach should be taken with regards to qualification of data and the reporting of false negative results should be avoided. Therefore, in order to avoid the reporting of false negative results, professional judgment was used to qualify the data in the following manner on the basis of laboratory method blank contamination. As allowed in the NFG, a blank action limit (BAL) was determined as five times the blank result:

- When the sample results were $<$ the blank result, the sample result was qualified as nondetect (U) at the sample result.
- When the sample result was \geq the blank result and \leq the BAL, the sample result was qualified as estimated and potentially biased high (J+).
- When the sample result was $>$ the BAL, the sample result was not qualified.

Qualified sample results are summarized in Table 1.

MS/MSD Results

MS/MSD analyses were not performed on a sample in this data set. No data validation actions were taken on this basis.

OPR Results

The OPR %Rs were reviewed for conformance with the method QC acceptance criteria. All method QC acceptance criteria were met.

The laboratory noted in the case narrative that the solvent went to dryness for the OPR and samples PDI-WS-T06-1811 and PDI-RB-XF-181129 during sample extraction. A new OPR was created and extracted, replacing the original OPR. The OPR was re-united with the batch and the clean-up of the samples continued. Additional sample was not available for the re-extraction of these samples. No data validation actions were taken on this basis.

Field Duplicate Results

A field duplicate pair was not submitted with this data set. No data validation actions were taken on this basis.

Labeled Compound Recoveries

The labeled compound %Rs were reviewed for conformance with the QC acceptance criteria. The laboratory noted in the case narrative that the solvent went dry during the extraction procedure of samples PDI-WS-T06-1811 and PDI-RB-XF-181129 along with the associated OPR. Additional sample was not available for the re-extraction of these samples. However, given that the extract went dry after the addition of the labeled extraction standards and that isotope dilution quantification corrects for such losses, data are not considered significantly impacted. All labeled extraction standards met criteria in sample PDI-WS-T06-1811.

Nonconformances noted for samples PDI-RB-XF-181129 and PDI-WS-T07-1811 are summarized in Attachment A in Table A-3. Samples were qualified as follows:

Actions: (Based on NFG 2016)

Criteria ¹		Actions ²	
		Detected	Nondetected
%R > Upper Acceptance Limit		J	UJ
%R >10% but < Lower Acceptance Limit		J	UJ
%R <10%		See below	
<10% and S/N >10:1		J	R
<10% and S/N <10:1		R	R
Ion abundance ratio criteria not met	Calibration compliant	J	UJ
	Calibration non-compliant	J	R
Clean-up Standard Recovery < Lower Acceptance Limit		J	UJ
¹ See Table 7 in method 1613B for acceptance criteria ² The dioxin method is performed using isotope dilution technique; therefore, professional judgment was applied and bias codes were not included in data qualification.			

Qualified sample results are summarized in Table 1.

Sample Results/Reporting Issues

All sample results detected at concentrations less than the lowest calibration standard but greater than the EDL are qualified by the laboratory as estimated (J). This “J” qualifier is retained during data validation.

Estimated Maximum Possible Concentrations (EMPCs)

The data were reviewed to identify sample results that were indicated by the laboratory to be estimated maximum possible concentrations (EMPCs) because of identification criteria not being met.

The laboratory identified the presence of EMPCs for the samples in this data set by qualifying affected results with a K laboratory qualifier. Samples were qualified as follows:

Actions: (Based on AECOM professional judgment)

Criteria	Actions
A native target compound was reported by the laboratory as an EMPC.	Report result as an EMPC and qualify as estimated and presumptively present (JN).
A labeled compound was flagged by the laboratory indicating all identification criteria were not met.	Qualify associated positive and nondetect results as estimated (J/UJ).

It should be noted that in instances of multiple nonconformances, the bias is considered indeterminate where there is a conflicting low and high bias or when a result does not exhibit a consistent bias. These results have an overall qualification of estimated (J) with the exception noted below.

When applicable, the "JN" qualifier was retained rather than replacement with the conventional overall "J" qualifier in instances where EMPC results were qualified for multiple quality control nonconformances.

Qualified sample results are summarized in Table 1.

QUALIFICATION ACTIONS

Sample results qualified as a result of validation actions are summarized in Table 1. All actions are described above.

ATTACHMENTS

Attachment A: Nonconformance Summary Tables

Attachment B: Qualifier Codes and Explanations

Attachment C: Reason Codes and Explanations

Table 1 - Data Validation Summary of Qualified Data

Sample ID	Matrix	Compound	Result	EDL	Units	Validation Qualifiers	Validation Reason
PDI-RB-XF-181129	WQ	Benz(a)anthracene	2.26	1.10	ng/sample	JN	bl,k
PDI-RB-XF-181129	WQ	Benzo(a)pyrene		9.06	ng/sample	UJ	lc
PDI-RB-XF-181129	WQ	Chrysene	2.10	1.00	ng/sample	J+	bl
PDI-WS-T01-1811	WS	Benz(a)anthracene	44.5	0.979	ng/sample	JN	k
PDI-WS-T01-1811	WS	Dibenz(a,h)anthracene	7.89	1.64	ng/sample	J+	bl
PDI-WS-T01-1811	WS	Indeno(1,2,3-cd)pyrene	45.8	1.24	ng/sample	JN	k
PDI-WS-T01-1811	WS	Naphthalene	35.6	3.24	ng/sample	J+	bl
PDI-WS-T02-1811	WS	Benz(a)anthracene	71.2	0.698	ng/sample	JN	k
PDI-WS-T02-1811	WS	Dibenz(a,h)anthracene	12.0	1.17	ng/sample	J+	bl
PDI-WS-T02-1811	WS	Naphthalene	53.2	5.11	ng/sample	J+	bl
PDI-WS-T03-1811	WS	Benz(a)anthracene	47.7	1.22	ng/sample	JN	k
PDI-WS-T03-1811	WS	Dibenz(a,h)anthracene	8.25	1.81	ng/sample	J+	bl
PDI-WS-T03-1811	WS	Indeno(1,2,3-cd)pyrene	46.7	1.76	ng/sample	JN	k
PDI-WS-T03-1811	WS	Naphthalene	34.4	4.05	ng/sample	J+	bl
PDI-WS-T04-1812	WS	Benz(a)anthracene	34.5	0.642	ng/sample	JN	k
PDI-WS-T04-1812	WS	Dibenz(a,h)anthracene	6.26	1.00	ng/sample	J+	bl
PDI-WS-T04-1812	WS	Indeno(1,2,3-cd)pyrene	33.7	2.65	ng/sample	JN	k
PDI-WS-T04-1812	WS	Naphthalene	28.4	3.96	ng/sample	J+	bl
PDI-WS-T05-1811	WS	Benz(a)anthracene	45.3	1.18	ng/sample	JN	k
PDI-WS-T05-1811	WS	Dibenz(a,h)anthracene	8.62	2.45	ng/sample	J+	bl
PDI-WS-T05-1811	WS	Naphthalene	44.0	3.57	ng/sample	J+	bl
PDI-WS-T06-1811	WS	Benz(a)anthracene	19.1	0.827	ng/sample	JN	k
PDI-WS-T06-1811	WS	Benzo(a)pyrene	17.6	2.85	ng/sample	JN	k
PDI-WS-T06-1811	WS	Dibenz(a,h)anthracene	4.24	2.44	ng/sample	JN	bl,k
PDI-WS-T06-1811	WS	Indeno(1,2,3-cd)pyrene	19.8	2.47	ng/sample	JN	k
PDI-WS-T07-1811	WS	Benz(a)anthracene	16.8	0.880	ng/sample	JN	k
PDI-WS-T07-1811	WS	Benzo(j,k)fluoranthene	24.4	1.65	ng/sample	JN	k
PDI-WS-T07-1811	WS	Dibenz(a,h)anthracene	5.46	2.43	ng/sample	J+	bl
PDI-WS-T07-1811	WS	Indeno(1,2,3-cd)pyrene	31.0	1.77	ng/sample	JN	k
PDI-WS-T07-1811	WS	Naphthalene	28.9	6.08	ng/sample	JN	bl,lc,k

Attachment A

Nonconformance Summary Tables

Table A-1 - Lab Blanks

Blank ID	Compound	Result	QL	BAL	Units	Associated Samples
WG66481-101	Indeno(1,2,3-cd)pyrene	2.21	77.6	11.0	ng/sample	PDI-RB-XF-181129
	Chrysene	1.33	77.5	6.65	ng/sample	PDI-WS-T01-1811
	Dibenz(a,h)anthracene	2.63	76.1	13.2	ng/sample	PDI-WS-T02-1811
	Benz(a)anthracene	0.720	77.7	3.60	ng/sample	PDI-WS-T03-1811
	Naphthalene	13.7	76.1	68.5	ng/sample	PDI-WS-T04-1812 PDI-WS-T05-1811 PDI-WS-T06-1811 PDI-WS-T07-1811

Table A-2 - Field Blanks

Blank ID	Compound	Result	QL	Units	Associated Samples
PDI-RB-XF-181129	Chrysene	2.10	80.0	ng/sample	PDI-WS-T01-1811
	Benz(a)anthracene	2.26	80.1	ng/sample	PDI-WS-T02-1811 PDI-WS-T03-1811
	Naphthalene	109	78.5	ng/sample	PDI-WS-T04-1812 PDI-WS-T05-1811 PDI-WS-T06-1811 PDI-WS-T07-1811

Table A-3 - Labeled Compound Recoveries

Sample ID	Labeled Compound	% Recovery	Lower Limit	Upper Limit
PDI-RB-XF-181129	Benzo[a]pyrene-D12	16.4	30	130
PDI-WS-T07-1811	Naphthalene-d8	14.4	15	130

Attachment B
Qualifier Codes and Explanations

Qualifier	Explanation
J	The analyte was positively identified; the associated numerical value is the approximate concentration of the analyte in the sample.
J-	The analyte was positively identified; the associated numerical value is the approximate concentration of the analyte in the sample with a potential low bias.
J+	The analyte was positively identified; the associated numerical value is the approximate concentration of the analyte in the sample with a potential high bias.
JN	The analyte was tentatively identified; the associated numerical value is the approximate concentration of the analyte in the sample.
UJ	The analyte was not detected above the reported sample quantitation limit. However, the reported quantitation limit is approximate and may or may not represent the actual limit of quantitation necessary to accurately and precisely measure the analyte in the sample.
U	The analyte was analyzed for, but was not detected above the reported sample quantitation limit.
R	The sample results are rejected due to serious deficiencies in the ability to analyze the sample and meet quality control criteria. The presence or absence of the analyte cannot be verified.

Attachment C

Reason Codes and Explanations

Reason Code	Explanation
be	Equipment blank contamination
bf	Field blank contamination
bl	Laboratory blank contamination
c	Calibration issue
cl	Clean-up standard recovery
d	Reporting limit raised due to chromatographic interference
fd	Field duplicate RPDs
h	Holding times
i	Internal standard areas
k	Estimated Maximum Possible Concentration (EMPC)
l	LCS or OPR recoveries
lc	Labeled compound recovery
ld	Laboratory duplicate RPDs
lp	Laboratory control sample/laboratory control sample duplicate RPDs
m	Matrix spike recovery
md	Matrix spike/matrix spike duplicate RPDs
nb	Negative laboratory blank contamination
p	Chemical preservation issue
r	Dual column RPD
q	Quantitation issue
s	Surrogate recovery
su	Ion suppression
t	Temperature preservation issue
x	Percent solids
y	Serial dilution results
z	ICS results